

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM062118\
 Data File : BM015729.D
 Acq On : 21 Jun 2018 11:35
 Operator : SJ/JU
 Sample : PB110366BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Jun 22 01:27:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060818MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 22 01:26:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.32	152	91214	20.00	ng/ul	0.00
18) Naphthalene-d8	11.15	136	402311	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.94	164	226137	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.66	188	510138	20.00	ng/ul	0.00
77) Chrysene-d12	21.77	240	549914	20.00	ng/ul	0.00
85) Perylene-d12	24.18	264	521971	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.54	96	12014	6.40	ng/uL	0.00
5) Phenol-d5	7.47	99	219870	26.55	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.63	67	148856	29.75	ng/ul	0.00
9) 2-Chlorophenol-d4	7.85	132	195221	30.83	ng/ul	0.00
13) 4-Methylphenol-d8	9.03	113	183976	25.79	ng/ul	0.00
19) Nitrobenzene-d5	9.50	128	88377	32.67	ng/ul	0.00
22) 2-Nitrophenol-d4	10.23	143	97483	33.51	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.77	165	182197	30.24	ng/ul	0.00
29) 4-Chloroaniline-d4	11.29	131	251581	37.47	ng/ul	0.00
43) Dimethylphthalate-d6	14.34	166	616844	32.95	ng/ul	0.00
46) Acenaphthylene-d8	14.64	160	709063	35.12	ng/ul	0.00
51) 4-Nitrophenol-d4	15.13	143	77944	21.57	ng/ul	0.00
57) Fluorene-d10	15.92	176	517415	32.66	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.05	200	70555	22.96	ng/ul	0.00
70) Anthracene-d10	17.76	188	816365	35.10	ng/ul	0.00
78) Pyrene-d10	20.02	212	924150	39.07	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.03	264	926429	35.41	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM062118\
Data File : BM015729.D
Acq On : 21 Jun 2018 11:35
Operator : SJ/JU
Sample : PB110366BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Jun 22 01:27:18 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060818MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 22 01:26:58 2018
Response via : Initial Calibration

