

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062119\
 Data File : BM021087.D
 Acq On : 21 Jun 2019 17:23
 Operator : HP/JU
 Sample : K3309-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 KY032MW006-190610

Manual Integrations
 APPROVED

mohammad
 6/25/2019 3:23:04 PM

Quant Time: Jun 27 08:45:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM062019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 20 15:39:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	320766	20.00	ng	0.00
21) Naphthalene-d8	10.58	136	1190225	20.00	ng	0.01
39) Acenaphthene-d10	14.43	164	723495	20.00	ng	0.00
64) Phenanthrene-d10	17.17	188	1563787	20.00	ng	0.00
76) Chrysene-d12	21.35	240	1440645	20.00	ng	0.00
87) Perylene-d12	23.63	264	1619599	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	1096210	61.88	ng	0.02
7) Phenol-d6	6.98	99	959724	37.20	ng	0.03
23) Nitrobenzene-d5	8.96	82	1844011	80.74	ng	0.02
42) 2,4,6-Tribromophenol	15.93	330	1607011	118.37	ng	0.01
45) 2-Fluorobiphenyl	13.05	172	4933721	83.81	ng	0.01
79) Terphenyl-d14	19.79	244	5856828	75.93	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.33	88	19653534	2944.752	ng	95
4) n-Nitrosodimethylamine	3.63	42	1293678	175.060	ng	# 1
10) Phenol	7.01	94	4290394	162.794	ng	99
11) bis(2-Chloroethyl)ether	7.23	93	6836917	327.666	ng	97
12) 1,3-Dichlorobenzene	7.67	146	276057	10.451	ng	# 94
13) 1,4-Dichlorobenzene	7.82	146	1347010	50.159	ng	99
14) 1,2-Dichlorobenzene	8.14	146	6357517	241.568	ng	98
17) 2-Methylphenol	8.26	107	5665590	300.958	ng	99
20) 3+4-Methylphenols	8.64	107	33683611m	1299.896	ng	
22) Acetophenone	8.63	105	632448m	20.852	ng	
27) 2,4-Dimethylphenol	9.85	122	14934534m	924.358	ng	
31) Naphthalene	10.63	128	2167687	34.415	ng	95
37) 2-Methylnaphthalene	12.25	142	196462	4.054	ng	93
38) 1-Methylnaphthalene	12.46	142	117811	2.549	ng	# 97
50) Dimethylphthalate	13.90	163	331176	5.148	ng	97
60) Diethylphthalate	15.26	149	383051	6.046	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062119\
Data File : BM021087.D
Acq On : 21 Jun 2019 17:23
Operator : HP/JU
Sample : K3309-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
KY032MW006-190610

Manual Integrations
APPROVED

mohammad
6/25/2019 3:23:04 PM

Quant Time: Jun 27 08:45:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM062019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 20 15:39:58 2019
Response via : Initial Calibration

