


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062121\
 Data File : BM030496.D
 Acq On : 21 Jun 2021 09:20
 Operator : CG/JU
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP051

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM060621.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jun 21 10:48:13 2021



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	93209	PASS
68	69	0.00	2	1.7	1500	PASS
69	198	0.00	100	37.5	89533	PASS
70	69	0.00	2	0.5	476	PASS
127	198	10	80	47.0	112357	PASS
197	198	0.00	2	0.5	1171	PASS
198	198	100	100	100.0	238976	PASS
199	198	5	9	6.6	15747	PASS
275	198	10	60	29.7	70904	PASS
365	198	1	100	4.2	10143	PASS
441	198	0.01	100	11.9	28483	PASS
442	198	50	100	77.7	185792	PASS
443	442	15	24	19.1	35507	PASS

m/z	Abundance
38.90	240.0
39.90	693.0
41.10	361.0
42.90	299.0
44.00	1606.0
45.00	217.0
49.00	363.0
51.00	252.0
54.90	220.0
57.00	196.0
69.00	150.0
73.00	907.0
77.00	215.0
83.90	196.0
85.80	176.0
96.10	161.0
133.00	285.0
134.90	194.0
147.00	279.0
191.00	283.0
193.00	354.0
207.00	2428.0
207.90	511.0
209.00	302.0
221.00	187.0
281.00	733.0
281.90	231.0
283.10	171.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

m/z	Abundance
37.10	658.0
38.00	1346.0
39.00	9992.0
40.00	1144.0
41.00	549.0
42.10	172.0
43.10	261.0
44.00	1625.0
44.90	348.0
47.10	176.0
49.00	1089.0
50.00	33040.0
51.00	126440.0
52.00	6558.0
53.00	334.0
54.90	785.0
56.00	3441.0
57.00	7638.0
58.00	483.0
59.00	230.0
60.00	177.0
61.10	1369.0
62.00	1776.0
63.00	4453.0
64.00	498.0
65.00	2418.0
65.90	223.0
66.90	229.0
68.00	1938.0
69.00	118040.0
70.00	569.0
70.90	185.0
73.10	1713.0
74.00	12938.0
75.00	20944.0
76.10	6625.0
77.00	137472.0
78.00	10116.0
79.00	9115.0
80.00	6894.0
81.00	9797.0
82.00	2004.0
83.00	2353.0
84.00	543.0
85.00	1627.0
85.90	2856.0
87.00	1248.0
87.90	502.0
89.10	296.0
91.00	2079.0
92.00	2540.0
93.00	15917.0
94.00	1158.0
95.00	328.0
96.00	949.0
97.10	271.0
98.00	12668.0
99.00	10675.0
100.00	900.0
101.00	5506.0
101.90	347.0
102.90	1817.0
104.00	3270.0
105.00	2944.0
106.10	872.0
107.00	40304.0
108.00	6214.0
109.00	1256.0
110.00	67600.0
111.00	10960.0
112.00	1499.0
112.90	505.0
115.00	176.0
116.00	1893.0
117.00	37584.0
118.00	2865.0
119.10	367.0
120.00	551.0
120.90	270.0
122.00	3158.0
123.00	4733.0
124.00	1945.0
125.00	1825.0
126.10	587.0
127.00	140928.0
128.00	10711.0
129.00	54696.0
130.00	4258.0
131.00	754.0
132.00	411.0
132.90	440.0
134.00	1786.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

135.00	4906.0
135.90	1874.0
137.00	2590.0
138.00	363.0
138.90	297.0
140.00	587.0
141.00	7095.0
142.00	2420.0
142.90	1167.0
143.90	275.0
144.90	319.0
145.90	1429.0
147.00	4094.0
148.00	9407.0
148.90	2016.0
149.90	553.0
151.00	993.0
151.80	440.0
153.00	2314.0
154.00	1577.0
155.00	3956.0
156.00	5981.0
157.00	1371.0
158.00	1476.0
159.00	1056.0
160.00	2170.0
161.10	3575.0
161.90	1022.0
163.00	553.0
163.90	452.0
164.90	2544.0
166.00	2056.0
167.00	15448.0
168.00	9464.0
169.00	1283.0
170.00	361.0
170.80	532.0
172.00	1302.0
173.00	1808.0
174.00	3304.0
175.00	5845.0
176.00	1688.0
177.00	2942.0
177.90	875.0
179.00	11134.0
180.00	8328.0
181.00	3772.0
182.00	923.0
183.00	294.0
183.90	735.0
185.00	5284.0
186.00	40992.0
187.00	12548.0
188.00	1367.0
189.00	2669.0
190.00	559.0
191.00	1409.0
192.00	3714.0
193.00	4363.0
194.00	944.0
195.10	584.0
196.00	8273.0
197.00	1199.0
198.00	281920.0
199.00	18952.0
200.00	1716.0
201.50	1136.0
203.00	2232.0
204.00	11089.0
205.00	18320.0
206.00	77888.0
207.10	13679.0
208.00	3205.0
209.00	1331.0
210.10	1315.0
211.10	3322.0
212.00	674.0
212.90	339.0
214.90	849.0
216.00	1783.0
217.00	22080.0
218.00	2792.0
219.00	379.0
221.00	11036.0
222.00	2461.0
223.00	4981.0
224.00	45432.0
225.00	10854.0
226.00	1244.0
227.00	19032.0
228.00	2693.0
229.00	3780.0
229.90	609.0
231.00	1495.0
231.90	261.0
232.90	456.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

234.00	1493.0
235.00	1466.0
236.00	814.0
237.00	2109.0
237.90	331.0
239.00	683.0
239.90	566.0
241.00	996.0
242.00	2389.0
243.00	2667.0
244.00	33712.0
245.00	4672.0
246.00	6687.0
246.90	1205.0
248.00	281.0
249.00	1470.0
249.90	272.0
251.00	358.0
252.00	547.0
253.00	593.0
254.10	893.0
255.00	180928.0
256.00	26944.0
257.00	2123.0
258.00	10402.0
259.00	1731.0
259.90	296.0
261.00	323.0
264.00	222.0
265.00	4217.0
265.90	694.0
267.00	354.0
269.90	164.0
270.90	295.0
272.00	506.0
273.00	5484.0
274.00	13196.0
275.00	78088.0
276.00	10914.0
277.00	6625.0
278.00	1357.0
278.90	325.0
281.10	589.0
282.00	440.0
282.90	838.0
284.00	537.0
285.00	1124.0
285.90	226.0
288.90	264.0
289.90	188.0
291.00	190.0
291.90	276.0
293.00	1689.0
293.90	537.0
294.90	344.0
296.00	22776.0
297.00	3594.0
297.90	205.0
300.80	278.0
301.90	326.0
303.00	2578.0
304.00	719.0
308.00	370.0
309.00	245.0
310.00	335.0
313.00	195.0
314.00	887.0
315.00	2553.0
316.00	1222.0
317.00	232.0
321.00	738.0
322.00	320.0
323.00	6467.0
324.00	1291.0
324.90	212.0
326.90	1557.0
328.00	795.0
328.90	212.0
332.00	438.0
333.00	567.0
334.00	4421.0
335.00	1342.0
341.00	993.0
341.90	240.0
346.00	1733.0
346.90	327.0
352.00	2132.0
353.00	1387.0
354.00	2026.0
355.00	626.0
365.00	10442.0
365.90	1463.0
371.00	449.0
372.00	3787.0
373.10	910.0
383.00	988.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

383.90	279.0
390.00	520.0
391.00	232.0
401.00	200.0
402.00	1398.0
403.00	1813.0
404.00	790.0
415.10	216.0
421.00	1517.0
422.10	1373.0
423.00	10276.0
424.00	3298.0
425.10	324.0
441.10	26360.0
442.10	170112.0
443.00	34000.0
444.00	3149.0
445.00	232.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

m/z	Abundance
36.90	443.0
38.00	1437.0
39.00	7680.0
40.00	1156.0
41.00	402.0
43.10	359.0
44.00	1729.0
44.90	370.0
46.90	168.0
49.10	977.0
50.00	27576.0
51.00	105048.0
52.00	5636.0
53.00	303.0
55.00	824.0
56.00	2751.0
57.00	7164.0
58.10	353.0
59.00	297.0
61.00	1241.0
62.00	1669.0
63.00	3801.0
64.00	558.0
65.00	2354.0
66.00	172.0
66.90	191.0
68.00	1746.0
69.00	102824.0
70.00	596.0
71.10	264.0
73.00	1771.0
74.00	11041.0
75.00	16648.0
76.00	5728.0
77.00	122272.0
78.10	8317.0
79.00	7977.0
80.00	5510.0
81.00	8794.0
82.00	2281.0
83.00	2096.0
83.90	460.0
85.00	1372.0
86.00	2583.0
87.00	1158.0
88.00	483.0
89.00	292.0
91.00	1742.0
92.00	2497.0
93.00	14418.0
93.90	899.0
94.90	221.0
96.10	790.0
97.00	246.0
98.00	10768.0
99.00	8366.0
100.00	608.0
101.00	5274.0
102.00	389.0
103.00	1634.0
103.90	3232.0
105.00	2682.0
105.90	941.0
107.00	36400.0
108.00	6095.0
109.10	985.0
110.00	61792.0
111.00	9886.0
112.00	922.0
113.00	524.0
115.10	259.0
116.00	2069.0
117.00	36224.0
118.00	2334.0
118.90	369.0
120.00	541.0
121.90	2596.0
123.00	4209.0
123.90	1665.0
125.00	1656.0
126.20	548.0
127.00	130880.0
128.00	9304.0
129.00	50464.0
130.00	4223.0
130.90	816.0
132.00	476.0
133.00	436.0
133.90	1269.0
135.00	4146.0
135.90	1742.0
137.00	2631.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

137.90	348.0
140.10	530.0
141.00	6610.0
142.00	2042.0
142.90	1821.0
144.00	387.0
144.90	523.0
146.00	1388.0
147.00	4046.0
148.00	9165.0
149.00	1852.0
150.00	540.0
151.00	1054.0
151.90	513.0
152.90	1928.0
153.90	2041.0
155.00	3944.0
156.00	5645.0
157.10	838.0
157.90	1457.0
158.90	1031.0
159.90	2348.0
161.00	3051.0
162.00	1051.0
162.80	494.0
164.00	531.0
165.00	2835.0
166.00	2142.0
167.00	15086.0
168.00	8836.0
169.00	1314.0
169.90	584.0
171.00	517.0
171.90	1165.0
173.00	1852.0
174.00	3204.0
175.00	5617.0
176.10	1647.0
177.00	2892.0
178.00	1003.0
179.00	11422.0
180.00	8434.0
181.00	3896.0
181.90	557.0
183.00	426.0
184.00	872.0
185.00	5170.0
186.00	41600.0
187.00	11344.0
188.10	1367.0
189.00	3067.0
190.00	391.0
191.00	1606.0
192.00	3907.0
193.00	4596.0
194.00	875.0
195.10	600.0
196.00	8229.0
197.00	1408.0
198.00	284672.0
199.00	18624.0
200.00	1325.0
201.40	1536.0
203.00	2322.0
204.00	11650.0
205.00	19432.0
206.00	78232.0
207.00	12728.0
208.00	3231.0
209.00	1295.0
210.00	1460.0
211.00	3179.0
211.80	717.0
213.00	230.0
214.90	973.0
216.00	1641.0
217.00	22536.0
218.00	3239.0
219.00	367.0
221.00	12240.0
221.80	2904.0
223.00	5060.0
224.00	44168.0
225.00	11104.0
226.10	1432.0
227.00	20720.0
228.00	3007.0
229.00	4188.0
230.00	497.0
231.00	1754.0
232.00	257.0
232.90	446.0
234.00	1139.0
235.00	1570.0
235.90	890.0
237.00	1845.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

238.00	226.0
239.00	768.0
240.00	843.0
240.90	1119.0
242.00	2464.0
243.00	2623.0
244.00	37080.0
245.00	4864.0
246.00	6546.0
247.00	1596.0
248.00	364.0
249.00	1615.0
249.90	316.0
250.90	582.0
252.00	495.0
253.00	791.0
254.10	903.0
255.00	188224.0
256.00	26152.0
257.00	2450.0
258.00	10849.0
259.00	1713.0
259.90	293.0
261.00	419.0
262.90	209.0
264.10	279.0
265.00	4591.0
265.90	589.0
267.10	216.0
270.00	218.0
271.00	355.0
271.90	493.0
273.00	5726.0
274.00	13742.0
275.00	83304.0
276.00	11104.0
277.00	7063.0
278.00	1307.0
279.00	297.0
281.00	779.0
281.90	233.0
283.00	1062.0
284.00	613.0
285.00	1337.0
285.80	191.0
289.00	293.0
290.00	265.0
291.00	188.0
292.10	239.0
293.00	1968.0
294.00	451.0
295.00	492.0
296.00	24472.0
297.00	3729.0
298.00	258.0
301.00	384.0
302.00	471.0
303.00	2668.0
304.00	753.0
307.90	330.0
308.90	274.0
310.10	295.0
312.90	209.0
314.00	1328.0
315.00	2838.0
316.00	1476.0
317.00	221.0
321.00	786.0
322.00	364.0
323.10	8083.0
324.00	1506.0
325.00	187.0
326.00	154.0
327.00	1849.0
327.90	800.0
328.90	196.0
332.00	591.0
333.00	655.0
334.00	5041.0
335.00	1080.0
339.00	230.0
340.00	189.0
341.00	1075.0
341.90	323.0
346.00	1886.0
346.90	312.0
351.00	199.0
352.00	2352.0
353.00	1811.0
354.00	2621.0
354.90	546.0
359.00	283.0
365.00	11981.0
365.90	1704.0
370.00	291.0
371.00	676.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

372.00	4007.0
373.00	1298.0
376.90	217.0
383.00	1070.0
384.10	407.0
385.00	157.0
390.10	491.0
391.00	421.0
392.00	257.0
400.90	265.0
402.00	1457.0
403.00	2360.0
404.00	997.0
405.10	176.0
415.00	180.0
421.00	1800.0
422.00	1756.0
423.00	13441.0
424.00	3404.0
425.00	440.0
441.00	33816.0
442.10	227392.0
443.10	42672.0
444.10	3822.0
445.00	230.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

m/z	Abundance
37.00	292.0
38.00	631.0
39.00	3923.0
40.00	991.0
41.00	565.0
43.10	386.0
44.00	1552.0
44.90	251.0
49.00	599.0
50.00	12902.0
51.00	48896.0
52.00	2515.0
55.00	583.0
56.00	1714.0
57.00	3648.0
58.00	281.0
60.90	606.0
62.00	620.0
63.00	2044.0
63.90	263.0
65.00	745.0
66.00	161.0
67.00	213.0
68.00	817.0
69.00	48184.0
70.00	264.0
73.00	1505.0
74.00	5171.0
75.00	7656.0
76.00	2736.0
77.00	56304.0
78.10	3993.0
79.00	3967.0
80.00	3088.0
81.00	4135.0
82.00	1006.0
82.90	951.0
83.90	469.0
85.00	750.0
85.90	1464.0
87.00	690.0
88.00	257.0
90.90	1039.0
92.00	1217.0
93.00	6430.0
93.90	426.0
95.10	187.0
96.00	446.0
97.00	172.0
98.00	5038.0
99.00	4063.0
99.90	355.0
101.00	2275.0
102.00	196.0
102.90	968.0
103.90	1461.0
105.00	1630.0
106.00	357.0
107.00	17056.0
108.00	2863.0
109.00	555.0
110.00	29520.0
111.00	4784.0
111.90	416.0
113.00	406.0
116.00	927.0
117.00	17416.0
117.90	1179.0
119.00	307.0
120.00	184.0
122.00	1405.0
123.00	2067.0
123.90	952.0
125.00	919.0
127.00	65264.0
128.00	4496.0
129.00	25000.0
130.00	2301.0
130.90	528.0
132.00	311.0
132.90	320.0
133.90	777.0
134.90	2331.0
136.00	771.0
137.00	1182.0
137.80	360.0
139.90	260.0
141.00	3063.0
141.90	1096.0
143.00	666.0
144.00	178.0
145.00	264.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

146.00	521.0
147.00	2587.0
148.00	4923.0
149.00	1036.0
150.00	294.0
150.90	617.0
151.90	447.0
152.90	1184.0
154.00	947.0
154.90	2087.0
156.00	3069.0
157.00	900.0
157.90	740.0
159.00	446.0
159.90	1182.0
161.00	1554.0
161.90	546.0
163.10	203.0
164.90	1599.0
166.00	964.0
167.00	8320.0
168.00	4830.0
169.00	653.0
170.10	318.0
171.00	377.0
172.00	677.0
173.00	703.0
174.00	1492.0
175.00	3050.0
176.00	867.0
177.00	1758.0
177.90	449.0
178.90	5934.0
180.00	3902.0
181.00	1951.0
181.90	395.0
183.00	202.0
183.90	540.0
185.00	2597.0
186.00	21504.0
187.00	5883.0
187.90	576.0
188.90	1523.0
189.80	216.0
191.00	1050.0
192.00	1960.0
193.00	2857.0
194.00	480.0
195.20	222.0
196.00	4859.0
197.00	907.0
198.00	150336.0
198.90	9664.0
199.90	773.0
201.50	983.0
203.00	1102.0
204.00	6376.0
205.00	10438.0
206.00	44704.0
207.00	8073.0
208.00	1978.0
209.00	664.0
210.20	694.0
211.10	1703.0
211.90	222.0
212.90	192.0
215.00	530.0
216.00	946.0
217.00	12821.0
218.00	1872.0
220.10	327.0
221.00	6884.0
221.90	1614.0
223.00	2785.0
224.00	25856.0
225.00	6791.0
226.00	531.0
227.00	11215.0
228.00	1580.0
229.00	2362.0
229.90	329.0
231.00	737.0
233.00	152.0
234.00	683.0
234.90	814.0
235.90	544.0
236.90	1201.0
237.90	173.0
238.90	597.0
240.00	512.0
240.90	565.0
242.00	1216.0
243.00	1468.0
244.00	20864.0
245.00	2456.0
246.00	5074.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051

246.90	625.0
247.90	164.0
249.00	810.0
249.90	152.0
251.00	276.0
252.00	248.0
253.10	465.0
254.00	982.0
255.00	111528.0
256.00	15200.0
257.00	783.0
258.00	6367.0
258.90	985.0
260.00	253.0
261.00	211.0
264.10	195.0
265.00	2930.0
266.00	459.0
267.00	222.0
269.90	214.0
271.00	302.0
272.00	357.0
273.00	3512.0
274.00	9098.0
275.00	51320.0
276.00	6528.0
277.00	4475.0
278.00	729.0
281.10	884.0
282.00	267.0
283.00	541.0
283.90	375.0
285.10	699.0
286.10	211.0
292.90	1079.0
293.90	244.0
295.00	232.0
296.00	15228.0
297.00	2116.0
298.00	152.0
301.00	187.0
302.10	230.0
303.00	1877.0
304.00	570.0
307.90	234.0
314.00	767.0
315.00	1730.0
316.00	778.0
321.00	452.0
322.00	220.0
323.00	5064.0
324.00	724.0
327.00	1182.0
328.00	706.0
328.90	195.0
331.90	418.0
333.00	502.0
334.00	3450.0
335.10	807.0
341.10	683.0
342.10	159.0
346.00	1418.0
347.00	163.0
352.00	1686.0
353.00	1085.0
354.00	1642.0
355.10	406.0
365.00	8006.0
365.90	1258.0
369.90	199.0
371.00	279.0
372.00	2685.0
373.00	560.0
383.00	724.0
389.90	304.0
391.10	284.0
400.90	247.0
402.00	1031.0
403.00	1783.0
404.00	550.0
405.00	161.0
421.00	1340.0
422.10	1302.0
423.00	8922.0
424.00	2379.0
424.90	195.0
441.00	25272.0
442.10	159872.0
443.10	29848.0
444.00	2720.0
444.90	186.0

Instrument :
BNA_M
ClientSampleId :
DFTPP051