


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062322\
 Data File : BM035618.D
 Acq On : 23 Jun 2022 10:27
 Operator : CG/JU
 Sample : DFTPP060
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP060

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM061822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 20 01:56:13 2022



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	165805	PASS
68	69	0.00	2	1.7	2740	PASS
69	198	0.00	100	39.8	159419	PASS
70	69	0.00	2	0.5	826	PASS
127	198	10	80	51.3	205525	PASS
197	198	0.00	2	0.5	1927	PASS
198	198	100	100	100.0	400448	PASS
199	198	5	9	6.7	26880	PASS
275	198	10	60	26.6	106600	PASS
365	198	1	100	3.2	12756	PASS
441	198	0.01	100	11.7	46808	PASS
442	198	50	100	77.2	309056	PASS
443	442	15	24	19.8	61165	PASS

m/z	Abundance
35.90	217.0
38.90	252.0
39.90	1188.0
41.00	435.0
42.00	332.0
43.00	465.0
44.00	2921.0
45.00	342.0
47.00	166.0
49.00	812.0
50.90	275.0
54.90	273.0
56.10	153.0
56.90	237.0
58.90	170.0
59.90	174.0
69.10	218.0
73.00	1293.0
74.80	152.0
76.90	390.0
78.00	191.0
83.10	189.0
84.00	473.0
85.80	351.0
95.90	273.0
105.10	153.0
115.00	161.0
133.00	275.0
135.00	191.0
147.10	346.0
191.10	309.0
192.90	469.0
207.00	2475.0
208.00	586.0
208.90	233.0
220.90	229.0
266.90	244.0
281.10	800.0
282.10	260.0
283.10	183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
35.90	243.0
37.10	701.0
38.00	1807.0
39.00	11386.0
39.90	2030.0
41.00	714.0
42.00	170.0
42.90	539.0
44.00	2791.0
44.90	619.0
46.90	262.0
49.00	1481.0
50.00	35456.0
51.00	145344.0
52.00	7948.0
53.00	322.0
55.00	920.0
56.00	4250.0
57.00	10035.0
58.00	395.0
58.90	339.0
60.00	270.0
60.90	2110.0
62.00	1923.0
63.00	5499.0
64.00	871.0
65.00	2887.0
66.00	205.0
67.00	250.0
68.00	2538.0
69.00	137536.0
70.00	625.0
71.00	163.0
73.00	2064.0
74.00	13380.0
75.00	22472.0
76.00	6819.0
77.00	152000.0
78.10	11321.0
79.00	9782.0
80.00	8070.0
81.00	10816.0
82.00	2938.0
83.00	2582.0
83.90	709.0
85.00	1962.0
86.00	2818.0
86.90	1357.0
87.90	555.0
88.90	206.0
91.00	2153.0
92.00	3042.0
93.00	17256.0
94.10	1253.0
94.90	404.0
96.00	1102.0
97.00	299.0
98.00	12205.0
99.00	11965.0
100.00	968.0
101.00	5877.0
102.00	453.0
103.00	2229.0
103.90	3748.0
105.00	3268.0
106.00	1347.0
107.00	43416.0
108.00	6845.0
109.00	1182.0
110.00	83792.0
111.00	12199.0
111.90	1482.0
113.00	446.0
115.00	194.0
116.00	2561.0
117.00	35024.0
117.90	3086.0
118.90	372.0
120.00	533.0
120.90	310.0
122.00	3116.0
123.00	4402.0
124.00	2153.0
125.00	1726.0
126.10	754.0
127.00	164480.0
128.00	12419.0
129.00	60384.0
130.00	6099.0
131.00	870.0
132.00	646.0
133.00	694.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

134.00	2050.0
135.00	4878.0
136.00	1814.0
137.00	2778.0
138.00	656.0
138.90	354.0
140.00	883.0
141.00	7670.0
142.00	2821.0
143.00	1639.0
144.00	546.0
145.00	427.0
145.90	1341.0
147.00	4444.0
148.00	8074.0
149.00	2292.0
150.00	601.0
151.10	1141.0
153.00	2783.0
154.00	1793.0
155.00	3617.0
156.00	6583.0
157.10	1318.0
157.90	1549.0
159.00	998.0
159.90	2432.0
161.00	3405.0
162.00	1058.0
162.90	239.0
163.90	433.0
164.90	2928.0
166.00	2031.0
167.00	14552.0
168.00	11178.0
169.00	1444.0
169.90	544.0
170.80	683.0
171.90	1194.0
173.00	1905.0
174.00	3279.0
175.00	6016.0
176.00	1875.0
177.00	3557.0
177.90	1139.0
179.00	11473.0
180.00	8930.0
181.00	3543.0
182.00	712.0
183.00	345.0
183.90	990.0
185.00	4897.0
186.00	39128.0
187.00	12081.0
188.00	1727.0
189.00	2518.0
190.00	614.0
190.90	1828.0
192.00	3360.0
193.00	4298.0
194.00	1053.0
195.00	644.0
196.00	8867.0
197.00	1530.0
198.00	293824.0
198.90	20368.0
200.00	1680.0
201.50	1394.0
202.90	1907.0
204.00	9779.0
205.00	18048.0
206.00	73616.0
207.00	12644.0
208.00	3055.0
209.00	1035.0
210.10	1239.0
211.00	2881.0
211.90	687.0
212.80	256.0
214.90	695.0
216.00	1528.0
217.00	19056.0
218.00	2578.0
219.00	186.0
221.00	12536.0
221.70	3058.0
223.00	4294.0
224.00	38824.0
225.00	10042.0
225.90	1055.0
227.00	15716.0
227.90	2438.0
229.00	3655.0
229.90	468.0
231.00	1439.0
231.90	382.0
233.10	354.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

234.00	1424.0
235.00	1005.0
235.90	819.0
237.00	2153.0
238.10	345.0
239.00	695.0
240.00	688.0
240.90	877.0
242.00	1895.0
243.10	2128.0
244.00	30336.0
245.00	4311.0
246.00	5462.0
247.00	1077.0
247.90	197.0
249.00	1211.0
250.00	193.0
251.10	275.0
251.90	258.0
253.10	716.0
254.10	1294.0
255.00	161152.0
256.00	24440.0
257.00	1891.0
258.00	8775.0
259.00	1376.0
260.10	329.0
260.90	191.0
264.00	290.0
265.00	3733.0
266.00	544.0
267.00	191.0
269.90	239.0
270.90	223.0
272.00	280.0
273.00	4866.0
274.00	11611.0
275.00	71624.0
276.00	9885.0
277.00	5750.0
278.00	1199.0
278.90	273.0
281.00	1035.0
282.00	352.0
283.00	778.0
284.00	596.0
285.00	1336.0
286.10	236.0
290.00	271.0
290.90	241.0
292.00	181.0
292.90	1099.0
294.10	379.0
295.00	439.0
296.00	18248.0
297.00	2602.0
298.00	231.0
300.90	274.0
301.90	362.0
303.00	2045.0
303.90	617.0
308.10	176.0
309.00	273.0
310.10	460.0
314.00	958.0
315.00	1718.0
316.00	1008.0
317.00	205.0
321.00	483.0
321.80	210.0
323.00	5473.0
324.00	923.0
324.90	212.0
326.00	258.0
327.00	1264.0
327.90	627.0
332.00	239.0
332.90	560.0
334.00	3447.0
335.00	1177.0
338.80	154.0
341.10	751.0
341.90	200.0
346.00	1581.0
351.00	174.0
352.00	1826.0
353.00	1266.0
354.10	1788.0
355.00	550.0
365.00	7794.0
366.00	1067.0
371.00	604.0
372.10	3138.0
373.10	727.0
374.00	182.0
383.00	780.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

383.80	176.0
390.00	407.0
391.00	325.0
391.80	331.0
401.00	236.0
402.00	1161.0
403.00	1669.0
404.00	541.0
405.00	226.0
414.80	169.0
421.10	1205.0
421.90	1138.0
423.00	9300.0
424.10	3730.0
425.10	600.0
441.10	25296.0
442.10	172800.0
443.10	34808.0
444.10	3418.0
445.00	172.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
35.90	431.0
37.00	786.0
38.00	2543.0
39.00	15120.0
40.00	2060.0
41.00	757.0
41.90	232.0
43.00	552.0
44.00	3070.0
44.90	714.0
46.80	227.0
47.80	175.0
49.00	2193.0
50.00	52096.0
51.00	199936.0
52.00	10533.0
53.00	510.0
54.90	1349.0
56.00	5897.0
57.00	14276.0
58.00	822.0
58.90	233.0
59.90	268.0
61.00	2590.0
62.00	3090.0
63.00	7703.0
64.00	1070.0
65.00	3877.0
66.00	219.0
67.00	453.0
68.00	3078.0
69.00	192192.0
70.00	949.0
73.00	3074.0
74.00	19768.0
75.00	30896.0
76.10	10958.0
77.00	218176.0
78.00	14611.0
79.00	13792.0
80.00	10809.0
81.00	14876.0
82.00	3609.0
83.00	3516.0
84.00	905.0
85.00	2895.0
86.00	4544.0
87.00	2016.0
87.90	776.0
88.90	457.0
91.00	3609.0
92.00	4234.0
93.00	23920.0
94.00	1675.0
95.00	450.0
96.00	1560.0
96.80	493.0
98.00	19632.0
99.00	15512.0
100.00	1485.0
101.00	8880.0
101.90	453.0
102.90	2837.0
104.00	5724.0
105.00	5251.0
106.00	1492.0
107.00	66256.0
108.00	9554.0
109.00	1920.0
110.00	123320.0
111.00	19888.0
112.00	2444.0
113.00	634.0
114.20	294.0
115.10	423.0
116.00	3507.0
117.00	54792.0
117.90	3880.0
119.00	481.0
120.00	804.0
121.10	350.0
122.00	4514.0
123.00	6712.0
124.00	3616.0
125.00	2807.0
126.10	1309.0
127.00	246592.0
128.00	19344.0
129.00	93864.0
130.00	8320.0
131.00	1792.0
132.00	966.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

132.80	650.0
133.90	2498.0
135.00	7592.0
136.00	3116.0
137.00	4383.0
138.00	1123.0
139.00	465.0
140.00	1221.0
141.00	11820.0
142.00	3954.0
142.90	2734.0
144.00	824.0
144.90	691.0
146.00	2199.0
147.00	6285.0
148.00	14070.0
149.00	3157.0
149.90	893.0
151.00	1750.0
151.70	1010.0
153.00	3723.0
154.00	2834.0
155.00	6732.0
156.00	10425.0
157.00	2221.0
157.90	2498.0
159.00	1562.0
160.00	3829.0
161.00	5372.0
162.00	1627.0
163.00	514.0
164.00	566.0
165.00	4423.0
166.00	3499.0
167.00	24200.0
168.00	14904.0
169.00	2169.0
170.00	888.0
170.90	931.0
171.90	2430.0
173.00	3084.0
174.00	4797.0
175.00	9229.0
176.00	2906.0
177.00	4606.0
178.00	1441.0
179.00	17208.0
180.00	13624.0
181.00	5902.0
182.10	1220.0
182.80	567.0
183.90	1255.0
185.00	8297.0
186.00	65896.0
187.00	19344.0
188.00	2404.0
189.00	3917.0
190.00	925.0
191.00	2258.0
192.00	5869.0
193.00	6385.0
194.00	1314.0
194.90	893.0
196.00	15569.0
197.00	2466.0
198.00	480960.0
199.00	32984.0
199.90	2433.0
201.50	2269.0
203.00	3185.0
204.00	16472.0
205.00	29312.0
206.00	120688.0
207.00	19056.0
208.00	4639.0
209.00	1482.0
211.10	4786.0
212.90	397.0
214.00	202.0
215.00	1519.0
216.10	2747.0
217.00	31096.0
218.00	3944.0
219.00	410.0
221.00	22048.0
221.80	4390.0
223.00	7598.0
224.00	67096.0
225.00	16936.0
226.00	2093.0
227.00	27904.0
228.00	3958.0
229.00	5765.0
230.00	670.0
231.00	2323.0
231.90	344.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

232.90	468.0
234.00	1609.0
234.90	2132.0
236.00	1409.0
237.00	3459.0
238.10	478.0
239.00	1101.0
240.00	769.0
241.00	1581.0
242.00	3273.0
243.00	3951.0
244.10	53392.0
245.00	6978.0
246.00	10031.0
247.00	1904.0
248.00	516.0
249.00	1883.0
250.00	340.0
250.90	458.0
251.90	470.0
253.00	947.0
254.10	2300.0
255.00	275776.0
256.00	39728.0
257.00	3665.0
258.00	15208.0
259.00	2448.0
259.80	431.0
260.90	438.0
263.90	468.0
265.00	6624.0
266.00	900.0
267.10	286.0
267.90	152.0
270.00	300.0
270.90	589.0
272.00	772.0
273.00	7709.0
274.00	22104.0
275.00	126552.0
276.00	16952.0
277.00	9938.0
278.00	2080.0
279.00	428.0
281.10	962.0
281.90	507.0
283.00	1502.0
284.00	824.0
285.00	1701.0
286.00	400.0
287.90	189.0
289.00	427.0
290.00	410.0
291.00	246.0
292.10	474.0
293.00	2219.0
294.00	752.0
295.00	737.0
296.00	31424.0
297.00	4648.0
297.90	353.0
301.00	556.0
302.00	497.0
303.00	3581.0
304.00	987.0
305.00	213.0
308.00	471.0
308.90	368.0
310.10	534.0
313.00	260.0
314.00	1508.0
315.00	3546.0
316.00	2325.0
317.00	229.0
321.00	848.0
322.00	570.0
323.10	10115.0
324.00	1662.0
325.00	266.0
326.00	222.0
327.00	2324.0
328.00	932.0
328.90	163.0
332.00	735.0
333.00	734.0
334.00	7058.0
335.00	2003.0
338.90	197.0
341.00	1352.0
341.90	276.0
346.00	2162.0
347.00	541.0
350.90	164.0
352.10	3289.0
353.00	1975.0
354.10	3476.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

355.00	719.0
356.00	171.0
358.90	224.0
365.00	14766.0
366.00	2007.0
369.90	261.0
371.10	791.0
372.10	5745.0
373.10	1179.0
373.90	172.0
377.00	219.0
383.00	1737.0
384.00	259.0
384.90	203.0
390.00	746.0
390.90	500.0
391.90	345.0
401.00	528.0
402.00	2350.0
403.00	3590.0
404.00	1253.0
405.00	332.0
415.00	241.0
421.00	2664.0
422.10	3028.0
423.10	19168.0
424.00	6383.0
425.10	1094.0
439.10	199.0
440.00	279.0
441.10	53616.0
442.10	349440.0
443.10	67704.0
444.00	6632.0
445.10	263.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
35.90	344.0
37.10	682.0
38.00	1818.0
39.00	11634.0
40.00	2002.0
40.90	709.0
42.00	241.0
43.00	696.0
44.00	3260.0
45.00	520.0
46.80	259.0
49.00	1328.0
50.00	40064.0
51.00	152960.0
52.10	7926.0
52.90	387.0
55.00	1082.0
56.00	4839.0
57.00	10496.0
58.00	472.0
58.90	255.0
59.80	162.0
61.00	2123.0
62.00	2439.0
63.00	5655.0
64.00	947.0
65.00	3189.0
65.90	259.0
67.00	310.0
68.10	2603.0
69.00	149184.0
70.00	905.0
73.00	2504.0
74.00	14960.0
75.00	23920.0
76.00	7700.0
77.00	165376.0
78.00	11240.0
79.00	10725.0
80.00	8238.0
81.00	12076.0
82.00	3006.0
83.00	3019.0
84.00	761.0
85.00	2335.0
86.00	3770.0
87.00	1658.0
87.90	596.0
88.90	459.0
91.00	2876.0
92.00	3060.0
93.00	18784.0
94.00	1365.0
95.00	412.0
96.00	1047.0
97.10	452.0
98.00	14785.0
99.00	13068.0
100.10	1144.0
101.00	7279.0
102.00	493.0
103.00	2553.0
104.00	4698.0
105.00	4154.0
106.00	1385.0
107.00	51952.0
108.00	8190.0
109.00	1575.0
110.00	100240.0
111.00	15032.0
112.00	1765.0
113.10	618.0
115.10	266.0
116.00	2696.0
117.00	44816.0
118.00	3281.0
119.00	564.0
119.90	647.0
121.00	277.0
122.00	3900.0
123.00	5481.0
124.00	2357.0
125.00	2753.0
127.00	205504.0
128.00	15166.0
129.00	75344.0
130.00	6185.0
131.00	1685.0
131.90	767.0
132.90	702.0
134.00	2445.0
135.00	5598.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

136.00	3044.0
137.00	3362.0
137.70	664.0
138.00	655.0
138.90	322.0
140.00	1021.0
141.00	10078.0
142.00	3420.0
143.00	2515.0
144.00	705.0
144.90	815.0
146.00	1988.0
147.00	5616.0
148.00	11693.0
148.90	2238.0
149.90	773.0
151.00	1383.0
151.90	854.0
153.00	3500.0
154.00	2610.0
155.00	5308.0
156.00	8299.0
157.00	1636.0
157.90	1978.0
158.90	1522.0
160.00	3079.0
161.00	4709.0
161.90	1294.0
163.00	352.0
164.00	755.0
165.00	4148.0
166.00	2910.0
167.00	19776.0
168.00	10880.0
169.00	1854.0
169.90	812.0
170.90	795.0
172.00	1930.0
173.00	2150.0
174.00	4226.0
175.00	8439.0
176.00	2370.0
176.90	3585.0
178.00	1625.0
179.00	15615.0
180.00	10604.0
181.00	4897.0
182.10	675.0
182.90	519.0
183.90	1262.0
185.00	6893.0
186.00	57752.0
187.00	16672.0
188.00	2067.0
189.00	3556.0
190.00	690.0
191.00	1954.0
191.90	4491.0
193.00	5652.0
194.00	1407.0
194.90	852.0
196.00	13475.0
197.00	1786.0
198.00	426560.0
199.00	27288.0
200.00	2212.0
201.60	2424.0
203.00	3185.0
204.00	14935.0
205.00	25792.0
206.00	108592.0
207.10	16984.0
208.00	4331.0
209.00	1645.0
210.00	1698.0
211.00	4299.0
213.00	334.0
214.00	159.0
214.90	1164.0
216.00	2185.0
217.00	26704.0
218.00	3735.0
219.00	357.0
221.00	18312.0
223.00	6110.0
224.00	60112.0
225.00	15382.0
226.00	1526.0
227.00	25200.0
228.00	3241.0
229.00	5343.0
230.10	720.0
231.00	2367.0
231.90	329.0
233.00	428.0
234.00	1688.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

235.10	1622.0
235.90	949.0
237.00	2488.0
238.00	337.0
238.90	787.0
239.90	882.0
241.00	1522.0
242.00	3010.0
243.00	3273.0
244.00	47856.0
245.00	6018.0
246.00	9176.0
247.00	1698.0
248.10	385.0
248.90	1824.0
249.80	297.0
250.90	397.0
251.90	387.0
252.90	1132.0
254.10	2039.0
255.00	253376.0
256.00	37456.0
257.00	2933.0
258.00	14386.0
259.00	2007.0
260.00	561.0
260.90	404.0
262.90	208.0
263.90	403.0
265.00	5537.0
266.00	994.0
266.90	259.0
269.90	316.0
271.00	481.0
272.00	790.0
273.00	7449.0
274.00	20160.0
275.00	121624.0
276.00	16067.0
277.00	8990.0
278.00	1606.0
279.00	427.0
281.00	1002.0
282.00	677.0
283.00	1362.0
284.00	706.0
285.10	1768.0
285.90	218.0
289.00	369.0
289.80	338.0
290.90	342.0
292.00	547.0
293.00	2160.0
294.00	745.0
295.00	542.0
296.00	30720.0
297.00	3896.0
298.00	399.0
301.10	663.0
302.00	708.0
303.00	3971.0
304.00	1108.0
308.00	365.0
309.00	231.0
309.80	414.0
313.00	306.0
314.00	1645.0
315.00	3439.0
316.00	1911.0
317.00	337.0
320.00	173.0
320.90	1113.0
321.80	482.0
323.00	11243.0
324.10	1910.0
325.00	180.0
326.00	301.0
327.00	1868.0
328.10	1000.0
329.00	202.0
332.00	636.0
332.90	1076.0
334.00	6810.0
335.00	1822.0
336.10	199.0
339.00	292.0
341.00	1360.0
341.90	280.0
346.00	2654.0
347.00	539.0
351.00	187.0
352.00	3121.0
353.00	2342.0
354.00	3740.0
355.00	998.0
358.90	242.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

364.10	246.0
365.00	15708.0
366.00	2369.0
370.00	279.0
370.90	690.0
372.10	5785.0
373.00	1636.0
376.90	178.0
383.00	1672.0
384.10	413.0
389.90	516.0
390.10	519.0
391.00	507.0
391.90	403.0
401.00	541.0
402.00	2590.0
403.00	4069.0
404.00	1328.0
405.00	312.0
414.80	210.0
421.00	3309.0
422.00	2849.0
423.10	21360.0
424.00	6585.0
425.00	927.0
440.00	244.0
441.10	61512.0
442.10	404928.0
443.10	80984.0
444.10	7050.0
445.10	495.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060