

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062323\
 Data File : BM040372.D
 Acq On : 23 Jun 2023 13:03
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4129

Quant Time: Jun 24 01:36:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 00:44:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.949	152	5970	0.400	ng/ul	0.00
4) Naphthalene-d8	10.762	136	21077	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.591	164	9983	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.337	188	18855	0.400	ng/ul	0.00
17) Chrysene-d12	21.515	240	14652	0.400	ng/ul	0.00
23) Perylene-d12	23.929	264	14661	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.333	96	3538	0.377	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.345	152	12205	0.413	ng/ul	0.00
18) Fluoranthene-d10	19.361	212	18551	0.395	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.367	88	3604	0.378	ng/ul#	92
5) Naphthalene	10.811	128	23477	0.404	ng/ul	100
7) 2-Methylnaphthalene	12.422	142	13865	0.406	ng/ul	99
8) 1-Methylnaphthalene	12.637	142	13927	0.408	ng/ul	100
10) Acenaphthylene	14.313	152	18120	0.406	ng/ul	100
11) Acenaphthene	14.651	153	14840	0.394	ng/ul	98
12) Fluorene	15.637	166	15817	0.393	ng/ul	98
14) Pentachlorophenol	16.995	266	454	0.125	ng/ul	90
15) Phenanthrene	17.379	178	23178	0.396	ng/ul	99
16) Anthracene	17.467	178	20153	0.409	ng/ul	99
19) Fluoranthene	19.389	202	24424	0.392	ng/ul	100
20) Pyrene	19.751	202	26157	0.392	ng/ul	98
21) Benzo(a)anthracene	21.497	228	21227	0.428	ng/ul	100
22) Chrysene	21.550	228	22888	0.405	ng/ul	99
24) Benzo(b)fluoranthene	23.192	252	23413	0.413	ng/ul	98
25) Benzo(k)fluoranthene	23.239	252	22841	0.408	ng/ul	99
26) Benzo(a)pyrene	23.821	252	20902	0.409	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	26.434	276	25690	0.361	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.454	278	20385	0.366	ng/ul	99
29) Benzo(g,h,i)perylene	27.205	276	21272	0.341	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062323\
Data File : BM040372.D
Acq On : 23 Jun 2023 13:03
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4129

Quant Time: Jun 24 01:36:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 00:44:24 2023
Response via : Initial Calibration

