


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM062618\
 Data File : BM015757.D
 Acq On : 26 Jun 2018 08:44
 Operator : SJ/JU
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060818MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jun 27 10:48:40 2018



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2503

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.9	90506	PASS
68	69	0.00	2	1.9	1562	PASS
69	198	0.00	100	49.6	81752	PASS
70	69	0.00	2	0.7	549	PASS
127	198	10	80	56.0	92336	PASS
197	198	0.00	2	0.4	578	PASS
198	198	100	100	100.0	164856	PASS
199	198	5	9	7.0	11619	PASS
275	198	10	60	29.8	49176	PASS
365	198	1	100	5.8	9621	PASS
441	443	0.01	100	79.4	19180	PASS
442	198	50	100	74.0	122008	PASS
443	442	15	24	19.8	24167	PASS

m/z	Abundance
36.90	539.0
38.00	1376.0
39.00	7189.0
40.00	361.0
41.00	563.0
43.10	367.0
44.00	353.0
49.00	803.0
50.10	24528.0
51.10	86144.0
52.10	4007.0
53.00	305.0
55.00	685.0
56.00	2545.0
57.00	4999.0
61.00	903.0
62.00	1318.0
63.00	3100.0
63.90	463.0
65.00	1418.0
68.00	1341.0
69.00	75392.0
69.90	530.0
73.00	1862.0
74.00	10372.0
75.00	12957.0
76.00	5440.0
77.10	84560.0
78.00	6257.0
79.00	7016.0
80.00	4562.0
81.00	6015.0
82.00	1583.0
83.00	1314.0
85.00	901.0
86.00	2231.0
87.00	1016.0
91.00	1655.0
92.00	1956.0
93.00	12321.0
93.90	830.0
95.00	315.0
96.10	337.0
98.00	10278.0
99.00	7238.0
100.10	648.0
101.00	3694.0
103.00	1270.0
104.00	2358.0
105.00	2063.0
106.00	757.0
107.00	26936.0
108.00	4077.0
109.10	906.0
110.00	39400.0
111.00	6938.0
112.00	845.0
116.00	1513.0
117.00	33968.0
118.00	2006.0
119.20	322.0
120.00	366.0
122.00	1936.0
123.00	2664.0
124.00	1408.0
125.00	1187.0
126.20	469.0
127.00	79944.0
128.00	6162.0
129.00	35080.0
130.00	2780.0
131.00	589.0
131.90	323.0
132.90	357.0
133.90	1087.0
135.00	3187.0
136.00	1605.0
137.00	1601.0
140.00	581.0
141.00	4928.0
142.00	1427.0
143.00	1126.0
144.90	363.0
146.00	1091.0
147.00	2786.0
148.00	7960.0
149.00	1609.0
151.00	698.0
151.60	379.0
152.10	379.0
153.00	1345.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

154.00	1197.0
155.00	2812.0
156.00	4086.0
157.00	760.0
158.00	853.0
158.90	793.0
160.00	1459.0
161.00	2111.0
161.90	704.0
165.00	2203.0
165.90	1606.0
167.00	11397.0
168.00	7258.0
168.90	1063.0
169.90	517.0
171.00	488.0
172.00	930.0
173.00	1433.0
174.00	2226.0
175.00	3657.0
176.00	1215.0
177.00	1677.0
178.00	612.0
178.90	7928.0
180.00	4966.0
181.00	2242.0
182.00	303.0
183.90	714.0
185.00	3518.0
186.00	23416.0
187.00	7800.0
188.00	730.0
188.90	1842.0
189.90	442.0
190.90	1115.0
192.00	2358.0
193.00	3079.0
194.00	621.0
195.10	326.0
196.00	4108.0
196.80	712.0
197.90	127592.0
198.90	10480.0
200.00	953.0
201.30	590.0
202.90	1646.0
204.00	6922.0
205.00	10621.0
206.00	42216.0
207.00	6949.0
208.00	1825.0
209.00	1221.0
210.00	1149.0
211.00	1837.0
211.90	300.0
214.90	777.0
215.90	1049.0
216.90	12569.0
217.90	2028.0
220.10	331.0
221.00	4838.0
222.00	1559.0
223.00	2725.0
224.00	25056.0
225.00	6968.0
226.00	704.0
227.00	13208.0
228.00	2168.0
228.90	2322.0
231.00	715.0
232.90	317.0
234.00	629.0
235.00	854.0
236.00	639.0
236.90	1566.0
238.90	511.0
240.00	315.0
241.00	785.0
242.00	1330.0
243.00	1277.0
244.00	15864.0
245.00	2531.0
246.00	5026.0
247.00	949.0
248.90	597.0
252.90	545.0
254.10	704.0
255.00	89184.0
256.00	14408.0
257.00	1610.0
258.00	6700.0
258.90	1318.0
264.90	2853.0
265.90	503.0
271.00	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

272.00	325.0
273.00	2537.0
274.00	6192.0
275.00	35976.0
276.00	4695.0
277.00	4125.0
278.00	1015.0
281.00	603.0
283.00	484.0
283.80	363.0
285.00	580.0
292.90	850.0
296.00	13952.0
297.00	1626.0
303.00	1198.0
304.00	472.0
313.90	733.0
315.00	1457.0
316.00	769.0
323.00	3554.0
324.00	602.0
326.90	732.0
328.00	377.0
333.00	493.0
334.00	2074.0
335.00	492.0
341.00	567.0
345.90	811.0
352.00	958.0
352.90	794.0
354.00	1062.0
355.00	367.0
365.00	6068.0
366.00	1068.0
371.00	310.0
372.00	1635.0
373.00	321.0
382.90	359.0
401.90	623.0
403.00	938.0
404.00	315.0
420.90	613.0
422.00	728.0
423.00	4607.0
424.00	1786.0
441.00	11661.0
442.00	71368.0
443.00	14382.0
444.10	1227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	637.0
38.00	1697.0
39.10	9158.0
40.00	638.0
41.00	519.0
44.90	324.0
49.20	1186.0
50.10	32304.0
51.10	110296.0
52.10	5893.0
52.90	322.0
55.00	800.0
56.00	3440.0
57.00	6539.0
58.00	432.0
60.90	1604.0
62.00	2054.0
63.10	4305.0
64.00	499.0
65.00	2041.0
68.10	2156.0
69.00	100544.0
69.90	687.0
73.10	1753.0
74.00	13637.0
75.00	18056.0
76.00	6556.0
77.10	116352.0
78.10	7946.0
79.00	8938.0
80.00	6479.0
81.00	9480.0
82.00	1995.0
83.00	1714.0
84.00	365.0
85.00	1472.0
86.00	2661.0
86.90	1284.0
88.00	345.0
91.00	1918.0
92.00	2527.0
93.00	17344.0
94.00	1082.0
94.90	363.0
96.00	638.0
97.20	411.0
98.00	13485.0
99.00	9239.0
100.00	719.0
101.00	4804.0
103.00	1618.0
104.00	3133.0
105.00	2894.0
106.00	846.0
107.00	37816.0
108.00	5470.0
109.10	834.0
110.00	55304.0
111.00	9091.0
112.00	1398.0
113.00	497.0
116.00	1927.0
117.00	46432.0
118.00	3477.0
119.00	400.0
120.00	429.0
122.00	2375.0
123.00	3669.0
124.00	1893.0
125.00	1648.0
126.10	475.0
127.00	113616.0
128.00	8571.0
129.00	48464.0
130.00	4304.0
131.00	1105.0
132.00	590.0
132.90	424.0
133.90	1489.0
135.00	4272.0
136.00	1737.0
137.00	2687.0
138.00	591.0
138.90	385.0
140.00	870.0
141.00	6989.0
142.00	2311.0
143.00	1763.0
143.90	359.0
144.80	305.0
145.10	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

146.00	1499.0
147.00	3788.0
148.00	10637.0
149.00	2214.0
150.10	497.0
150.90	848.0
151.80	440.0
153.00	2211.0
154.00	1626.0
155.00	3707.0
156.00	5459.0
157.00	1062.0
157.90	1491.0
159.00	1082.0
160.00	2434.0
161.00	3122.0
162.00	1009.0
163.00	315.0
163.90	413.0
165.00	3125.0
166.00	2224.0
167.00	17368.0
168.00	9824.0
169.00	1416.0
170.10	632.0
171.00	717.0
171.90	1172.0
173.00	1561.0
174.00	2926.0
175.00	5475.0
176.00	1243.0
176.90	2648.0
178.00	1130.0
179.00	11011.0
180.00	8190.0
181.00	3015.0
182.00	523.0
182.90	395.0
183.90	1020.0
185.00	6031.0
186.00	35672.0
187.00	10694.0
188.00	1080.0
189.00	3068.0
190.00	626.0
190.90	1465.0
192.00	4080.0
193.00	4618.0
194.00	1092.0
196.00	5385.0
197.10	1457.0
197.90	202688.0
198.90	13634.0
199.90	1381.0
201.00	477.0
201.60	821.0
202.90	2165.0
204.00	10677.0
205.00	17192.0
206.00	67536.0
207.00	9428.0
208.00	2872.0
209.00	1263.0
210.00	1164.0
211.00	2769.0
214.90	838.0
216.00	1669.0
217.00	20008.0
217.90	2366.0
219.00	334.0
220.20	369.0
221.00	6462.0
221.90	2841.0
223.00	4417.0
224.00	38680.0
225.00	10528.0
225.90	1164.0
227.00	20096.0
228.00	2842.0
228.90	3922.0
229.90	586.0
231.00	1259.0
233.00	503.0
233.90	1304.0
234.90	1405.0
235.90	1204.0
237.00	2198.0
238.90	899.0
240.00	690.0
240.90	1188.0
242.00	2262.0
243.00	1913.0
244.00	27344.0
245.00	3732.0
246.00	7261.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

247.00	1520.0
248.00	339.0
248.90	891.0
250.90	380.0
252.00	363.0
253.00	729.0
254.10	1069.0
255.00	148096.0
256.00	20776.0
257.00	1620.0
258.00	11593.0
259.00	1741.0
265.00	5504.0
265.90	807.0
267.00	308.0
270.90	315.0
272.00	570.0
273.00	4346.0
274.00	11176.0
275.00	57904.0
276.00	7701.0
277.00	6834.0
278.00	1295.0
281.00	545.0
281.90	356.0
282.90	855.0
283.90	455.0
284.90	1075.0
293.00	1341.0
293.90	470.0
294.80	351.0
296.00	25328.0
297.10	3195.0
297.90	307.0
302.00	378.0
303.00	2205.0
303.90	506.0
308.10	383.0
312.90	304.0
313.90	911.0
314.90	2653.0
316.00	1017.0
321.00	606.0
322.10	384.0
323.00	5903.0
324.00	1159.0
327.00	1484.0
327.90	956.0
332.00	401.0
333.00	686.0
334.00	4492.0
335.00	941.0
341.00	833.0
346.00	1488.0
347.00	381.0
352.00	1973.0
352.90	1008.0
354.00	1683.0
355.00	469.0
365.00	11331.0
365.90	1408.0
370.90	309.0
372.00	2823.0
373.00	656.0
383.00	956.0
389.90	442.0
402.00	1047.0
403.00	1611.0
404.00	497.0
421.00	1193.0
422.10	1457.0
423.00	8116.0
424.10	2958.0
425.00	300.0
441.00	22088.0
442.00	140288.0
443.10	27992.0
444.10	2945.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	394.0
38.00	1234.0
39.00	6243.0
40.00	481.0
41.00	387.0
43.90	390.0
49.00	796.0
50.10	22232.0
51.10	75080.0
52.00	3565.0
55.00	809.0
56.00	2597.0
57.00	4671.0
61.00	865.0
62.10	915.0
63.00	3137.0
64.10	340.0
65.00	1164.0
68.00	1190.0
69.00	69320.0
69.90	432.0
73.10	1626.0
74.00	8708.0
75.00	12107.0
76.00	4474.0
77.10	80392.0
78.00	5587.0
79.00	6688.0
80.00	4493.0
81.00	6692.0
82.00	1688.0
83.10	1305.0
85.00	1092.0
86.00	1727.0
87.00	912.0
87.90	419.0
90.90	1575.0
92.00	1491.0
93.00	11671.0
93.90	638.0
96.10	542.0
97.00	337.0
98.00	10777.0
99.00	6537.0
100.00	583.0
101.00	3664.0
102.90	1143.0
104.00	2532.0
105.00	2383.0
106.00	683.0
107.00	26976.0
108.00	3950.0
109.10	767.0
110.00	40800.0
111.00	6414.0
111.90	933.0
112.90	337.0
116.00	1798.0
117.00	34896.0
118.00	2678.0
119.00	368.0
119.90	488.0
121.00	319.0
122.00	1797.0
123.00	2930.0
124.00	1619.0
124.90	1159.0
126.20	537.0
127.00	83448.0
128.00	6169.0
129.00	36960.0
130.00	3130.0
130.90	485.0
131.90	314.0
134.00	1249.0
135.00	3427.0
136.00	1224.0
137.00	1721.0
137.90	420.0
139.00	357.0
140.20	451.0
141.00	5602.0
142.00	1732.0
143.00	1320.0
143.90	340.0
145.00	457.0
146.00	1244.0
147.00	3186.0
148.00	8731.0
149.00	1699.0
149.90	563.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

150.90	532.0
153.00	1843.0
154.00	1590.0
155.00	3150.0
156.00	4162.0
157.00	687.0
157.90	1183.0
159.00	742.0
159.90	1857.0
161.00	2576.0
161.90	882.0
162.90	344.0
163.90	401.0
165.00	2611.0
165.90	1656.0
167.00	14044.0
168.00	7677.0
168.90	983.0
170.00	474.0
171.00	491.0
171.90	1304.0
173.00	1205.0
174.00	2430.0
175.00	3908.0
176.00	1007.0
177.00	2188.0
178.00	791.0
178.90	9267.0
180.00	5521.0
181.00	2487.0
182.00	506.0
183.90	868.0
185.00	3450.0
186.00	28808.0
187.00	8557.0
188.00	1084.0
188.90	2351.0
190.00	374.0
191.00	1157.0
192.00	3185.0
193.00	3436.0
194.10	559.0
196.00	4745.0
196.90	1022.0
197.90	164288.0
199.00	10744.0
200.00	877.0
201.00	326.0
201.60	495.0
203.00	1834.0
204.00	8884.0
205.00	14443.0
206.00	54672.0
207.00	8371.0
208.00	2498.0
209.00	975.0
210.00	1216.0
211.00	2291.0
212.10	356.0
214.90	655.0
216.00	1525.0
217.00	16416.0
218.00	2118.0
221.00	5193.0
221.90	1698.0
223.00	3490.0
224.00	33096.0
225.00	8320.0
225.90	1016.0
227.00	16424.0
228.00	2103.0
229.00	3368.0
230.00	361.0
231.00	1032.0
232.90	351.0
233.90	1031.0
234.90	1233.0
236.00	940.0
237.00	1881.0
238.00	303.0
238.90	807.0
239.90	786.0
240.90	856.0
242.00	1620.0
243.00	1822.0
244.00	23944.0
245.00	2990.0
246.00	6386.0
247.00	1350.0
248.90	802.0
253.10	539.0
254.10	1057.0
255.00	131776.0
256.00	19432.0
257.00	1733.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

258.00	9739.0
259.00	1550.0
265.00	4302.0
266.00	622.0
270.90	354.0
272.00	362.0
273.00	3704.0
274.00	10086.0
275.00	53648.0
276.00	7370.0
277.00	6192.0
278.00	1236.0
281.00	609.0
283.00	673.0
283.90	348.0
285.00	992.0
292.90	1066.0
294.00	372.0
295.10	430.0
296.00	23064.0
297.00	3498.0
300.90	362.0
302.00	367.0
303.00	2030.0
304.00	480.0
314.00	1076.0
314.90	2505.0
316.00	1450.0
321.00	494.0
323.10	5255.0
324.00	1010.0
326.90	1219.0
328.00	736.0
332.00	478.0
333.00	453.0
334.00	3811.0
335.00	960.0
341.10	761.0
345.90	1635.0
352.00	1487.0
353.00	1147.0
354.00	1732.0
355.10	476.0
365.00	11465.0
366.00	1472.0
371.00	517.0
372.00	2958.0
373.10	628.0
383.00	911.0
384.00	362.0
390.10	408.0
390.90	357.0
402.00	1279.0
403.00	1713.0
404.00	679.0
421.00	1122.0
422.00	1370.0
423.00	8694.0
424.00	2374.0
424.90	331.0
441.00	23792.0
442.00	154368.0
443.00	30128.0
444.00	2258.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

Scan 2503 (18.010 min): BM015757.D
DFTPP66

m/z	Abundance
41.00	332.0
73.00	830.0
193.10	374.0
207.00	1682.0
207.90	370.0
280.90	377.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66