

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062623\
 Data File : BM040425.D
 Acq On : 27 Jun 2023 05:38
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4135

Quant Time: Jun 27 06:19:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 00:44:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.945	152	5439	0.400	ng/ul	0.00
4) Naphthalene-d8	10.750	136	20386	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.582	164	10038	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.332	188	19642	0.400	ng/ul	0.00
17) Chrysene-d12	21.512	240	15256	0.400	ng/ul	0.00
23) Perylene-d12	23.926	264	13840	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.329	96	2892	0.339	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.340	152	11737	0.411	ng/ul	-0.01
18) Fluoranthene-d10	19.356	212	19466	0.398	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	3086	0.355	ng/ul#	78
5) Naphthalene	10.800	128	22653	0.403	ng/ul	99
7) 2-Methylnaphthalene	12.411	142	13447	0.407	ng/ul	98
8) 1-Methylnaphthalene	12.631	142	13448	0.407	ng/ul	100
10) Acenaphthylene	14.304	152	18064	0.403	ng/ul	99
11) Acenaphthene	14.646	153	14632	0.386	ng/ul	98
12) Fluorene	15.632	166	15706	0.389	ng/ul	99
14) Pentachlorophenol	16.986	266	1532	0.405	ng/ul	93
15) Phenanthrene	17.374	178	23272	0.381	ng/ul	100
16) Anthracene	17.463	178	20606	0.402	ng/ul	99
19) Fluoranthene	19.389	202	25214	0.389	ng/ul	96
20) Pyrene	19.751	202	27169	0.391	ng/ul	99
21) Benzo(a)anthracene	21.494	228	21976	0.426	ng/ul	99
22) Chrysene	21.547	228	22441	0.381	ng/ul	99
24) Benzo(b)fluoranthene	23.186	252	21286	0.397	ng/ul	95
25) Benzo(k)fluoranthene	23.236	252	20296	0.385	ng/ul	98
26) Benzo(a)pyrene	23.818	252	18758	0.389	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	26.434	276	23248	0.346	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.447	278	18346	0.349	ng/ul	97
29) Benzo(g,h,i)perylene	27.198	276	19328	0.328	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062623\
Data File : BM040425.D
Acq On : 27 Jun 2023 05:38
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4135

Quant Time: Jun 27 06:19:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 00:44:24 2023
Response via : Initial Calibration

