

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070219\
 Data File : BM021338.D
 Acq On : 02 Jul 2019 12:08
 Operator : HP/JU
 Sample : PB121052BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK52

Quant Time: Jul 02 16:52:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 02 01:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	269057	20.00	ng/ul	0.00
18) Naphthalene-d8	10.53	136	1197949	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.39	164	760770	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.14	188	1561087	20.00	ng/ul	0.00
77) Chrysene-d12	21.33	240	1181001	20.00	ng/ul	0.00
85) Perylene-d12	23.60	264	1368277	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.27	96	32910	5.81	ng/uL	0.00
5) Phenol-d5	6.92	99	738529	31.54	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	440295	31.56	ng/ul	0.00
9) 2-Chlorophenol-d4	7.28	132	619068	32.53	ng/ul	0.00
13) 4-Methylphenol-d8	8.46	113	638057	32.64	ng/ul	0.00
19) Nitrobenzene-d5	8.92	128	337210	34.67	ng/ul	0.00
22) 2-Nitrophenol-d4	9.63	143	393776	36.47	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.16	165	740439	33.72	ng/ul	0.00
29) 4-Chloroaniline-d4	10.68	131	930869	40.65	ng/ul	0.00
43) Dimethylphthalate-d6	13.81	166	2356505	34.28	ng/ul	0.00
46) Acenaphthylene-d8	14.09	160	2804541	33.33	ng/ul	0.00
51) 4-Nitrophenol-d4	14.61	143	258133	23.09	ng/ul	0.00
57) Fluorene-d10	15.39	176	1989644	33.28	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.52	200	212228	22.12	ng/ul	0.00
70) Anthracene-d10	17.24	188	2767814	34.12	ng/ul	0.00
78) Pyrene-d10	19.54	212	2538524	35.82	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.46	264	2782944	34.47	ng/ul	0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070219\
Data File : BM021338.D
Acq On : 02 Jul 2019 12:08
Operator : HP/JU
Sample : PB121052BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK52

Quant Time: Jul 02 16:52:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 02 01:54:12 2019
Response via : Initial Calibration

