

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070323\
 Data File : BM040566.D
 Acq On : 03 Jul 2023 09:16
 Operator : MA/JU
 Sample : PB153767BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK767

Quant Time: Jul 03 13:55:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 13:40:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.923	152	5643	0.400	ng/ul	0.00
4) Naphthalene-d8	10.728	136	20366	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.568	164	10746	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.315	188	20151	0.400	ng/ul	0.00
17) Chrysene-d12	21.506	240	12353	0.400	ng/ul	0.00
23) Perylene-d12	23.914	264	14237	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.316	96	40269	4.543	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.323	152	8205	0.287	ng/ul	0.00
18) Fluoranthene-d10	19.347	212	12398	0.313	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070323\
Data File : BM040566.D
Acq On : 03 Jul 2023 09:16
Operator : MA/JU
Sample : PB153767BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK767

Quant Time: Jul 03 13:55:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 13:40:38 2023
Response via : Initial Calibration

