


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $3.40 \times 10^{-1} \times \text{Amt} - 1.61 \times 10^{-1}$
Coef of Det (r^2) = 0.990 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM070518.M
Calibration Table Last Updated: Thu Jul 05 17:39:50 2018

2,4-Dinitrophenol

Response Ratio



Resp Ratio = $1.89 \times 10^{-1} \times \text{Amt} - 6.60 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM070518.M
Calibration Table Last Updated: Thu Jul 05 17:39:50 2018