

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070519\
 Data File : BM021396.D
 Acq On : 05 Jul 2019 14:38
 Operator : HP/JU
 Sample : K3675-06
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C5AF7

Quant Time: Jul 05 23:08:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 01:58:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	197806	20.00	ng/ul	0.00
18) Naphthalene-d8	10.53	136	927491	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.39	164	618384	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.14	188	1433179	20.00	ng/ul	0.00
77) Chrysene-d12	21.32	240	1060565	20.00	ng/ul	0.00
85) Perylene-d12	23.59	264	1189680	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.27	96	22524	5.41	ng/uL	0.00
5) Phenol-d5	6.92	99	525353	30.51	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	296033	28.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.28	132	433088	30.96	ng/ul	0.00
13) 4-Methylphenol-d8	8.45	113	468803	32.62	ng/ul	0.00
19) Nitrobenzene-d5	8.90	128	241327	32.04	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	279580	33.44	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.16	165	546266	32.14	ng/ul	0.00
29) 4-Chloroaniline-d4	10.67	131	683549	38.55	ng/ul	0.00
43) Dimethylphthalate-d6	13.80	166	1756559	31.44	ng/ul	0.00
46) Acenaphthylene-d8	14.08	160	2084774	30.48	ng/ul	0.00
51) 4-Nitrophenol-d4	14.61	143	191642	21.09	ng/ul	0.00
57) Fluorene-d10	15.38	176	1566320	32.23	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.52	200	186963	21.22	ng/ul	0.00
70) Anthracene-d10	17.23	188	2336860	31.38	ng/ul	0.00
78) Pyrene-d10	19.53	212	2262866	35.56	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.44	264	2152200	30.66	ng/ul	0.00
Target Compounds						
6) Phenol	6.95	94	27419	1.684	ng/ul#	88
44) Dimethylphthalate	13.85	163	376831	7.409	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070519\
Data File : BM021396.D
Acq On : 05 Jul 2019 14:38
Operator : HP/JU
Sample : K3675-06
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C5AF7

Quant Time: Jul 05 23:08:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 03 01:58:18 2019
Response via : Initial Calibration

