

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040639.D
Acq On : 05 Jul 2023 13:56
Operator : MA/JU
Sample : PB153521BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK521

Quant Time: Jul 05 16:25:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	6721	0.400	ng/u1	0.00
4) Naphthalene-d8	10.716	136	24409	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.557	164	11855	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.305	188	20158	0.400	ng/u1	0.00
17) Chrysene-d12	21.496	240	11713	0.400	ng/u1	0.00
23) Perylene-d12	23.896	264	15923	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.312	96	60272	5.709	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.311	152	11545	0.337	ng/u1	0.00
18) Fluoranthene-d10	19.336	212	14453	0.385	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040639.D
Acq On : 05 Jul 2023 13:56
Operator : MA/JU
Sample : PB153521BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK521

Quant Time: Jul 05 16:25:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration

