

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
 Data File : BM040690.D
 Acq On : 06 Jul 2023 23:27
 Operator : MA/JU
 Sample : 03257-13
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXYY2

Quant Time: Jul 07 00:57:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 09:30:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	6973	0.400	ng/ul	0.00
4) Naphthalene-d8	10.708	136	26829	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.546	164	14389	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.295	188	24274	0.400	ng/ul	-0.01
17) Chrysene-d12	21.486	240	13117	0.400	ng/ul	0.00
23) Perylene-d12	23.883	264	16533	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	40997	3.743	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.297	152	9395	0.250	ng/ul	-0.01
18) Fluoranthene-d10	19.325	212	15767	0.375	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	530	0.048	ng/ul#	16

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040690.D
Acq On : 06 Jul 2023 23:27
Operator : MA/JU
Sample : 03257-13
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXYY2

Quant Time: Jul 07 00:57:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration

