

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
 Data File : BM040691.D
 Acq On : 07 Jul 2023 00:05
 Operator : MA/JU
 Sample : 03257-14
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXYY6

Quant Time: Jul 07 00:57:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 09:30:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.903	152	6240	0.400	ng/ul	0.00
4) Naphthalene-d8	10.707	136	22724	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.546	164	13379	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.299	188	22491	0.400	ng/ul	-0.01
17) Chrysene-d12	21.486	240	12055	0.400	ng/ul	0.00
23) Perylene-d12	23.886	264	15339	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	35628	3.635	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.297	152	10629	0.334	ng/ul	-0.01
18) Fluoranthene-d10	19.329	212	17431	0.451	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040691.D
Acq On : 07 Jul 2023 00:05
Operator : MA/JU
Sample : 03257-14
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXYY6

Quant Time: Jul 07 00:57:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration

