

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
 Data File : BM040693.D
 Acq On : 07 Jul 2023 01:20
 Operator : MA/JU
 Sample : 03257-16
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXYY9

Quant Time: Jul 07 02:04:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 09:30:35 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.903	152	6758	0.400	ng/u1	0.00
4) Naphthalene-d8	10.708	136	25383	0.400	ng/u1	# 0.00
9) Acenaphthene-d10	14.546	164	13417	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.295	188	22814	0.400	ng/u1	-0.01
17) Chrysene-d12	21.486	240	12741	0.400	ng/u1	0.00
23) Perylene-d12	23.883	264	16451	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	33409	3.147	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.297	152	8993	0.253	ng/u1	-0.01
18) Fluoranthene-d10	19.329	212	14183	0.347	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.343	88	8586	0.795	ng/u1	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 07 02:04:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.795 ng/ul
 RT: 3.343 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BM040693.D
 Acq: 07 Jul 2023 01:20

Instrument :
 BNA_M
 ClientSampleId :
 EXYY9

Tgt Ion:	88	Resp:	8586
Ion	Ratio	Lower	Upper
88	100		
43	46.7	36.1	54.1
58	69.3	49.8	74.8

