

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
 Data File : BM040699.D
 Acq On : 07 Jul 2023 05:43
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4158

Quant Time: Jul 07 06:13:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 09:30:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	6349	0.400	ng/ul	0.00
4) Naphthalene-d8	10.708	136	23669	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.546	164	10416	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.299	188	16983	0.400	ng/ul	-0.01
17) Chrysene-d12	21.486	240	10514	0.400	ng/ul	0.00
23) Perylene-d12	23.883	264	10740	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.309	96	3411	0.342	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.297	152	12372	0.373	ng/ul	-0.01
18) Fluoranthene-d10	19.329	212	13604	0.404	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	3310	0.326	ng/ul#	91
5) Naphthalene	10.757	128	24370	0.373	ng/ul	100
7) 2-Methylnaphthalene	12.374	142	14100	0.368	ng/ul	99
8) 1-Methylnaphthalene	12.588	142	13397	0.350	ng/ul	98
10) Acenaphthylene	14.268	152	17022	0.366	ng/ul	99
11) Acenaphthene	14.611	153	14134	0.360	ng/ul	99
12) Fluorene	15.601	166	14631	0.349	ng/ul	99
14) Pentachlorophenol	16.957	266	750	0.229	ng/ul	93
15) Phenanthrene	17.346	178	18964	0.359	ng/ul	99
16) Anthracene	17.434	178	15967	0.360	ng/ul	98
19) Fluoranthene	19.362	202	17173	0.385	ng/ul	100
20) Pyrene	19.720	202	17749	0.371	ng/ul	97
21) Benzo(a)anthracene	21.469	228	12845	0.361	ng/ul	99
22) Chrysene	21.521	228	13868	0.342	ng/ul	99
24) Benzo(b)fluoranthene	23.152	252	14518	0.349	ng/ul	94
25) Benzo(k)fluoranthene	23.199	252	14306	0.349	ng/ul	95
26) Benzo(a)pyrene	23.778	252	12947	0.346	ng/ul	92
27) Indeno(1,2,3-cd)pyrene	26.378	276	17777	0.341	ng/ul	94
28) Dibenzo(a,h)anthracene	26.384	278	13964	0.342	ng/ul	98
29) Benzo(g,h,i)perylene	27.139	276	15759	0.345	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040699.D
Acq On : 07 Jul 2023 05:43
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4158

Quant Time: Jul 07 06:13:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 09:30:35 2023
Response via : Initial Calibration

