

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
 Data File : BM040720.D
 Acq On : 07 Jul 2023 18:54
 Operator : MA/JU
 Sample : PB153603BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK603

Quant Time: Jul 07 20:07:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 10:22:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	4671	0.400	ng/ul	0.00
4) Naphthalene-d8	10.702	136	17907	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.541	164	10065	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.291	188	18939	0.400	ng/ul	0.00
17) Chrysene-d12	21.486	240	11029	0.400	ng/ul	0.00
23) Perylene-d12	23.877	264	12487	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.301	96	36540	4.980	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.297	152	7516	0.299	ng/ul	0.00
18) Fluoranthene-d10	19.324	212	12173	0.344	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070523\
Data File : BM040720.D
Acq On : 07 Jul 2023 18:54
Operator : MA/JU
Sample : PB153603BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK603

Quant Time: Jul 07 20:07:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 07 10:22:24 2023
Response via : Initial Calibration

