

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM070618\  
 Data File : BM015824.D  
 Acq On : 07 Jul 2018 04:52  
 Operator : SJ/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Quant Time: Jul 07 06:29:17 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM070518.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jul 05 17:39:50 2018  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 8.30  | 152  | 133226   | 20.00 | ng    | -0.01    |
| 21) Naphthalene-d8        | 11.13 | 136  | 644725   | 20.00 | ng    | -0.01    |
| 38) Acenaphthene-d10      | 14.93 | 164  | 400378   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10      | 17.66 | 188  | 935010   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 21.77 | 240  | 1111991  | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 24.16 | 264  | 972850   | 20.00 | ng    | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 5) 2-Fluorophenol           | 5.80  | 112  | 611316   | 78.69 | ng    | -0.01    |
| 7) Phenol-d6                | 7.46  | 99   | 883330   | 83.06 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.50  | 82   | 1064013  | 80.52 | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 16.41 | 330  | 469922   | 89.83 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 13.56 | 172  | 2544460  | 78.90 | ng    | 0.00     |
| 78) Terphenyl-d14           | 20.22 | 244  | 4864966  | 81.15 | ng    | 0.00     |

| Target Compounds               | R.T.  | QIon | Response | Conc   | Units | Qvalue |
|--------------------------------|-------|------|----------|--------|-------|--------|
| 2) 1,4-Dioxane                 | 3.56  | 88   | 113831   | 35.430 | ng    | # 100  |
| 3) Pyridine                    | 4.00  | 79   | 318753   | 37.597 | ng    | # 77   |
| 4) n-Nitrosodimethylamine      | 3.92  | 42   | 268060   | 40.606 | ng    | # 42   |
| 6) Aniline                     | 7.63  | 93   | 526546   | 40.837 | ng    | 90     |
| 8) 2-Chlorophenol              | 7.86  | 128  | 377480   | 40.487 | ng    | 96     |
| 9) Benzaldehyde                | 7.44  | 77   | 287613   | 42.908 | ng    | 90     |
| 10) Phenol                     | 7.49  | 94   | 437940   | 40.982 | ng    | 93     |
| 11) bis(2-Chloroethyl)ether    | 7.72  | 93   | 343900   | 39.561 | ng    | 89     |
| 12) 1,3-Dichlorobenzene        | 8.19  | 146  | 407378   | 39.409 | ng    | # 92   |
| 13) 1,4-Dichlorobenzene        | 8.34  | 146  | 420931   | 39.377 | ng    | # 94   |
| 14) 1,2-Dichlorobenzene        | 8.66  | 146  | 414707   | 40.114 | ng    | # 95   |
| 15) Benzyl Alcohol             | 8.56  | 79   | 393490   | 42.618 | ng    | # 88   |
| 16) 2,2'-oxybis(1-Chloropropan | 8.83  | 45   | 374135   | 39.409 | ng    | 97     |
| 17) 2-Methylphenol             | 8.75  | 107  | 312359   | 42.191 | ng    | # 87   |
| 18) Hexachloroethane           | 9.39  | 117  | 174997   | 41.401 | ng    | 95     |
| 19) n-Nitroso-di-n-propylamine | 9.13  | 70   | 369195   | 43.195 | ng    | # 79   |
| 20) 3+4-Methylphenols          | 9.09  | 107  | 444092   | 43.184 | ng    | 89     |
| 22) Acetophenone               | 9.15  | 105  | 666393   | 39.677 | ng    | # 88   |
| 24) Nitrobenzene               | 9.54  | 77   | 508333   | 39.520 | ng    | 98     |
| 25) Isophorone                 | 10.06 | 82   | 828525   | 41.080 | ng    | # 94   |
| 26) 2-Nitrophenol              | 10.25 | 139  | 219703   | 39.666 | ng    | # 57   |
| 27) 2,4-Dimethylphenol         | 10.30 | 122  | 329285   | 39.134 | ng    | # 85   |
| 28) bis(2-Chloroethoxy)methane | 10.53 | 93   | 516275   | 39.290 | ng    | 98     |
| 29) 2,4-Dichlorophenol         | 10.79 | 162  | 389846   | 41.833 | ng    | 97     |
| 30) 1,2,4-Trichlorobenzene     | 10.99 | 180  | 426647   | 39.417 | ng    | 96     |
| 31) Naphthalene                | 11.19 | 128  | 1324567  | 39.626 | ng    | 100    |
| 32) Benzoic acid               | 10.50 | 122  | 253104   | 33.975 | ng    | 90     |
| 33) 4-Chloroaniline            | 11.30 | 127  | 566798   | 41.590 | ng    | # 88   |
| 34) Hexachlorobutadiene        | 11.45 | 225  | 284259   | 39.351 | ng    | 97     |
| 35) Caprolactam                | 12.13 | 113  | 133173   | 43.090 | ng    | # 86   |
| 36) 4-Chloro-3-methylphenol    | 12.41 | 107  | 463728   | 43.368 | ng    | 89     |
| 37) 2-Methylnaphthalene        | 12.77 | 142  | 981724   | 41.286 | ng    | # 93   |
| 39) 1,2,4,5-Tetrachlorobenzene | 13.13 | 216  | 497862   | 38.743 | ng    | # 100  |
| 40) Hexachlorocyclopentadiene  | 13.10 | 237  | 193316   | 37.874 | ng    | 97     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM070618\  
 Data File : BM015824.D  
 Acq On : 07 Jul 2018 04:52  
 Operator : SJ/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1


Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Quant Time: Jul 07 06:29:17 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM070518.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jul 05 17:39:50 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 13.37 | 196  | 334755   | 41.211 | ng    | 99       |
| 43) 2,4,5-Trichlorophenol      | 13.45 | 196  | 353452   | 40.281 | ng #  | 86       |
| 45) 1,1'-Biphenyl              | 13.77 | 154  | 1344633  | 38.844 | ng    | 99       |
| 46) 2-Chloronaphthalene        | 13.82 | 162  | 1010709  | 39.051 | ng #  | 92       |
| 47) 2-Nitroaniline             | 14.03 | 65   | 371691   | 41.181 | ng #  | 79       |
| 48) Acenaphthylene             | 14.66 | 152  | 1697761  | 40.323 | ng    | 99       |
| 49) Dimethylphthalate          | 14.39 | 163  | 1412632  | 40.006 | ng    | 100      |
| 50) 2,6-Dinitrotoluene         | 14.52 | 165  | 285180   | 41.422 | ng #  | 68       |
| 51) Acenaphthene               | 14.99 | 154  | 1046767  | 39.954 | ng    | 97       |
| 52) 3-Nitroaniline             | 14.85 | 138  | 312153   | 41.423 | ng #  | 79       |
| 53) 2,4-Dinitrophenol          | 15.06 | 184  | 126367   | 40.468 | ng #  | 87       |
| 54) Dibenzofuran               | 15.33 | 168  | 1626321  | 40.284 | ng #  | 87       |
| 55) 4-Nitrophenol              | 15.14 | 139  | 235227   | 42.295 | ng #  | 74       |
| 56) 2,4-Dinitrotoluene         | 15.30 | 165  | 419920   | 42.247 | ng    | 93       |
| 57) Fluorene                   | 15.97 | 166  | 1385987  | 41.632 | ng    | 100      |
| 58) 2,3,4,6-Tetrachlorophenol  | 15.54 | 232  | 315669   | 42.632 | ng #  | 100      |
| 59) Diethylphthalate           | 15.72 | 149  | 1569496  | 41.478 | ng    | 96       |
| 60) 4-Chlorophenyl-phenylether | 15.95 | 204  | 705457   | 41.181 | ng #  | 89       |
| 61) 4-Nitroaniline             | 16.01 | 138  | 318630   | 40.758 | ng #  | 38       |
| 62) Azobenzene                 | 16.24 | 77   | 1504898  | 42.427 | ng    | 82       |
| 64) 4,6-Dinitro-2-methylphenol | 16.06 | 198  | 230025   | 40.468 | ng    | 85       |
| 65) n-Nitrosodiphenylamine     | 16.17 | 169  | 1259951  | 39.326 | ng    | 99       |
| 66) 4-Bromophenyl-phenylether  | 16.84 | 248  | 421703   | 40.394 | ng #  | 84       |
| 67) Hexachlorobenzene          | 16.96 | 284  | 471772   | 40.441 | ng #  | 82       |
| 68) Atrazine                   | 17.10 | 200  | 419810   | 39.837 | ng    | 99       |
| 69) Pentachlorophenol          | 17.30 | 266  | 295893   | 40.991 | ng    | 98       |
| 70) Phenanthrene               | 17.70 | 178  | 2132339  | 39.862 | ng    | 99       |
| 71) Anthracene                 | 17.79 | 178  | 2146219  | 40.855 | ng    | 99       |
| 72) Carbazole                  | 18.06 | 167  | 2013390  | 41.129 | ng    | 98       |
| 73) Di-n-butylphthalate        | 18.57 | 149  | 2797838  | 42.404 | ng #  | 98       |
| 74) Fluoranthene               | 19.68 | 202  | 2568131  | 41.761 | ng    | 99       |
| 76) Benzidine                  | 19.85 | 184  | 1192672  | 36.607 | ng    | 98       |
| 77) Pyrene                     | 20.04 | 202  | 2699369  | 38.970 | ng    | 99       |
| 79) Butylbenzylphthalate       | 20.89 | 149  | 1375395  | 41.077 | ng    | 88       |
| 80) Benzo(a)anthracene         | 21.75 | 228  | 2783062  | 39.474 | ng    | 99       |
| 81) 3,3'-Dichlorobenzidine     | 21.67 | 252  | 1013616  | 40.014 | ng    | 99       |
| 82) Chrysene                   | 21.80 | 228  | 2541663  | 38.699 | ng    | 99       |
| 83) Bis(2-ethylhexyl)phthalate | 21.63 | 149  | 2035691  | 41.147 | ng    | 96       |
| 84) Di-n-octyl phthalate       | 22.55 | 149  | 3416855  | 42.356 | ng #  | 91       |
| 85) Indeno(1,2,3-cd)pyrene     | 26.66 | 276  | 2364052  | 38.743 | ng #  | 93       |
| 87) Benzo(b)fluoranthene       | 23.44 | 252  | 2597054  | 39.984 | ng #  | 96       |
| 88) Benzo(k)fluoranthene       | 23.49 | 252  | 2496847  | 39.573 | ng    | 98       |
| 89) Benzo(a)pyrene             | 24.07 | 252  | 2287265  | 40.004 | ng    | 100      |
| 90) Dibenzo(a,h)anthracene     | 26.65 | 278  | 2091995  | 40.667 | ng    | 97       |
| 91) Benzo(g,h,i)perylene       | 27.41 | 276  | 1754172  | 38.810 | ng #  | 93       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1




1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 8.30 min Scan# 853  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

Tgt Ion:152 Resp: 133226  
Ion Ratio Lower Upper  
152 100  
150 156.8 119.5 179.3  
115 63.1 43.0 64.4




Abundance Ion 152.00 (151.70 to 152.70): E  
Ion 150.00 (149.70 to 150.70): E  
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane  
Concen: 35.430 ng  
RT: 3.56 min Scan# 47  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion: 88 Resp: 113831  
Ion Ratio Lower Upper  
88 100  
58 79.4 0.0 0.0#  
43 46.6 0.0 0.0#



Abundance Ion 88.00 (87.70 to 88.70): BM0  
Ion 58.00 (57.70 to 58.70): BM0  
Ion 43.00 (42.70 to 43.70): BM0





#3

Pyridine

Concen: 37.597 ng

RT: 4.00 min Scan# 122

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 318753

Ion Ratio Lower Upper

79 100

52 78.3 51.1 76.7#

51 50.0 26.1 39.1#



#4

n-Nitrosodimethylamine

Concen: 40.606 ng

RT: 3.92 min Scan# 107

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52




Tgt Ion: 42 Resp: 268060


Ion Ratio Lower Upper

42 100

74 73.2 119.3 178.9#

44 4.4 5.4 8.2#





#5

2-Fluorophenol

Concen: 40.606 ng

RT: 5.80 min Scan# 428

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 611316

Ion Ratio Lower Upper

112 100

64 64.6 38.6 57.8#




63 41.5 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#6

Aniline

Concen: 40.837 ng

RT: 7.63 min Scan# 738

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion: 93 Resp: 526546

Ion Ratio Lower Upper

93 100

66 44.8 30.8 46.2



65 23.6 16.0 24.0




Abundance Ion 93.00 (92.70 to 93.70): B

Ion 66.00 (65.70 to 66.70): B

Ion 65.00 (64.70 to 65.70): B





#7

Phenol-d6

Concen: 40.837 ng

RT: 7.46 min Scan# 710

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 883330

Ion Ratio Lower Upper

99 100

42 30.4 11.0 16.4#




71 45.4 23.4 35.2#



Abundance Ion 99.00 (98.70 to 99.70): BM015824.D (-693) (-)

Ion 42.00 (41.70 to 42.70): BM015824.D (-693) (-)

Ion 71.00 (70.70 to 71.70): BM015824.D (-693) (-)



#8

2-Chlorophenol

Concen: 40.487 ng

RT: 7.86 min Scan# 778

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion: 128 Resp: 377480

Ion Ratio Lower Upper

128 100

130 32.4 12.0 52.0



64 49.0 24.9 64.9




Abundance Ion 128.00 (127.70 to 128.70): BM015824.D (-763) (-)

Ion 130.00 (129.70 to 130.70): BM015824.D (-763) (-)

Ion 64.00 (63.70 to 64.70): BM015824.D (-763) (-)







#9

Benzaldehyde

Concen: 42.908 ng

RT: 7.44 min Scan# 706

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 287613

Ion Ratio Lower Upper

77 100

105 89.8 78.8 118.8

106 83.9 73.7 113.7



#10

Phenol

Concen: 40.982 ng

RT: 7.49 min Scan# 714

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52




Tgt Ion: 94 Resp: 437940

Ion Ratio Lower Upper


94 100

65 37.4 18.8 58.8

66 52.4 39.9 79.9











#11  
bis(2-Chloroethyl)ether  
Concen: 39.561 ng  
RT: 7.72 min Scan# 754  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 93 Resp: 343900  
Ion Ratio Lower Upper  
93 100  
63 82.2 49.5 89.5  
95 32.0 13.5 53.5



#12  
1,3-Dichlorobenzene  
Concen: 39.409 ng  
RT: 8.19 min Scan# 834  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion: 146 Resp: 407378  
Ion Ratio Lower Upper  
146 100  
148 63.8 50.9 76.3  
75 39.8 21.4 32.2#





#13

1,4-Dichlorobenzene

Concen: 39.377 ng

RT: 8.34 min Scan# 859

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040





Tgt Ion:146 Resp: 420931

Ion Ratio Lower Upper

146 100

148 66.3 51.9 77.9

111 45.1 29.2 43.8#



#14

1,2-Dichlorobenzene

Concen: 40.114 ng

RT: 8.66 min Scan# 914

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52




Tgt Ion:146 Resp: 414707


Ion Ratio Lower Upper

146 100

148 65.0 52.3 78.5

111 46.5 30.6 45.8#





#15

Benzyl Alcohol

Concen: 42.618 ng

RT: 8.56 min Scan# 896

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 393490

Ion Ratio Lower Upper

79 100

108 63.5 65.0 97.4#




77 68.2 53.0 79.6



Abundance Ion 79.00 (78.70 to 79.70): BM015767.D (-889) (-)

Ion 108.00 (107.70 to 108.70): BM015767.D (-889) (-)

Ion 77.00 (76.70 to 77.70): BM015767.D (-889) (-)



#16

2,2'-oxybis(1-Chloropropane)

Concen: 39.409 ng

RT: 8.83 min Scan# 942

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion: 45 Resp: 374135

Ion Ratio Lower Upper

45 100

77 14.7 0.0 35.2



79 10.2 0.0 32.0




Abundance Ion 45.00 (44.70 to 45.70): BM015767.D (-935) (-)

Ion 77.00 (76.70 to 77.70): BM015767.D (-935) (-)

Ion 79.00 (78.70 to 79.70): BM015767.D (-935) (-)





#17

2-Methylphenol

Concen: 42.191 ng

RT: 8.75 min Scan# 929

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tgt Ion:107 Resp: 312359


Ion Ratio Lower Upper

107 100

108 110.9 89.8 134.8

77 61.3 32.6 48.8#

79 58.1 32.8 49.2#






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0



#18

Hexachloroethane

Concen: 41.401 ng

RT: 9.39 min Scan# 1038

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:117 Resp: 174997

Ion Ratio Lower Upper

117 100

119 95.8 79.2 118.8

201 93.2 69.8 104.6








Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 43.195 ng

RT: 9.13 min Scan# 993

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 369195





Ion Ratio Lower Upper

70 100

42 78.2 44.5 66.7#

101 9.2 7.4 11.2

130 20.3 15.3 22.9



#20

3+4-Methylphenols

Concen: 43.184 ng

RT: 9.09 min Scan# 987

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Tgt Ion:107 Resp: 444092


Ion Ratio Lower Upper


107 100

108 87.3 73.2 113.2

77 42.4 10.7 50.7

79 37.0 9.6 49.6









#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 11.13 min Scan# 1334  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 644725 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 11.1  | 8.7   | 13.1   |
| 54       | 9.5   | 4.6   | 6.8#   |
| 68       | 4.7   | 3.7   | 5.5    |



#22  
Acetophenone  
Concen: 39.677 ng  
RT: 9.15 min Scan# 997  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 105   | Resp: | 666393 |
| Ion      | Ratio | Lower | Upper  |
| 105      | 100   |       |        |
| 71       | 1.0   | 0.6   | 1.0#   |
| 51       | 37.0  | 21.6  | 32.4#  |
| 120      | 21.7  | 16.1  | 24.1   |









#23  
 Nitrobenzene-d5  
 Concen: 39.677 ng  
 RT: 9.50 min Scan# 1056  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Tgt Ion: 82 Resp: 1064013  
 Ion Ratio Lower Upper  
 82 100  
 128 38.2 32.8 49.2  
 54 60.3 40.6 60.8




#24  
 Nitrobenzene  
 Concen: 39.520 ng  
 RT: 9.54 min Scan# 1063  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion: 77 Resp: 508333  
 Ion Ratio Lower Upper  
 77 100  
 123 40.5 32.6 49.0  
 65 15.4 10.9 16.3







#25

Isophorone

Concen: 41.080 ng

RT: 10.06 min Scan# 1152

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 828525

Ion Ratio Lower Upper

82 100

95 9.1 5.8 8.6#

138 17.5 16.3 24.5






Abundance Ion 82.00 (81.70 to 82.70): BM015767.D (-1144) (-)

Ion 95.00 (94.70 to 95.70): BM015767.D (-1144) (-)

Ion 138.00 (137.70 to 138.70): BM015767.D (-1144) (-)

Time-->



#26

2-Nitrophenol

Concen: 39.666 ng

RT: 10.25 min Scan# 1184

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion: 139 Resp: 219703

Ion Ratio Lower Upper

139 100

109 45.5 20.1 30.1#

65 66.5 31.5 47.3#






Abundance Ion 139.00 (138.70 to 139.70): BM015767.D (-1178) (-)

Ion 109.00 (108.70 to 109.70): BM015767.D (-1178) (-)

Ion 65.00 (64.70 to 65.70): BM015767.D (-1178) (-)

Time-->





#27

2,4-Dimethylphenol

Concen: 39.134 ng

RT: 10.30 min Scan# 1192

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:122 Resp: 329285

Ion Ratio Lower Upper

122 100

107 128.9 84.7 127.1#




121 60.5 46.6 69.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 39.290 ng

RT: 10.53 min Scan# 1232

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion: 93 Resp: 516275

Ion Ratio Lower Upper

93 100

95 32.3 25.5 38.3



123 12.2 8.6 13.0




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 95.00 (94.70 to 95.70): BM0

Ion 123.00 (122.70 to 123.70): E





#29

2,4-Dichlorophenol

Concen: 41.833 ng

RT: 10.79 min Scan# 1275

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 389846

Ion Ratio Lower Upper

162 100

164 63.8 42.8 82.8

98 37.3 14.5 54.5



#30

1,2,4-Trichlorobenzene

Concen: 39.417 ng

RT: 10.99 min Scan# 1310

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52




Tgt Ion:180 Resp: 426647


Ion Ratio Lower Upper

180 100

182 94.1 77.6 116.4

145 32.9 23.4 35.2









#31  
Naphthalene  
Concen: 39.626 ng  
RT: 11.19 min Scan# 1343  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion:128 Resp: 1324567  
Ion Ratio Lower Upper  
128 100  
129 10.7 8.9 13.3  
127 13.0 10.4 15.6



#32  
Benzoic acid  
Concen: 33.975 ng  
RT: 10.50 min Scan# 1226  
Delta R.T. 0.02 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:122 Resp: 253104  
Ion Ratio Lower Upper  
122 100  
105 126.6 99.9 139.9  
77 107.2 73.7 113.7









#33  
4-Chloroaniline  
Concen: 41.590 ng  
RT: 11.30 min Scan# 1363  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 127   | Resp: | 566798 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 33.2  | 25.3  | 37.9   |
| 65       | 35.6  | 17.6  | 26.4   |
| 92       | 19.9  | 13.1  | 19.7   |




Abundance Ion 127.00 (126.70 to 127.70): E  
Ion 129.00 (128.70 to 129.70): E  
Ion 65.00 (64.70 to 65.70): BM  
Ion 92.00 (91.70 to 92.70): BM






#34  
Hexachlorobutadiene  
Concen: 39.351 ng  
RT: 11.45 min Scan# 1387  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 284259 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 62.9  | 47.9  | 71.9   |
| 227      | 64.5  | 50.1  | 75.1   |



Abundance Ion 225.00 (224.70 to 225.70): E  
Ion 223.00 (222.70 to 223.70): E  
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 43.090 ng

RT: 12.13 min Scan# 1504

Delta R.T. 0.02 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:113 Resp: 133173

Ion Ratio Lower Upper

113 100

55 177.3 145.9 185.9




56 143.8 101.0 141.0#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 43.368 ng

RT: 12.41 min Scan# 1551

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:107 Resp: 463728

Ion Ratio Lower Upper

107 100

144 25.7 23.3 34.9



142 79.0 72.6 109.0




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 41.286 ng

RT: 12.77 min Scan# 1613

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:142 Resp: 981724

Ion Ratio Lower Upper

142 100

141 90.7 71.9 107.9




115 44.8 25.4 38.0#



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.93 min Scan# 1979

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:164 Resp: 400378

Ion Ratio Lower Upper

164 100

162 105.5 82.4 123.6



160 47.7 35.8 53.6




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 38.743 ng  
 RT: 13.13 min Scan# 1674  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 216   | Resp: | 497862 |
| Ion      | Ratio | Lower | Upper  |
| 216      | 100   |       |        |
| 214      | 78.1  | 0.0   | 0.0#   |
| 179      | 23.9  | 0.0   | 0.0#   |
| 108      | 23.6  | 0.0   | 0.0#   |




Abundance Ion 216.00 (215.70 to 216.70): E  
 Ion 214.00 (213.70 to 214.70): E  
 Ion 179.00 (178.70 to 179.70): E  
 Ion 108.00 (107.70 to 108.70): E





#40  
 Hexachlorocyclopentadiene  
 Concen: 37.874 ng  
 RT: 13.10 min Scan# 1668  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 237   | Resp: | 193316 |
| Ion      | Ratio | Lower | Upper  |
| 237      | 100   |       |        |
| 235      | 63.5  | 42.0  | 82.0   |
| 272      | 11.3  | 0.0   | 33.4   |



Abundance Ion 236.80 (236.50 to 237.50): E  
 Ion 234.90 (234.60 to 235.60): E  
 Ion 271.75 (271.45 to 272.45): E











#41  
 2,4,6-Tribromophenol  
 Concen: 37.874 ng  
 RT: 16.41 min Scan# 2231  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Tgt Ion:330 Resp: 469922  
 Ion Ratio Lower Upper  
 330 100  
 332 97.1 0.0 0.0#  
 141 50.1 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E  
 Ion 331.80 (331.50 to 332.50): E  
 Ion 141.00 (140.70 to 141.70): E






#42  
 2,4,6-Trichlorophenol  
 Concen: 41.211 ng  
 RT: 13.37 min Scan# 1715  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:196 Resp: 334755  
 Ion Ratio Lower Upper  
 196 100  
 198 95.3 76.3 114.5  
 200 30.5 25.6 38.4



Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 200.00 (199.70 to 200.70): E









#43  
 2,4,5-Trichlorophenol  
 Concen: 40.281 ng  
 RT: 13.45 min Scan# 1728  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:196 Resp: 353452  
 Ion Ratio Lower Upper  
 196 100  
 198 97.3 79.4 119.0  
 97 63.1 26.8 40.2#  
 132 29.1 18.6 28.0#



#44  
 2-Fluorobiphenyl  
 Concen: 40.281 ng  
 RT: 13.56 min Scan# 1746  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:172 Resp: 2544460  
 Ion Ratio Lower Upper  
 172 100  
 171 35.7 28.5 42.7  
 170 24.1 18.8 28.2





#45

1,1'-Biphenyl

Concen: 38.844 ng

RT: 13.77 min Scan# 1782

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 1344633

Ion Ratio Lower Upper

154 100

153 40.7 20.7 60.7




76 11.8 0.0 30.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BM



#46

2-Chloronaphthalene

Concen: 39.051 ng

RT: 13.82 min Scan# 1791

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:162 Resp: 1010709

Ion Ratio Lower Upper

162 100

127 41.8 26.9 40.3#



164 32.1 26.8 40.2




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47  
2-Nitroaniline  
Concen: 41.181 ng  
RT: 14.03 min Scan# 1826  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 65 Resp: 371691  
Ion Ratio Lower Upper  
65 100  
92 60.9 59.1 88.7  
138 89.8 93.9 140.9#



#48  
Acenaphthylene  
Concen: 40.323 ng  
RT: 14.66 min Scan# 1933  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion: 152 Resp: 1697761  
Ion Ratio Lower Upper  
152 100  
151 20.3 15.9 23.9  
153 13.3 10.6 16.0





#49

Dimethylphthalate

Concen: 40.006 ng

RT: 14.39 min Scan# 1887

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 1412632

Ion Ratio Lower Upper

163 100

194 3.8 2.7 4.1




164 10.2 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 41.422 ng

RT: 14.52 min Scan# 1909

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:165 Resp: 285180

Ion Ratio Lower Upper

165 100

63 81.6 44.1 66.1#



89 74.9 44.3 66.5#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 39.954 ng

RT: 14.99 min Scan# 1990

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 1046767

Ion Ratio Lower Upper

154 100

153 115.7 90.0 135.0




152 54.9 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 41.423 ng

RT: 14.85 min Scan# 1966

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:138 Resp: 312153

Ion Ratio Lower Upper

138 100

108 10.5 7.3 10.9



92 117.6 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53  
2,4-Dinitrophenol  
Concen: 40.468 ng  
RT: 15.06 min Scan# 2002  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Instrument :  
BNA\_M  
ClientSampleID :  
SSTDCCC040

Tgt Ion:184 Resp: 126367  
Ion Ratio Lower Upper  
184 100  
63 93.2 69.2 103.8  
154 82.8 53.3 79.9#




Abundance Ion 184.00 (183.70 to 184.70): E  
Ion 63.00 (62.70 to 63.70): BM0  
Ion 154.00 (153.70 to 154.70): E






#54  
Dibenzofuran  
Concen: 40.284 ng  
RT: 15.33 min Scan# 2047  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:168 Resp: 1626321  
Ion Ratio Lower Upper  
168 100  
139 44.1 27.3 40.9#  
169 13.2 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E  
Ion 139.00 (138.70 to 139.70): E  
Ion 169.00 (168.70 to 169.70): E









#55  
4-Nitrophenol  
Concen: 42.295 ng  
RT: 15.14 min Scan# 2016  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

Tgt Ion:139 Resp: 235227  
Ion Ratio Lower Upper  
139 100  
109 111.3 130.0 170.0#  
65 122.8 130.0 170.0#




#56  
2,4-Dinitrotoluene  
Concen: 42.247 ng  
RT: 15.30 min Scan# 2043  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:165 Resp: 419920  
Ion Ratio Lower Upper  
165 100  
63 60.4 52.4 78.6  
89 83.3 72.4 108.6  
182 2.6 2.0 3.0







#57

Fluorene

Concen: 41.632 ng

RT: 15.97 min Scan# 2156

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:166 Resp: 1385987

Ion Ratio Lower Upper

166 100

165 99.0 79.0 118.4




167 13.4 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 42.632 ng

RT: 15.54 min Scan# 2084

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Tgt Ion:232 Resp: 315669


Ion Ratio Lower Upper

232 100

131 56.9 0.0 0.0#

130 3.7 0.0 0.0#

166 35.1 0.0 0.0#






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59  
Diethylphthalate  
Concen: 41.478 ng  
RT: 15.72 min Scan# 2114  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

Tgt Ion:149 Resp: 1569496  
Ion Ratio Lower Upper  
149 100  
177 20.2 18.3 27.5  
150 12.6 9.8 14.8




Abundance Ion 149.00 (148.70 to 149.70): E  
Ion 177.00 (176.70 to 177.70): E  
Ion 150.00 (149.70 to 150.70): E






#60  
4-Chlorophenyl-phenylether  
Concen: 41.181 ng  
RT: 15.95 min Scan# 2153  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:204 Resp: 705457  
Ion Ratio Lower Upper  
204 100  
206 33.1 26.6 39.8  
141 68.7 45.2 67.8#



Abundance Ion 204.00 (203.70 to 204.70): E  
Ion 206.00 (205.70 to 206.70): E  
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 40.758 ng

RT: 16.01 min Scan# 2163

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:138 Resp: 318630

Ion Ratio Lower Upper

138 100

92 56.3 16.7 56.7




108 117.5 37.6 77.6#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BM0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 42.427 ng

RT: 16.24 min Scan# 2203

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Tgt Ion: 77 Resp: 1504898


Ion Ratio Lower Upper

77 100

182 22.2 6.5 46.5

105 16.7 3.8 43.8

51 41.5 5.5 45.5






Abundance Ion 77.00 (76.70 to 77.70): BM0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BM0









#63  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.66 min Scan# 2443  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion:188 Resp: 935010  
Ion Ratio Lower Upper  
188 100  
94 8.3 8.7 13.1#  
80 9.0 9.3 13.9#



#64  
4,6-Dinitro-2-methylphenol  
Concen: 40.468 ng  
RT: 16.06 min Scan# 2172  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:198 Resp: 230025  
Ion Ratio Lower Upper  
198 100  
51 65.8 28.8 68.8  
105 47.3 30.5 70.5





#65

n-Nitrosodiphenylamine

Concen: 39.326 ng

RT: 16.17 min Scan# 2190

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:169 Resp: 1259951

Ion Ratio Lower Upper

169 100

168 68.3 53.9 80.9




167 36.9 28.9 43.3



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 40.394 ng

RT: 16.84 min Scan# 2304

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:248 Resp: 421703

Ion Ratio Lower Upper

248 100

250 95.2 77.4 116.0



141 87.2 45.2 67.8#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67  
Hexachlorobenzene  
Concen: 40.441 ng  
RT: 16.96 min Scan# 2325  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion:284 Resp: 471772  
Ion Ratio Lower Upper  
284 100  
142 41.6 20.4 30.6#  
249 33.3 23.8 35.6



#68  
Atrazine  
Concen: 39.837 ng  
RT: 17.10 min Scan# 2349  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:200 Resp: 419810  
Ion Ratio Lower Upper  
200 100  
173 25.2 4.8 44.8  
215 46.0 26.9 66.9









#69  
 Pentachlorophenol  
 Concen: 40.991 ng  
 RT: 17.30 min Scan# 2383  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:266 Resp: 295893  
 Ion Ratio Lower Upper  
 266 100  
 268 64.9 52.0 78.0  
 264 63.6 49.0 73.4



#70  
 Phenanthrene  
 Concen: 39.862 ng  
 RT: 17.70 min Scan# 2450  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:178 Resp: 2132339  
 Ion Ratio Lower Upper  
 178 100  
 176 19.6 15.2 22.8  
 179 15.1 12.1 18.1









#71  
 Anthracene  
 Concen: 40.855 ng  
 RT: 17.79 min Scan# 2465  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Tgt Ion:178 Resp: 2146219  
 Ion Ratio Lower Upper  
 178 100  
 176 19.2 14.6 22.0  
 179 15.7 12.6 19.0




#72  
 Carbazole  
 Concen: 41.129 ng  
 RT: 18.06 min Scan# 2511  
 Delta R.T. -0.00 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:167 Resp: 2013390  
 Ion Ratio Lower Upper  
 167 100  
 166 21.9 17.2 25.8  
 139 14.8 10.8 16.2











#73  
 Di-n-butylphthalate  
 Concen: 42.404 ng  
 RT: 18.57 min Scan# 2599  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:149 Resp: 2797838  
 Ion Ratio Lower Upper  
 149 100  
 150 9.1 7.5 11.3  
 104 6.5 3.7 5.5#



#74  
 Fluoranthene  
 Concen: 41.761 ng  
 RT: 19.68 min Scan# 2787  
 Delta R.T. -0.00 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:202 Resp: 2568131  
 Ion Ratio Lower Upper  
 202 100  
 101 9.3 0.0 28.7  
 203 17.4 0.0 37.2









#75  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.77 min Scan# 3142  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion:240 Resp: 1111991  
Ion Ratio Lower Upper  
240 100  
120 8.9 8.2 12.2  
236 26.6 20.0 30.0



#76  
Benzidine  
Concen: 36.607 ng  
RT: 19.85 min Scan# 2816  
Delta R.T. -0.01 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:184 Resp: 1192672  
Ion Ratio Lower Upper  
184 100  
185 14.0 11.3 16.9  
183 12.3 8.9 13.3









#77  
 Pyrene  
 Concen: 38.970 ng  
 RT: 20.04 min Scan# 2848  
 Delta R.T. -0.00 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:202 Resp: 2699369  
 Ion Ratio Lower Upper  
 202 100  
 200 21.6 16.9 25.3  
 203 17.3 14.1 21.1



#78  
 Terphenyl-d14  
 Concen: 38.970 ng  
 RT: 20.22 min Scan# 2878  
 Delta R.T. -0.01 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:244 Resp: 4864966  
 Ion Ratio Lower Upper  
 244 100  
 212 8.1 4.9 7.3#  
 122 11.0 6.2 9.4#









#79  
Butylbenzylphthalate  
Concen: 41.077 ng  
RT: 20.89 min Scan# 2992  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040


Tgt Ion:149 Resp: 1375395  
Ion Ratio Lower Upper  
149 100  
91 74.4 50.1 75.1  
206 19.5 16.2 24.2



#80  
Benzo(a)anthracene  
Concen: 39.474 ng  
RT: 21.75 min Scan# 3139  
Delta R.T. -0.00 min  
Lab File: BM015824.D  
Acq: 07 Jul 2018 04:52

Tgt Ion:228 Resp: 2783062  
Ion Ratio Lower Upper  
228 100  
226 27.3 21.7 32.5  
229 19.7 16.0 24.0









#81  
 3,3'-Dichlorobenzidine  
 Concen: 40.014 ng  
 RT: 21.67 min Scan# 3126  
 Delta R.T. -0.00 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:252 Resp: 1013616  
 Ion Ratio Lower Upper  
 252 100  
 254 64.8 51.9 77.9  
 126 10.8 9.8 14.8



#82  
 Chrysene  
 Concen: 38.699 ng  
 RT: 21.80 min Scan# 3148  
 Delta R.T. -0.00 min  
 Lab File: BM015824.D  
 Acq: 07 Jul 2018 04:52

Tgt Ion:228 Resp: 2541663  
 Ion Ratio Lower Upper  
 228 100  
 226 30.6 24.1 36.1  
 229 19.6 15.5 23.3





#83

Bis(2-ethylhexyl)phthalate

Concen: 41.147 ng

RT: 21.63 min Scan# 3118

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 2035691

Ion Ratio Lower Upper

149 100

167 25.8 22.4 33.6




279 4.0 3.4 5.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 42.356 ng

RT: 22.55 min Scan# 3275

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:149 Resp: 3416855

Ion Ratio Lower Upper

149 100

167 1.5 1.2 1.8



43 12.8 7.3 10.9#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BM





#85

Indeno(1,2,3-cd)pyrene

Concen: 38.743 ng

RT: 26.66 min Scan# 3973

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040





Tgt Ion:276 Resp: 2364052

Ion Ratio Lower Upper

276 100

138 22.4 12.0 18.0#

277 25.5 20.0 30.0



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.16 min Scan# 3549

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52




Tgt Ion:264 Resp: 972850


Ion Ratio Lower Upper

264 100

260 24.6 18.6 27.8

265 22.1 17.3 25.9





#87

Benzo(b)fluoranthene

Concen: 39.984 ng

RT: 23.44 min Scan# 3426

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 2597054

Ion Ratio Lower Upper

252 100

253 22.3 18.0 27.0




125 8.8 9.9 14.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 39.573 ng

RT: 23.49 min Scan# 3434

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:252 Resp: 2496847

Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8



125 9.1 8.1 12.1




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E







#89

Benzo(a)pyrene

Concen: 40.004 ng

RT: 24.07 min Scan# 3533

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 2287265

Ion Ratio Lower Upper




252 100

253 22.0 17.6 26.4

125 9.4 7.4 11.2



Abundance Ion 252.00 (251.70 to 252.70): E  
Ion 253.00 (252.70 to 253.70): E  
Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 40.667 ng

RT: 26.65 min Scan# 3972

Delta R.T. -0.01 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52


Tgt Ion:278 Resp: 2091995

Ion Ratio Lower Upper



278 100


139 14.1 13.4 20.0

279 23.8 19.0 28.4



Abundance Ion 278.00 (277.70 to 278.70): E  
Ion 139.00 (138.70 to 139.70): E  
Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 38.810 ng

RT: 27.41 min Scan# 4102

Delta R.T. -0.00 min

Lab File: BM015824.D

Acq: 07 Jul 2018 04:52

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 276 Resp: 1754172

Ion Ratio Lower Upper

276 100

277 23.8 18.5 27.7

138 18.1 19.0 28.6



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

