

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
 Data File : BM040737.D
 Acq On : 08 Jul 2023 12:52
 Operator : MA/JU
 Sample : 03332-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YBTX6

Quant Time: Jul 08 22:58:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 08 22:50:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.891	152	3423	0.400	ng/u1	0.00
4) Naphthalene-d8	10.697	136	12789	0.400	ng/u1	# 0.00
9) Acenaphthene-d10	14.537	164	6302	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.286	188	12492	0.400	ng/u1	-0.01
17) Chrysene-d12	21.483	240	8215	0.400	ng/u1	0.00
23) Perylene-d12	23.874	264	7862	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.301	96	21290	3.960	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.291	152	4554	0.254	ng/u1	0.00
18) Fluoranthene-d10	19.320	212	7211	0.274	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
Data File : BM040737.D
Acq On : 08 Jul 2023 12:52
Operator : MA/JU
Sample : 03332-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YBTX6

Quant Time: Jul 08 22:58:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 08 22:50:20 2023
Response via : Initial Calibration

