

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
 Data File : BM040769.D
 Acq On : 09 Jul 2023 10:09
 Operator : MA/JU
 Sample : 03356-05
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YBW34

Quant Time: Jul 09 14:34:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jul 09 14:07:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	3709	0.400	ng/ul	0.00
4) Naphthalene-d8	10.691	136	13430	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.532	164	8169	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.282	188	16631	0.400	ng/ul	#-0.01
17) Chrysene-d12	21.475	240	11167	0.400	ng/ul	0.00
23) Perylene-d12	23.866	264	10260	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.296	96	39069	6.706	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.286	152	9448	0.502	ng/ul	0.00
18) Fluoranthene-d10	19.315	212	18661	0.521	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
Data File : BM040769.D
Acq On : 09 Jul 2023 10:09
Operator : MA/JU
Sample : 03356-05
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YBW34

Quant Time: Jul 09 14:34:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 09 14:07:58 2023
Response via : Initial Calibration

