

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
 Data File : BM040778.D
 Acq On : 09 Jul 2023 15:45
 Operator : MA/JU
 Sample : 03356-06
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YBW35

Quant Time: Jul 09 17:52:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jul 09 14:07:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	4291	0.400	ng/ul	0.00
4) Naphthalene-d8	10.691	136	15242	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.532	164	7706	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.282	188	13769	0.400	ng/ul	-0.01
17) Chrysene-d12	21.477	240	7798	0.400	ng/ul	0.00
23) Perylene-d12	23.865	264	7601	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.296	96	35247	5.229	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.286	152	7286	0.341	ng/ul	0.00
18) Fluoranthene-d10	19.315	212	9959	0.398	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070823\
Data File : BM040778.D
Acq On : 09 Jul 2023 15:45
Operator : MA/JU
Sample : 03356-06
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YBW35

Quant Time: Jul 09 17:52:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 09 14:07:58 2023
Response via : Initial Calibration

