


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM071118\
 Data File : BM015884.D
 Acq On : 11 Jul 2018 12:10
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM070518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 05 17:39:50 2018



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.5	41261	PASS
68	69	0.00	2	2.0	745	PASS
69	198	0.00	100	48.7	36816	PASS
70	69	0.00	2	0.4	130	PASS
127	198	10	80	54.9	41552	PASS
197	198	0.00	2	0.8	601	PASS
198	198	100	100	100.0	75640	PASS
199	198	5	9	6.7	5080	PASS
275	198	10	60	29.3	22138	PASS
365	198	1	100	6.2	4715	PASS
441	443	0.01	100	79.7	8931	PASS
442	198	50	100	77.4	58565	PASS
443	442	15	24	19.1	11206	PASS