

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071123\
 Data File : BM040807.D
 Acq On : 11 Jul 2023 11:34
 Operator : MA/JU
 Sample : 03477-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 HOA41

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/12/2023
 Supervised By :mohammad ahmed 07/12/2023

Quant Time: Jul 12 04:33:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 03:54:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	3350	0.400	ng/ul	0.00
4) Naphthalene-d8	10.686	136	11863	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.527	164	5382	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.278	188	9160	0.400	ng/ul	-0.01
17) Chrysene-d12	21.475	240	5546	0.400	ng/ul	-0.01
23) Perylene-d12	23.860	264	8334m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.297	96	23066	4.383	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.275	152	4734	0.285	ng/ul	-0.01
18) Fluoranthene-d10	19.311	212	6343	0.357	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.330	88	404568	75.531	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071123\
Data File : BM040807.D
Acq On : 11 Jul 2023 11:34
Operator : MA/JU
Sample : 03477-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
HOA41

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/12/2023
Supervised By :mohammad ahmed 07/12/2023

Quant Time: Jul 12 04:33:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 03:54:24 2023
Response via : Initial Calibration

