

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071223\
 Data File : BM040834.D
 Acq On : 12 Jul 2023 12:55
 Operator : MA/JU
 Sample : 03491-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GWOW-BR-05-167-187-062923

Quant Time: Jul 13 00:12:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 12 05:37:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.882	152	2748	0.400	ng/ul	0.00
4) Naphthalene-d8	10.685	136	10679	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.527	164	4815	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.282	188	8514	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.480	240	5252	0.400	ng/ul	0.00
23) Perylene-d12	23.868	264	4273	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.280	152	4932	0.329	ng/ul	0.00
18) Fluoranthene-d10	19.310	212	6767	0.402	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.326	88	149341	33.989	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071223\
Data File : BM040834.D
Acq On : 12 Jul 2023 12:55
Operator : MA/JU
Sample : 03491-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GWOW-BR-05-167-187-062923

Quant Time: Jul 13 00:12:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 12 05:37:29 2023
Response via : Initial Calibration

