


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071318\
 Data File : BM015937.D
 Acq On : 13 Jul 2018 10:36
 Operator : SJ/JU
 Sample : DFTPP28
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP28

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM071318MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jul 13 23:45:44 2018



AutoFind: Scans 2502, 2503, 2504; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.3	55392	PASS
68	69	0.00	2	2.0	942	PASS
69	198	0.00	100	48.9	47264	PASS
70	69	0.00	2	0.2	109	PASS
127	198	10	80	57.0	55093	PASS
197	198	0.00	2	0.6	621	PASS
198	198	100	100	100.0	96672	PASS
199	198	5	9	6.7	6522	PASS
275	198	10	60	29.2	28216	PASS
365	198	1	100	6.7	6475	PASS
441	198	0.01	100	11.7	11357	PASS
442	198	50	100	70.8	68434	PASS
443	442	15	24	19.7	13464	PASS

m/z	Abundance
44.00	478.0
73.10	525.0
192.90	301.0
206.90	817.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

m/z	Abundance
38.10	662.0
39.10	2546.0
39.90	385.0
44.00	482.0
49.00	410.0
50.00	11004.0
51.00	33744.0
52.00	1405.0
56.00	1090.0
56.90	1933.0
61.00	468.0
62.00	526.0
63.10	1179.0
65.00	493.0
68.00	596.0
69.00	30816.0
73.00	1031.0
74.00	4459.0
75.00	6041.0
76.10	1966.0
77.10	38120.0
78.00	2665.0
79.00	3124.0
80.00	1826.0
81.00	2730.0
82.10	690.0
83.00	533.0
85.00	639.0
85.90	995.0
86.90	344.0
91.00	801.0
92.00	850.0
93.00	6585.0
94.00	564.0
98.00	5315.0
99.00	3884.0
101.00	1502.0
103.00	479.0
104.00	947.0
105.00	1120.0
107.00	13313.0
108.00	1720.0
109.00	401.0
110.00	17568.0
111.00	3261.0
111.90	372.0
116.00	823.0
117.00	19472.0
118.00	1319.0
121.90	1126.0
122.90	1646.0
123.90	706.0
124.90	558.0
127.00	37416.0
128.00	2985.0
129.00	17168.0
130.00	1499.0
133.90	569.0
135.00	1486.0
135.90	784.0
137.00	1126.0
139.90	320.0
141.00	2884.0
142.00	731.0
143.00	617.0
146.00	602.0
147.00	1548.0
148.00	4656.0
149.00	632.0
151.10	414.0
152.90	932.0
154.00	495.0
155.00	1646.0
156.00	2328.0
158.00	683.0
158.90	359.0
159.90	869.0
160.90	1246.0
161.90	376.0
165.00	1265.0
165.90	730.0
167.00	7494.0
168.00	4366.0
169.00	604.0
171.90	552.0
173.00	909.0
174.00	1122.0
175.00	2350.0
176.00	362.0
176.90	985.0
178.00	544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

178.90	4583.0
180.00	2893.0
180.90	1462.0
182.00	308.0
183.90	395.0
185.00	2167.0
186.00	15579.0
187.00	4167.0
188.00	359.0
188.90	1291.0
190.90	789.0
192.00	1502.0
192.90	1881.0
195.90	2107.0
196.90	518.0
197.90	72688.0
198.90	4362.0
199.90	423.0
202.90	1113.0
204.00	4210.0
205.00	7163.0
206.00	25376.0
207.00	4135.0
207.90	1408.0
208.90	769.0
210.00	684.0
211.00	1250.0
214.90	322.0
216.00	744.0
216.90	9865.0
217.90	973.0
221.00	2328.0
222.00	1127.0
222.90	2287.0
224.00	16504.0
225.00	4447.0
226.00	609.0
227.00	8942.0
227.90	1626.0
228.90	1654.0
231.00	507.0
233.90	656.0
235.00	548.0
235.90	454.0
237.00	962.0
238.90	454.0
239.90	309.0
240.90	461.0
242.00	1019.0
243.00	688.0
244.00	10489.0
245.00	1540.0
246.00	3250.0
247.00	700.0
248.90	344.0
252.80	505.0
254.10	518.0
255.00	61800.0
256.00	8543.0
257.00	824.0
258.00	5968.0
258.90	1028.0
265.00	2215.0
272.00	312.0
273.00	1850.0
274.00	5280.0
275.00	22720.0
276.00	3109.0
277.00	3365.0
277.90	809.0
281.10	389.0
282.90	432.0
285.00	751.0
292.90	512.0
296.00	12027.0
297.00	1570.0
303.00	1261.0
313.90	619.0
315.00	1525.0
316.00	553.0
320.80	335.0
323.00	2866.0
324.00	406.0
327.00	816.0
333.00	301.0
334.00	2018.0
335.00	472.0
340.90	341.0
345.90	865.0
352.00	846.0
353.00	554.0
354.00	661.0
365.00	5974.0
365.90	733.0
372.00	1434.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

383.00	443.0
401.90	672.0
403.00	874.0
421.00	603.0
421.90	443.0
423.00	3937.0
424.00	1144.0
425.00	312.0
441.10	10883.0
442.00	67664.0
443.10	13124.0
444.10	808.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

m/z	Abundance
37.00	360.0
38.00	916.0
39.10	4813.0
40.00	366.0
43.90	612.0
49.00	802.0
50.10	19528.0
51.00	66432.0
52.00	3363.0
56.00	1868.0
57.00	3839.0
61.00	911.0
62.00	794.0
63.00	2411.0
64.00	415.0
64.90	941.0
68.10	1002.0
69.00	57104.0
73.00	1130.0
74.00	8392.0
75.00	11512.0
76.00	3935.0
77.00	71512.0
78.10	4675.0
79.00	6166.0
79.90	4235.0
81.00	5245.0
82.00	1251.0
83.00	934.0
84.90	1014.0
85.90	1765.0
86.90	656.0
88.00	316.0
91.00	1270.0
92.00	1379.0
93.00	12501.0
94.00	500.0
96.00	435.0
98.00	9828.0
99.00	6948.0
99.90	434.0
101.00	2949.0
103.00	887.0
104.00	2011.0
105.00	1639.0
106.00	697.0
107.00	23712.0
108.00	3666.0
108.90	528.0
110.00	32792.0
111.00	5968.0
112.00	991.0
116.00	1664.0
117.00	36280.0
118.00	2739.0
122.00	2143.0
123.00	2505.0
124.00	1324.0
124.90	1280.0
127.00	66848.0
128.00	5224.0
129.00	31400.0
130.00	2680.0
131.00	683.0
132.00	303.0
133.90	1042.0
135.00	2362.0
136.00	1088.0
137.00	1947.0
137.90	358.0
139.90	444.0
141.00	4779.0
141.90	1234.0
143.00	1207.0
144.00	362.0
145.90	1107.0
147.00	2592.0
148.00	8365.0
148.90	1732.0
149.90	534.0
151.10	673.0
153.00	1509.0
153.90	1134.0
155.00	2612.0
156.00	3804.0
156.90	536.0
157.90	917.0
158.90	862.0
160.00	1322.0
161.00	2231.0
162.00	536.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

164.90	1812.0
165.90	1441.0
167.00	13222.0
168.00	7429.0
169.00	789.0
169.90	364.0
171.90	890.0
173.00	1148.0
174.00	1958.0
175.00	3854.0
175.90	578.0
176.90	1981.0
178.00	703.0
179.00	7583.0
180.00	5584.0
181.00	2121.0
182.00	478.0
184.00	598.0
185.00	3556.0
186.00	23648.0
187.00	7036.0
188.00	748.0
189.00	1905.0
189.90	389.0
190.90	818.0
192.00	2290.0
193.00	3099.0
194.00	630.0
195.00	354.0
195.90	4046.0
197.00	643.0
197.90	118800.0
199.00	8191.0
199.90	808.0
201.50	478.0
202.90	1606.0
204.00	7193.0
205.00	11821.0
206.00	40896.0
207.00	6545.0
208.00	2240.0
209.00	800.0
210.10	823.0
210.90	2162.0
215.00	695.0
215.90	1433.0
216.90	14650.0
217.90	1847.0
220.90	3134.0
221.90	1865.0
222.90	3437.0
224.00	26016.0
225.00	6425.0
226.00	1028.0
227.00	14669.0
227.90	1803.0
229.00	2238.0
230.10	315.0
231.00	716.0
233.90	799.0
234.90	937.0
235.90	772.0
236.90	1375.0
238.90	595.0
240.00	530.0
240.90	746.0
242.00	1681.0
243.00	1447.0
244.00	17512.0
245.00	2257.0
246.00	5405.0
246.90	844.0
249.00	615.0
251.90	309.0
253.00	654.0
254.00	632.0
255.00	90688.0
256.00	13321.0
257.00	1005.0
257.90	8138.0
259.00	1385.0
264.90	3123.0
265.90	381.0
273.00	2726.0
274.00	7225.0
275.00	35024.0
276.00	4249.0
277.00	5314.0
278.00	1030.0
281.00	420.0
282.00	319.0
283.00	573.0
285.00	642.0
292.90	1122.0
293.90	341.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

296.00	17176.0
297.00	2407.0
303.00	1552.0
304.00	428.0
314.10	483.0
314.90	2337.0
316.00	723.0
320.90	315.0
323.00	3554.0
324.00	652.0
326.90	1000.0
327.90	366.0
332.00	386.0
333.00	445.0
334.00	2307.0
335.00	701.0
341.10	538.0
346.00	841.0
352.00	1299.0
352.90	682.0
354.00	1002.0
355.10	454.0
365.00	7794.0
366.00	1126.0
371.00	377.0
372.00	1696.0
372.90	493.0
382.90	455.0
401.90	518.0
403.00	925.0
404.00	440.0
420.90	860.0
422.00	754.0
423.00	4912.0
424.00	1472.0
441.00	14326.0
442.00	84208.0
443.10	16544.0
444.00	1249.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

m/z	Abundance
38.00	1028.0
39.00	5108.0
40.00	348.0
43.90	356.0
44.90	309.0
48.90	746.0
50.00	21144.0
51.00	66000.0
52.00	3226.0
54.90	448.0
56.00	1675.0
57.00	3592.0
61.00	784.0
62.00	930.0
63.00	2409.0
63.90	342.0
65.00	1055.0
68.00	1230.0
69.00	53872.0
70.00	327.0
73.10	1237.0
74.00	8537.0
75.00	9955.0
76.10	3998.0
77.00	68592.0
78.10	4888.0
79.00	5513.0
80.00	3926.0
81.00	5087.0
82.00	1170.0
83.00	1074.0
85.00	1108.0
86.00	1780.0
87.00	706.0
90.90	1169.0
92.00	1433.0
93.00	10369.0
94.00	667.0
96.00	347.0
98.00	9152.0
99.00	6480.0
99.90	434.0
100.90	2689.0
102.90	990.0
104.00	2313.0
105.00	2089.0
105.90	510.0
107.00	22088.0
108.00	2960.0
109.10	509.0
110.00	30400.0
111.00	5920.0
111.90	697.0
116.00	1254.0
117.00	32560.0
118.00	2001.0
119.90	301.0
122.00	1790.0
123.00	2404.0
123.90	1182.0
125.00	888.0
127.00	61016.0
128.00	4921.0
129.00	28640.0
129.90	2505.0
130.90	559.0
133.90	893.0
135.00	2706.0
136.00	1153.0
136.90	1635.0
137.90	372.0
139.90	469.0
141.00	4027.0
142.00	1493.0
143.00	1042.0
143.90	338.0
146.00	817.0
147.00	2463.0
148.00	7532.0
149.00	1602.0
151.00	525.0
151.90	359.0
152.90	1413.0
153.90	1045.0
154.90	2616.0
156.00	3350.0
156.90	510.0
157.90	691.0
158.90	782.0
159.90	1230.0
160.90	1895.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

162.00	362.0
165.00	1775.0
166.00	1210.0
167.00	11282.0
167.90	5901.0
169.00	952.0
169.90	315.0
170.80	363.0
171.90	794.0
173.00	1073.0
173.90	1537.0
175.00	2981.0
175.90	481.0
177.00	1389.0
177.90	639.0
178.90	6696.0
180.00	4480.0
181.00	2138.0
182.10	315.0
183.90	666.0
185.00	3009.0
186.00	19472.0
187.00	6164.0
187.90	951.0
189.00	1789.0
189.90	300.0
190.90	799.0
192.00	2416.0
193.00	2790.0
194.00	560.0
196.00	3144.0
196.90	703.0
197.90	98528.0
198.90	7015.0
199.90	730.0
200.90	307.0
203.00	1278.0
204.00	5878.0
205.00	9044.0
206.00	36200.0
207.00	5690.0
208.00	1664.0
208.90	733.0
210.10	667.0
211.00	1618.0
215.00	620.0
216.00	1282.0
216.90	11881.0
217.90	1703.0
220.90	2617.0
221.90	1433.0
222.90	2258.0
224.00	21312.0
225.00	5768.0
226.00	774.0
227.00	11910.0
228.00	1747.0
229.00	2406.0
229.90	362.0
231.00	533.0
233.90	820.0
234.90	607.0
235.90	514.0
237.00	1449.0
238.90	475.0
239.80	377.0
241.00	507.0
242.00	1366.0
243.00	1279.0
244.00	13163.0
245.00	2318.0
245.90	3995.0
246.90	869.0
249.00	320.0
252.90	340.0
254.10	603.0
255.00	74624.0
256.00	11019.0
257.00	1088.0
258.00	6799.0
259.10	1263.0
264.90	2963.0
272.90	2209.0
274.00	5738.0
275.00	26904.0
276.00	3545.0
276.90	3913.0
278.00	864.0
281.00	369.0
283.00	517.0
284.00	304.0
285.00	474.0
293.00	557.0
296.00	12381.0
296.90	1891.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28

303.00	1058.0
304.10	323.0
314.00	526.0
315.00	1557.0
315.90	539.0
323.00	2543.0
324.00	473.0
326.90	756.0
332.90	350.0
334.00	2014.0
335.00	399.0
341.10	416.0
345.90	681.0
351.90	822.0
353.00	530.0
354.00	740.0
365.00	5657.0
365.90	628.0
372.00	1548.0
373.00	338.0
382.90	303.0
402.00	412.0
402.90	694.0
420.90	526.0
421.90	400.0
423.00	3651.0
424.00	981.0
441.10	8862.0
442.00	53432.0
443.00	10726.0
444.00	1026.0

Instrument :
BNA_M
ClientSampleId :
DFTPP28