

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071321\
 Data File : BM030970.D
 Acq On : 14 Jul 2021 21:14
 Operator : CG/JU
 Sample : M2910-05
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AW4

Manual Integrations
 APPROVED

mohammad
 7/15/2021 11:31:22 AM

Quant Time: Jul 15 04:07:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 13 14:01:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	1166	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.428	136	4021	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.292	164	2976	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.037	188	5941m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.222	240	5416	0.400	ng/ul	-0.01
23) Perylene-d12	23.411	264	5480	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	5976	4.984	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.020	152	2742	0.392	ng/ul	-0.01
18) Fluoranthene-d10	19.066	212	8802	0.483	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.269	88	295	0.221	ng/ul#	59
5) Naphthalene	10.478	128	3633	0.303	ng/ul	98
8) 1-Methylnaphthalene	12.317	142	1078	0.130	ng/ul	98
11) Acenaphthene	14.352	153	961	0.091	ng/ul	98
12) Fluorene	15.346	166	1596	0.128	ng/ul#	96
14) Pentachlorophenol	16.690	266	220543	105.088	ng/ul	99
15) Phenanthrene	17.079	178	1127	0.060	ng/ul#	70
16) Anthracene	17.169	178	590	0.033	ng/ul#	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071321\
Data File : BM030970.D
Acq On : 14 Jul 2021 21:14
Operator : CG/JU
Sample : M2910-05
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AW4

Manual Integrations
APPROVED

mohammad
7/15/2021 11:31:22 AM

Quant Time: Jul 15 04:07:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 13 14:01:01 2021
Response via : Initial Calibration

