

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071423\
 Data File : BM040907.D
 Acq On : 15 Jul 2023 16:35
 Operator : MA/JU
 Sample : 03414-03
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4646

Quant Time: Jul 16 00:29:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 15 03:58:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	4056	0.400	ng/ul	0.00
4) Naphthalene-d8	10.658	136	13215	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.504	164	5267	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.261	188	8455	0.400	ng/ul	0.00
17) Chrysene-d12	21.451	240	6009	0.400	ng/ul	0.00
23) Perylene-d12	23.824	264	6313	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.280	96	16655	2.614	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.253	152	2184	0.118	ng/ul	0.00
18) Fluoranthene-d10	19.287	212	2675	0.139	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.708	128	5197	0.143	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071423\
Data File : BM040907.D
Acq On : 15 Jul 2023 16:35
Operator : MA/JU
Sample : 03414-03
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4646

Quant Time: Jul 16 00:29:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 15 03:58:02 2023
Response via : Initial Calibration

