

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM071519\
 Data File : BM021456.D
 Acq On : 15 Jul 2019 12:21
 Operator : HP/JU
 Sample : K3762-10
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF4B6

Quant Time: Jul 15 13:19:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM070819MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 12 05:46:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	163100	20.00	ng/ul	0.00
18) Naphthalene-d8	10.51	136	632646	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.37	164	365066	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.12	188	844597	20.00	ng/ul	0.00
77) Chrysene-d12	21.31	240	806003	20.00	ng/ul	0.00
85) Perylene-d12	23.57	264	922707	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.26	96	21633	6.53	ng/uL	0.00
5) Phenol-d5	6.90	99	103644	8.17	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.07	67	256384	34.10	ng/ul	0.00
9) 2-Chlorophenol-d4	7.26	132	308585	28.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.43	113	193549	19.16	ng/ul	0.00
19) Nitrobenzene-d5	8.89	128	188300	36.13	ng/ul	0.00
22) 2-Nitrophenol-d4	9.61	143	207350	35.19	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.14	165	385038	32.76	ng/ul	0.00
29) 4-Chloroaniline-d4	10.69	131	18986	1.79	ng/ul	0.02
43) Dimethylphthalate-d6	13.79	166	1289888	40.34	ng/ul	0.00
46) Acenaphthylene-d8	14.06	160	1444874	35.99	ng/ul	0.00
51) 4-Nitrophenol-d4	14.60	143	24524	4.84	ng/ul	0.00
57) Fluorene-d10	15.37	176	1120232	39.25	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.50	200	160502	28.27	ng/ul	0.00
70) Anthracene-d10	17.22	188	1769721	40.34	ng/ul	0.00
78) Pyrene-d10	19.52	212	1991967	45.03	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.43	264	2187463	41.22	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM071519\
Data File : BM021456.D
Acq On : 15 Jul 2019 12:21
Operator : HP/JU
Sample : K3762-10
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BF4B6

Quant Time: Jul 15 13:19:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM070819MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 12 05:46:23 2019
Response via : Initial Calibration

