


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071619\
Data File : BM021460.D
Acq On : 16 Jul 2019 08:15
Operator : HP/JU
Sample : DFTPP26
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP26

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM070819MA.M
Title : SVOA CALIBRATION
Last Update : Fri Jul 12 05:46:23 2019



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.9	108154	PASS
68	69	0.00	2	1.9	2393	PASS
69	198	0.00	100	33.9	126749	PASS
70	69	0.00	2	0.5	601	PASS
127	198	10	80	42.9	160725	PASS
197	198	0.00	2	0.6	2269	PASS
198	198	100	100	100.0	374442	PASS
199	198	5	9	6.7	24933	PASS
275	198	10	60	30.4	114000	PASS
365	198	1	100	4.0	15104	PASS
441	198	0.01	100	13.6	51040	PASS
442	198	50	100	87.6	328106	PASS
443	442	15	24	19.5	63917	PASS

Scan 2470 (17.515 min): BM021460.D
DFTPP26

m/z	Abundance
40.00	1051.0
44.00	4915.0
48.90	372.0
73.10	428.0
191.00	316.0
192.90	325.0
207.00	2587.0
207.90	496.0
209.10	443.0
281.00	642.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

m/z	Abundance
37.00	476.0
38.10	1219.0
39.10	7655.0
40.10	1163.0
41.00	356.0
44.00	5143.0
44.90	304.0
49.10	984.0
50.00	25608.0
51.10	96728.0
52.10	5332.0
55.10	532.0
56.00	3037.0
57.00	7157.0
58.00	340.0
61.00	1587.0
62.00	1549.0
63.10	4303.0
64.00	590.0
65.00	2088.0
68.00	2050.0
69.00	113288.0
70.00	438.0
73.00	1208.0
74.10	12277.0
75.00	18128.0
76.10	6351.0
77.10	123184.0
78.10	9042.0
79.00	8626.0
80.00	6630.0
81.00	9568.0
82.00	2075.0
83.00	1985.0
84.10	449.0
85.00	1592.0
86.00	2943.0
86.90	1263.0
88.00	508.0
91.00	2258.0
92.00	2179.0
93.00	15303.0
93.90	1003.0
96.00	708.0
98.00	11277.0
99.00	9889.0
100.00	1039.0
101.00	5218.0
103.00	1652.0
104.00	3433.0
105.00	2988.0
106.10	945.0
107.00	39288.0
108.00	5455.0
109.00	1319.0
110.00	67520.0
111.00	10318.0
112.00	1307.0
112.90	332.0
113.10	331.0
116.00	1900.0
117.00	33384.0
118.00	2714.0
118.90	397.0
119.80	438.0
122.00	2535.0
123.00	4215.0
124.00	2151.0
125.00	1809.0
126.20	826.0
127.00	135040.0
128.00	11155.0
129.00	51960.0
130.00	4911.0
131.00	992.0
131.90	429.0
133.00	330.0
134.00	1767.0
135.00	4552.0
136.00	1788.0
137.00	2990.0
137.90	618.0
139.90	695.0
141.00	7234.0
141.90	2262.0
143.00	1651.0
143.90	474.0
145.00	388.0
146.00	1593.0
147.00	4261.0
148.00	8638.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

149.00	2203.0
149.90	405.0
151.20	817.0
151.90	449.0
152.90	2928.0
154.00	1918.0
155.00	4148.0
156.00	6426.0
157.00	1154.0
157.90	1094.0
159.00	1269.0
160.00	2712.0
161.00	3270.0
162.00	1053.0
162.90	316.0
164.00	449.0
165.00	3243.0
166.00	2251.0
167.00	14123.0
168.00	11384.0
169.00	1436.0
170.00	460.0
171.00	627.0
172.00	1295.0
173.00	1691.0
174.00	3206.0
175.00	6335.0
176.00	1650.0
177.00	2931.0
177.90	1064.0
179.00	11374.0
180.00	8934.0
181.00	3782.0
182.00	681.0
184.00	971.0
185.00	5584.0
186.00	41080.0
187.00	12999.0
188.00	1348.0
189.00	2721.0
189.90	389.0
190.90	1607.0
192.00	3984.0
193.00	4366.0
194.00	815.0
194.90	516.0
196.00	8124.0
197.10	1963.0
197.90	288000.0
198.90	18880.0
200.00	1516.0
201.40	1417.0
202.90	2235.0
204.00	10173.0
205.00	18712.0
206.00	77136.0
207.00	13099.0
208.00	3473.0
209.00	1263.0
210.00	1368.0
211.00	3097.0
211.70	1129.0
215.00	956.0
216.00	1697.0
217.00	21960.0
218.00	2888.0
221.00	12170.0
221.90	2691.0
223.00	4615.0
224.00	41976.0
225.00	11231.0
226.00	1467.0
227.00	18184.0
228.00	2798.0
229.00	4084.0
229.90	412.0
231.00	1539.0
232.00	331.0
233.90	1051.0
234.90	1479.0
235.90	876.0
237.00	2002.0
238.90	732.0
240.00	625.0
240.90	1074.0
242.00	2340.0
243.00	2697.0
244.00	32848.0
245.00	4828.0
246.00	6980.0
247.00	1504.0
247.90	397.0
248.90	1441.0
252.00	333.0
252.90	783.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

254.20	1320.0
255.00	177728.0
256.00	26632.0
257.00	2337.0
258.00	9960.0
259.00	1892.0
259.90	385.0
260.90	346.0
265.00	4438.0
266.00	837.0
271.00	687.0
272.00	533.0
273.00	5337.0
274.00	13273.0
275.00	77808.0
276.00	11808.0
277.00	7031.0
278.00	1300.0
281.00	713.0
281.90	434.0
283.00	881.0
283.90	404.0
285.00	1511.0
285.90	363.0
290.80	363.0
292.00	373.0
293.00	1678.0
293.90	644.0
295.00	549.0
296.00	21944.0
297.00	3012.0
302.00	583.0
303.00	2767.0
304.00	635.0
307.90	334.0
314.00	1103.0
315.00	2488.0
316.00	1400.0
317.00	347.0
321.10	684.0
322.10	392.0
323.00	6698.0
324.00	1269.0
327.00	1210.0
328.00	635.0
331.90	474.0
333.00	833.0
334.00	4736.0
335.00	1474.0
341.00	947.0
345.90	1236.0
347.00	333.0
352.00	2187.0
353.00	1476.0
354.10	2053.0
355.00	456.0
365.00	9665.0
366.00	1725.0
371.10	540.0
372.00	3722.0
373.00	957.0
383.00	913.0
390.00	430.0
391.00	329.0
392.10	303.0
402.00	1366.0
403.10	1606.0
404.00	835.0
421.00	1643.0
422.00	1677.0
423.10	11285.0
424.00	3416.0
425.10	413.0
441.00	27816.0
442.10	179968.0
443.10	36048.0
444.00	3469.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

m/z	Abundance
37.00	595.0
38.10	1841.0
39.10	10204.0
40.00	1463.0
44.00	5284.0
44.90	423.0
49.10	1303.0
50.10	35384.0
51.10	132608.0
52.10	6837.0
53.00	441.0
55.00	858.0
56.00	4364.0
57.00	8937.0
58.00	510.0
61.00	1833.0
62.10	1915.0
63.10	6038.0
63.90	714.0
65.10	3168.0
68.10	3246.0
69.00	155712.0
70.00	766.0
73.10	1609.0
74.00	15684.0
75.00	25232.0
76.10	8105.0
77.10	171264.0
78.10	11025.0
79.00	11988.0
80.00	8665.0
81.00	12356.0
82.00	2913.0
83.00	2653.0
84.00	408.0
85.00	2480.0
86.10	4217.0
87.00	1413.0
88.00	692.0
89.00	446.0
91.10	2556.0
92.00	3263.0
93.00	21328.0
94.00	1410.0
95.00	378.0
96.00	859.0
97.10	412.0
98.00	15474.0
99.00	13516.0
100.00	1193.0
101.00	7200.0
102.00	351.0
103.00	2298.0
104.00	4959.0
105.00	4235.0
106.10	1260.0
107.00	56224.0
108.00	8366.0
109.10	1697.0
110.00	96584.0
111.00	15581.0
112.00	1758.0
113.00	599.0
115.00	398.0
116.00	3500.0
117.00	47192.0
118.00	3579.0
119.00	422.0
120.10	725.0
122.00	4266.0
123.00	6313.0
124.00	2322.0
125.00	2289.0
126.20	1113.0
127.00	196288.0
128.00	14460.0
129.00	74624.0
130.00	6958.0
131.00	1124.0
132.10	493.0
132.80	351.0
134.00	2640.0
135.00	6733.0
136.10	2709.0
137.00	3570.0
137.90	685.0
139.00	401.0
140.00	1203.0
141.00	10367.0
142.00	3272.0
143.00	2585.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

144.00	543.0
145.10	641.0
146.00	1993.0
147.00	6181.0
148.00	13777.0
149.00	3015.0
150.00	734.0
151.00	1174.0
151.80	1063.0
153.00	3698.0
154.00	2297.0
155.00	5911.0
156.00	9680.0
157.00	1851.0
158.00	2149.0
159.00	1649.0
160.00	3302.0
161.00	5010.0
161.90	1531.0
163.00	458.0
164.00	372.0
165.00	4330.0
166.00	3514.0
167.00	23264.0
168.00	15115.0
169.00	2435.0
169.90	939.0
170.90	1092.0
172.00	2345.0
173.00	2820.0
174.00	4765.0
175.00	9495.0
176.00	2444.0
177.00	4982.0
178.00	1469.0
179.00	18400.0
180.00	12439.0
181.00	5384.0
182.00	966.0
182.90	562.0
184.00	1510.0
185.00	8996.0
186.00	66088.0
187.00	20088.0
188.10	1724.0
189.00	4413.0
190.00	807.0
191.00	2302.0
192.00	6205.0
193.00	6495.0
194.10	1486.0
195.00	673.0
196.00	12833.0
196.80	2590.0
197.10	2599.0
197.90	462720.0
198.90	30776.0
199.90	2781.0
201.50	2193.0
203.00	3762.0
204.00	18224.0
205.00	30696.0
206.00	130224.0
207.00	19624.0
208.00	5195.0
209.00	1812.0
210.30	2262.0
211.10	5260.0
212.00	839.0
213.00	475.0
214.90	1322.0
216.00	3031.0
217.00	34456.0
218.00	4355.0
219.00	409.0
221.00	21576.0
221.90	4532.0
223.00	7456.0
224.00	70408.0
225.00	17792.0
226.10	2290.0
227.00	29720.0
228.00	4553.0
229.00	5865.0
229.90	684.0
231.00	2888.0
231.90	443.0
233.00	502.0
234.00	1886.0
235.00	2391.0
235.90	1504.0
237.00	2911.0
237.90	491.0
239.00	1298.0
239.90	1243.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

241.00	1748.0
242.00	3736.0
243.00	4277.0
244.00	55152.0
245.00	7346.0
246.00	11464.0
247.00	2184.0
248.00	643.0
249.00	2193.0
249.90	412.0
250.90	569.0
252.00	472.0
253.00	1312.0
254.10	2410.0
255.00	310528.0
256.00	44016.0
257.00	3564.0
258.00	17008.0
259.00	3056.0
260.00	500.0
260.90	602.0
264.00	385.0
265.00	7122.0
266.00	1239.0
270.00	346.0
271.00	1287.0
271.90	861.0
273.00	9106.0
274.00	22776.0
275.00	139008.0
276.00	18912.0
277.00	11609.0
278.00	1982.0
279.00	611.0
281.10	744.0
282.00	614.0
283.00	1624.0
283.90	1161.0
285.00	2131.0
286.00	370.0
288.90	579.0
289.90	389.0
290.90	332.0
292.20	527.0
293.00	3182.0
293.90	1041.0
295.00	701.0
296.00	41680.0
297.00	5188.0
297.90	355.0
301.00	541.0
302.00	669.0
303.00	4750.0
304.00	1305.0
307.90	634.0
309.10	369.0
309.90	635.0
312.90	426.0
314.00	1946.0
315.00	4415.0
316.00	2603.0
317.00	441.0
321.00	1468.0
322.10	778.0
323.00	12671.0
324.00	2177.0
327.00	2752.0
328.00	1129.0
329.00	308.0
332.00	970.0
333.00	1258.0
334.00	8102.0
335.00	1960.0
341.00	1641.0
341.90	448.0
346.00	2834.0
347.00	536.0
352.00	3821.0
353.10	2719.0
354.00	3790.0
354.90	776.0
358.90	365.0
365.00	17488.0
366.00	2777.0
371.00	970.0
372.00	6403.0
373.00	1776.0
383.00	2046.0
384.10	435.0
390.10	957.0
391.00	649.0
392.00	516.0
400.90	538.0
402.00	2838.0
403.00	3596.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

404.10	1347.0
421.00	3148.0
422.00	3234.0
423.10	21904.0
424.00	6135.0
425.00	716.0
441.10	59000.0
442.10	379776.0
443.10	74760.0
444.00	6490.0
445.00	348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

m/z	Abundance
37.10	420.0
38.00	1184.0
39.10	6837.0
40.00	1140.0
41.00	311.0
44.00	5521.0
49.10	1076.0
50.10	24560.0
51.10	95128.0
52.00	4546.0
55.00	579.0
56.00	2881.0
57.00	6957.0
57.90	347.0
61.00	1402.0
62.00	1390.0
63.10	4019.0
64.10	552.0
65.10	1806.0
68.10	1885.0
69.00	111248.0
70.00	600.0
73.00	1566.0
74.10	10927.0
75.00	18824.0
76.10	6217.0
77.10	123768.0
78.00	8565.0
79.00	9071.0
80.00	6381.0
81.00	9128.0
82.00	2126.0
83.00	2380.0
83.90	403.0
85.00	1881.0
86.00	3017.0
87.00	1255.0
88.00	451.0
91.00	2094.0
92.00	2010.0
93.00	16109.0
93.90	865.0
95.90	736.0
97.10	460.0
98.00	11231.0
99.00	9854.0
100.00	921.0
101.00	4968.0
102.00	509.0
103.00	1809.0
104.00	3499.0
105.00	3078.0
106.00	1136.0
107.00	42280.0
108.00	6183.0
109.00	1216.0
110.00	73648.0
111.00	11054.0
112.00	1624.0
113.00	457.0
116.00	2431.0
117.00	35808.0
118.00	2630.0
118.90	408.0
119.90	506.0
122.00	2886.0
122.90	4068.0
124.00	2258.0
125.10	1677.0
126.20	946.0
127.00	150848.0
128.00	10885.0
129.00	60608.0
130.00	4531.0
131.00	978.0
132.10	506.0
132.90	415.0
134.00	1858.0
135.00	4501.0
136.10	1834.0
137.00	2217.0
137.90	671.0
139.10	301.0
139.90	1021.0
141.00	8959.0
142.00	2655.0
143.00	1890.0
144.00	611.0
145.10	336.0
146.00	1827.0
147.00	4403.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

148.00	10011.0
149.00	2327.0
149.90	612.0
151.10	1075.0
151.90	736.0
153.00	3247.0
154.00	1926.0
155.00	5210.0
156.00	7286.0
157.10	1661.0
157.90	1380.0
159.00	1295.0
159.90	3050.0
161.00	4071.0
162.00	1263.0
162.90	362.0
164.00	797.0
165.00	3279.0
166.00	2708.0
167.00	18456.0
168.00	11025.0
169.00	1630.0
170.00	666.0
170.90	684.0
172.00	1770.0
173.00	2405.0
174.00	4029.0
175.00	7692.0
175.90	1759.0
177.00	3715.0
178.00	1329.0
179.00	14204.0
180.00	9643.0
181.00	4257.0
182.00	679.0
183.00	525.0
184.00	1063.0
185.00	7039.0
186.00	54352.0
187.00	14940.0
188.00	1600.0
188.90	3188.0
190.00	608.0
191.00	2033.0
192.00	4230.0
193.00	5862.0
194.00	1286.0
195.00	663.0
196.00	11003.0
197.10	2247.0
197.90	372608.0
199.00	25144.0
199.90	2021.0
201.50	2488.0
203.00	3472.0
204.00	14891.0
205.00	24704.0
206.00	104376.0
207.00	15848.0
208.00	4116.0
209.00	1544.0
209.90	1356.0
211.00	3836.0
213.10	337.0
215.00	1486.0
216.00	2178.0
217.00	27608.0
218.00	3519.0
221.00	17088.0
221.80	3933.0
223.00	6584.0
224.00	59272.0
225.00	16025.0
226.10	2077.0
227.00	26336.0
228.00	3671.0
229.00	5524.0
229.90	732.0
231.00	2250.0
232.00	420.0
233.00	444.0
234.00	1707.0
235.00	1946.0
236.00	1608.0
237.00	2493.0
237.90	350.0
239.00	1114.0
240.10	743.0
241.00	1308.0
242.00	3365.0
243.00	3511.0
244.00	49152.0
245.00	6957.0
246.00	9754.0
247.00	2055.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26

247.90	492.0
249.00	1666.0
250.90	603.0
252.00	424.0
252.90	1228.0
254.20	2147.0
255.00	265792.0
256.00	39136.0
257.00	3366.0
258.00	14556.0
259.00	2991.0
260.00	328.0
260.80	452.0
264.00	521.0
265.00	6086.0
265.90	710.0
269.90	320.0
271.10	1021.0
272.10	767.0
273.00	8395.0
274.00	21656.0
275.00	125184.0
276.00	16848.0
277.00	10086.0
278.00	1790.0
279.00	366.0
281.00	863.0
282.00	424.0
283.10	1284.0
284.10	708.0
285.00	1566.0
285.90	361.0
289.00	431.0
291.10	307.0
291.90	709.0
292.90	2130.0
293.90	699.0
295.00	691.0
296.00	36816.0
297.00	4836.0
297.90	374.0
301.00	593.0
302.00	602.0
303.00	4387.0
304.00	1242.0
308.00	416.0
309.00	475.0
309.90	590.0
314.00	2014.0
315.00	4577.0
316.10	2235.0
317.00	469.0
321.00	1339.0
322.00	811.0
323.00	12293.0
324.00	2029.0
326.90	2383.0
328.00	1164.0
329.00	369.0
332.00	779.0
333.00	925.0
334.00	7949.0
335.00	2359.0
336.00	322.0
340.90	1400.0
342.00	407.0
346.00	3108.0
347.00	564.0
352.10	3593.0
353.00	2680.0
354.00	4285.0
355.10	707.0
359.00	362.0
365.00	18160.0
366.00	2705.0
370.10	341.0
371.00	1121.0
372.10	7545.0
373.10	1601.0
374.10	322.0
377.00	311.0
383.00	1878.0
384.00	448.0
390.00	959.0
391.00	782.0
391.90	366.0
401.10	456.0
402.10	2899.0
403.10	3639.0
404.00	1459.0
421.00	3432.0
422.00	3050.0
423.10	23984.0
424.10	6778.0
425.00	615.0

Instrument :

BNA_M

ClientSampleId :

DFTPP26

438.00	313.0
440.20	575.0
441.10	66304.0
442.10	424576.0
443.10	80944.0
444.10	7058.0
445.00	442.0

Instrument :
BNA_M
ClientSampleId :
DFTPP26