


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071718\
Data File : BM015971.D
Acq On : 17 Jul 2018 16:38
Operator : SJ/JU
Sample : DFTPP30
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP30

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM071318MA.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 18 01:01:35 2018



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2495

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.3	69498	PASS
68	69	0.00	2	2.0	1213	PASS
69	198	0.00	100	50.2	59776	PASS
70	69	0.00	2	0.6	353	PASS
127	198	10	80	57.6	68640	PASS
197	198	0.00	2	0.7	863	PASS
198	198	100	100	100.0	119157	PASS
199	198	5	9	7.1	8438	PASS
275	198	10	60	30.2	35962	PASS
365	198	1	100	7.0	8337	PASS
441	443	0.01	100	82.9	13525	PASS
442	198	50	100	71.7	85464	PASS
443	442	15	24	19.1	16324	PASS

m/z	Abundance
37.00	634.0
38.00	1306.0
39.00	7279.0
40.00	601.0
43.90	382.0
49.10	858.0
50.00	25648.0
51.00	85712.0
52.00	4331.0
54.90	632.0
56.00	2198.0
57.00	4509.0
61.00	1077.0
62.00	1255.0
63.00	3175.0
64.00	363.0
65.00	1178.0
68.00	1500.0
69.00	74848.0
69.90	526.0
73.00	1577.0
74.10	10628.0
75.00	14197.0
76.00	5154.0
77.10	91520.0
78.00	5956.0
79.00	6821.0
80.00	5061.0
81.00	7048.0
82.00	1505.0
83.00	1317.0
84.90	1220.0
86.00	2760.0
87.00	1089.0
87.90	331.0
91.00	1435.0
92.00	1680.0
93.00	14700.0
94.00	858.0
96.00	617.0
97.00	390.0
98.00	13029.0
99.00	8676.0
99.90	721.0
101.00	3489.0
103.00	1089.0
104.00	2542.0
105.00	2235.0
105.90	910.0
107.00	30072.0
108.00	4466.0
109.00	842.0
110.00	39264.0
111.00	7640.0
111.90	853.0
113.00	378.0
116.00	1826.0
117.00	43272.0
118.00	3208.0
119.00	312.0
120.10	310.0
121.90	2008.0
123.00	3174.0
124.00	1541.0
124.90	1372.0
127.00	84784.0
128.00	5981.0
129.00	38480.0
130.00	3567.0
130.90	585.0
133.90	1300.0
135.00	3340.0
135.90	1570.0
137.00	2091.0
138.00	339.0
139.90	573.0
141.00	5646.0
141.90	1935.0
143.00	1147.0
146.00	1153.0
147.00	3507.0
148.00	9660.0
149.00	1916.0
150.00	360.0
151.10	624.0
152.90	1608.0
154.00	1537.0
155.00	2727.0
156.00	4208.0
156.90	622.0
157.90	1016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

158.90	856.0
160.00	1856.0
160.90	3015.0
162.00	763.0
162.90	385.0
165.00	2347.0
166.00	1578.0
167.00	14951.0
168.00	8855.0
169.00	1110.0
169.90	498.0
171.00	596.0
171.90	1141.0
173.00	1295.0
174.00	2543.0
175.00	4631.0
176.10	711.0
176.90	2161.0
178.00	686.0
178.90	9411.0
180.00	6276.0
180.90	2564.0
182.10	314.0
184.00	616.0
185.00	4055.0
186.00	27376.0
187.00	8287.0
188.00	1025.0
189.00	2330.0
190.00	336.0
190.90	1148.0
192.00	3430.0
193.00	3286.0
193.90	547.0
195.00	312.0
195.90	4193.0
197.10	1261.0
197.90	130024.0
198.90	9471.0
199.90	867.0
201.70	444.0
203.00	1807.0
204.00	8510.0
205.00	14214.0
206.00	46144.0
207.00	6518.0
208.00	1889.0
208.90	969.0
209.90	960.0
211.00	2337.0
214.90	1144.0
216.00	1617.0
217.00	15895.0
218.00	2657.0
220.90	3511.0
222.00	1770.0
223.00	3413.0
224.00	28904.0
225.00	7285.0
225.90	1027.0
227.00	15995.0
228.00	2390.0
229.00	2590.0
230.00	499.0
230.90	891.0
233.90	948.0
234.90	1213.0
235.90	952.0
236.90	1571.0
238.90	661.0
240.00	775.0
241.00	758.0
242.00	1601.0
242.90	1381.0
244.00	18920.0
245.00	2648.0
246.00	5726.0
246.90	1254.0
248.90	627.0
252.90	600.0
254.10	767.0
255.00	98152.0
256.00	14748.0
256.90	1238.0
258.00	8414.0
259.00	1535.0
264.90	4310.0
266.00	498.0
273.00	3016.0
274.00	7757.0
275.00	36712.0
276.00	4869.0
277.00	4977.0
278.00	1196.0
281.00	353.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

283.00	728.0
284.00	469.0
285.00	936.0
293.00	1162.0
293.80	356.0
295.00	331.0
296.00	18936.0
297.00	2721.0
303.00	1801.0
313.90	711.0
315.00	2005.0
315.90	735.0
320.90	385.0
323.00	3405.0
324.00	635.0
326.90	1013.0
327.90	446.0
331.90	408.0
333.00	389.0
334.00	2538.0
334.90	547.0
341.00	569.0
345.90	1110.0
352.00	1103.0
353.00	856.0
354.00	1093.0
365.00	7898.0
365.90	988.0
370.90	428.0
372.00	1646.0
373.00	392.0
383.00	487.0
389.90	365.0
401.90	586.0
403.10	848.0
421.00	676.0
422.00	776.0
423.00	4317.0
424.00	1428.0
441.00	12061.0
442.00	71272.0
443.00	14313.0
444.00	1312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

m/z	Abundance
36.90	561.0
38.00	1212.0
39.10	6100.0
40.00	350.0
44.00	353.0
49.00	773.0
50.00	25936.0
51.00	81496.0
52.00	3971.0
54.90	378.0
56.00	2270.0
57.00	4204.0
61.00	1094.0
62.00	1380.0
63.00	3016.0
64.00	376.0
65.00	1163.0
68.00	1380.0
69.00	69648.0
70.00	533.0
73.00	1657.0
74.00	9878.0
75.00	12809.0
76.00	4988.0
77.00	85648.0
78.00	5872.0
79.00	6443.0
80.00	4634.0
81.00	6770.0
81.90	1482.0
83.00	1390.0
84.90	1276.0
86.00	2619.0
87.00	812.0
90.90	1601.0
91.90	1849.0
93.00	13682.0
94.00	905.0
96.10	328.0
98.00	11991.0
99.00	7634.0
100.00	569.0
101.00	3474.0
102.90	956.0
104.00	2740.0
104.90	2679.0
106.00	730.0
107.00	28800.0
108.00	3808.0
109.00	739.0
110.00	39904.0
111.00	6444.0
112.00	996.0
116.00	1698.0
117.00	43104.0
118.00	3012.0
121.90	2179.0
123.00	3618.0
123.90	1342.0
125.00	1281.0
127.00	79200.0
128.00	6420.0
129.00	38176.0
130.00	3666.0
130.90	692.0
132.00	318.0
133.90	1259.0
134.90	3356.0
136.00	1682.0
137.00	1618.0
138.00	356.0
140.00	604.0
141.00	6013.0
142.00	1440.0
142.90	1317.0
143.90	313.0
145.00	303.0
146.00	1255.0
147.00	3424.0
148.00	10349.0
149.00	2001.0
149.90	306.0
151.00	782.0
153.00	1581.0
154.00	1151.0
155.00	3767.0
156.00	4229.0
156.90	498.0
157.90	914.0
159.00	775.0
160.00	1914.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

160.90	2584.0
161.90	1025.0
162.90	313.0
165.00	2656.0
166.00	1640.0
167.00	15582.0
168.00	8697.0
169.00	1095.0
169.90	372.0
170.90	584.0
171.90	924.0
173.00	1769.0
174.00	2989.0
175.00	4828.0
176.00	948.0
176.90	2227.0
178.00	768.0
178.90	9590.0
179.90	6466.0
181.00	2638.0
182.00	336.0
183.90	492.0
185.00	4331.0
186.00	27504.0
187.00	8653.0
188.00	1098.0
189.00	2868.0
189.90	459.0
190.90	1014.0
192.00	3618.0
193.00	3711.0
194.00	791.0
196.00	4332.0
197.00	903.0
197.90	143360.0
198.90	9979.0
200.00	1373.0
201.40	509.0
201.90	309.0
202.90	2023.0
204.00	9258.0
205.00	14561.0
206.00	52096.0
207.00	7619.0
208.00	2745.0
208.90	1033.0
209.90	959.0
211.00	2619.0
212.00	314.0
214.90	786.0
216.00	1844.0
217.00	17776.0
217.90	2576.0
221.00	3786.0
221.90	2124.0
223.00	3246.0
224.00	30632.0
225.00	8503.0
226.00	1031.0
227.00	18504.0
228.00	2878.0
228.90	3479.0
229.90	467.0
230.90	1334.0
233.90	1431.0
235.00	1009.0
236.00	956.0
237.00	1491.0
238.90	786.0
240.00	668.0
241.00	1024.0
242.00	2083.0
243.00	1684.0
244.00	21272.0
245.00	2861.0
246.00	6745.0
246.90	1428.0
248.90	749.0
253.00	749.0
253.90	626.0
255.00	114584.0
256.00	16028.0
257.00	1612.0
258.00	9884.0
259.00	1438.0
260.00	306.0
265.00	4731.0
265.90	604.0
271.00	305.0
272.00	446.0
273.00	3819.0
274.00	10136.0
275.00	43824.0
276.00	5377.0
277.00	6525.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

277.90	1264.0
281.00	314.0
282.90	541.0
284.10	561.0
285.00	1036.0
293.00	1275.0
293.90	318.0
294.90	348.0
296.00	22288.0
297.00	3292.0
297.90	349.0
300.80	324.0
303.00	2204.0
303.90	555.0
313.90	763.0
315.00	2541.0
316.00	833.0
320.90	396.0
322.00	329.0
323.00	4982.0
324.00	920.0
326.90	1096.0
327.90	626.0
332.10	433.0
333.00	605.0
334.00	3403.0
335.00	927.0
341.00	743.0
345.90	1461.0
352.00	1553.0
353.00	796.0
353.90	1175.0
365.00	9710.0
365.90	1321.0
371.00	438.0
372.00	1890.0
373.00	609.0
383.00	643.0
401.90	883.0
403.00	1003.0
404.00	495.0
421.00	1175.0
422.00	1131.0
423.10	6064.0
423.90	1677.0
441.00	16760.0
442.00	104016.0
443.00	19840.0
444.00	2021.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

m/z	Abundance
37.90	554.0
39.00	2889.0
39.90	308.0
48.90	392.0
50.00	13226.0
51.00	41288.0
52.00	2191.0
54.90	319.0
56.00	1022.0
57.00	2485.0
60.90	589.0
61.90	594.0
63.00	1543.0
65.10	760.0
68.10	759.0
69.00	34832.0
73.00	1170.0
74.00	5307.0
75.00	6771.0
76.00	2246.0
77.00	46296.0
78.00	2716.0
79.00	3803.0
80.00	2729.0
81.00	3267.0
81.90	840.0
83.00	619.0
85.00	628.0
86.00	1198.0
87.00	419.0
91.00	646.0
92.00	832.0
93.00	8165.0
93.90	550.0
98.00	6529.0
99.00	3967.0
100.90	1577.0
102.90	668.0
104.00	1239.0
104.90	1274.0
107.00	15726.0
108.00	1970.0
109.00	349.0
110.00	20696.0
111.00	4476.0
112.00	719.0
116.00	899.0
117.00	24672.0
117.90	1851.0
122.00	1152.0
123.00	1934.0
123.90	781.0
124.90	573.0
126.20	378.0
127.00	41936.0
128.00	3152.0
129.00	21016.0
130.00	1602.0
131.00	300.0
133.90	714.0
135.00	2073.0
136.00	796.0
137.10	986.0
139.90	309.0
141.00	3503.0
141.90	974.0
143.00	912.0
146.00	741.0
147.00	1935.0
148.00	5804.0
148.90	1236.0
149.90	316.0
150.90	534.0
152.90	792.0
154.00	677.0
155.00	1829.0
156.00	2119.0
156.90	369.0
158.00	795.0
158.90	514.0
160.00	1109.0
160.90	1229.0
162.00	433.0
164.90	1798.0
165.90	858.0
167.00	9509.0
168.00	5292.0
168.90	903.0
170.90	366.0
172.00	548.0
173.00	839.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

174.00	1469.0
175.10	2780.0
175.90	483.0
176.90	1430.0
177.90	516.0
178.90	5247.0
180.00	3565.0
181.00	1425.0
183.90	489.0
185.00	2535.0
186.00	16816.0
187.00	5181.0
188.00	539.0
189.00	1252.0
191.00	832.0
191.90	1977.0
193.00	2376.0
193.90	416.0
195.90	2612.0
197.00	425.0
197.90	84088.0
199.00	5864.0
199.90	682.0
201.80	335.0
203.00	1331.0
204.00	5198.0
205.00	8315.0
206.00	28728.0
207.00	4466.0
207.90	1506.0
209.00	606.0
210.00	643.0
211.00	1258.0
214.90	491.0
216.00	1042.0
216.90	10372.0
217.90	1653.0
220.90	2329.0
221.90	1391.0
223.00	2269.0
224.00	19136.0
225.00	4702.0
225.90	604.0
227.00	11335.0
228.00	1445.0
229.00	1897.0
231.10	680.0
233.90	718.0
234.90	832.0
235.90	662.0
237.00	993.0
238.90	442.0
240.00	476.0
240.90	693.0
241.90	1337.0
243.00	1163.0
244.00	14849.0
245.00	1742.0
245.90	4612.0
246.90	853.0
249.00	434.0
252.90	479.0
254.10	550.0
255.00	71560.0
256.00	9782.0
257.00	897.0
258.00	6005.0
258.90	1161.0
265.00	3223.0
265.90	326.0
273.00	2165.0
274.00	5741.0
275.00	27352.0
276.00	3316.0
277.00	4113.0
277.90	769.0
282.90	494.0
284.00	466.0
285.00	635.0
292.90	615.0
296.00	15203.0
297.00	1928.0
303.00	1233.0
304.00	397.0
313.90	594.0
315.00	1604.0
316.00	886.0
321.00	371.0
323.00	3357.0
324.00	493.0
327.00	1016.0
327.90	459.0
333.00	318.0
334.00	2297.0
334.90	604.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

341.00	581.0
345.90	770.0
352.00	1131.0
353.00	687.0
353.90	1006.0
365.00	7403.0
365.90	1015.0
372.00	1633.0
383.00	373.0
402.00	780.0
403.00	769.0
421.10	767.0
422.00	749.0
423.00	4124.0
423.90	1186.0
441.00	11755.0
442.00	81104.0
443.10	14820.0
444.10	1323.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

Scan 2495 (17.962 min): BM015971.D
DFTPP30

m/z	Abundance
73.00	586.0
207.00	654.0
281.00	395.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30