

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071723\
 Data File : BM040923.D
 Acq On : 17 Jul 2023 15:18
 Operator : MA/JU
 Sample : PB153976BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS976

Quant Time: Jul 18 08:02:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 15 03:58:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	3937	0.400	ng/ul	0.00
4) Naphthalene-d8	10.658	136	13717	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.504	164	6105	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.253	188	10013	0.400	ng/ul	0.00
17) Chrysene-d12	21.448	240	7501	0.400	ng/ul	# 0.00
23) Perylene-d12	23.827	264	6866	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	38302	6.194	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.214	152	25691	1.336	ng/ul	-0.04
18) Fluoranthene-d10	19.287	212	7015	0.292	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.313	88	78023	12.395	ng/ul#	85
5) Naphthalene	10.708	128	1266787	33.483	ng/ul	99
7) 2-Methylnaphthalene	12.319	142	745397	33.546	ng/ul	99
8) 1-Methylnaphthalene	12.539	142	737465	33.209	ng/ul	99
10) Acenaphthylene	14.222	152	1102665	40.432	ng/ul	99
11) Acenaphthene	14.564	153	755128	32.787	ng/ul	98
12) Fluorene	15.554	166	881920	35.870	ng/ul	97
14) Pentachlorophenol	16.906	266	104306	54.101	ng/ul	93
15) Phenanthrene	17.295	178	1102954	35.462	ng/ul	98
16) Anthracene	17.388	178	1088677	41.653	ng/ul	99
19) Fluoranthene	19.315	202	1052526	33.033	ng/ul	96
20) Pyrene	19.682	202	1160947	33.986	ng/ul	95
21) Benzo(a)anthracene	21.431	228	983792	38.774	ng/ul	99
22) Chrysene	21.486	228	936487	32.357	ng/ul	98
24) Benzo(b)fluoranthene	23.102	252	984937	37.068	ng/ul	86
25) Benzo(k)fluoranthene	23.149	252	952291	36.366	ng/ul#	86
26) Benzo(a)pyrene	23.722	252	854319	35.693	ng/ul#	81
27) Indeno(1,2,3-cd)pyrene	26.287	276	1192795	35.783	ng/ul#	94
28) Dibenzo(a,h)anthracene	26.297	278	978445	37.489	ng/ul#	91
29) Benzo(g,h,i)perylene	27.038	276	933076	31.937	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071723\
Data File : BM040923.D
Acq On : 17 Jul 2023 15:18
Operator : MA/JU
Sample : PB153976BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS976

Quant Time: Jul 18 08:02:55 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 15 03:58:02 2023
Response via : Initial Calibration

