

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071923\
 Data File : BM040967.D
 Acq On : 19 Jul 2023 22:06
 Operator : MA/JU
 Sample : SSTD3.299
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2034

Quant Time: Jul 20 05:00:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM071923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 20 04:54:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.870	152	2610	0.400	ng/ul	0.00
4) Naphthalene-d8	10.675	136	9763	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.523	164	5036	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.278	188	9901	0.400	ng/ul	0.00
17) Chrysene-d12	21.466	240	7077	0.400	ng/ul	0.00
23) Perylene-d12	23.862	264	5815	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.279	96	13914	3.394	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.313	88	14533	3.483	ng/ul#	71
14) Pentachlorophenol	16.936	266	8221	4.312	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071923\
Data File : BM040967.D
Acq On : 19 Jul 2023 22:06
Operator : MA/JU
Sample : SST3.299
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2034

Quant Time: Jul 20 05:00:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM071923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 20 04:54:10 2023
Response via : Initial Calibration

