

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072318\
 Data File : BM016048.D
 Acq On : 24 Jul 2018 04:49
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 24 06:12:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 23 16:08:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.23	152	41778	20.00	ng	0.00
21) Naphthalene-d8	11.05	136	177736	20.00	ng	0.00
38) Acenaphthene-d10	14.85	164	98142	20.00	ng	0.00
63) Phenanthrene-d10	17.58	188	213049	20.00	ng	0.00
75) Chrysene-d12	21.70	240	242257	20.00	ng	0.00
86) Perylene-d12	24.07	264	242654	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.74	112	209556	83.32	ng	0.00
7) Phenol-d6	7.38	99	266902	81.26	ng	0.00
23) Nitrobenzene-d5	9.42	82	327230	85.63	ng	0.00
41) 2,4,6-Tribromophenol	16.34	330	97709	78.50	ng	0.00
44) 2-Fluorobiphenyl	13.47	172	674448	83.62	ng	0.00
78) Terphenyl-d14	20.15	244	971177	83.92	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.52	88	46659	39.596	ng	# 100
3) Pyridine	3.96	79	109720	38.651	ng	# 72
4) n-Nitrosodimethylamine	3.87	42	88728	40.073	ng	# 35
6) Aniline	7.55	93	151024	39.935	ng	# 89
8) 2-Chlorophenol	7.79	128	121251	40.564	ng	96
9) Benzaldehyde	7.37	77	93467	40.503	ng	92
10) Phenol	7.41	94	132714	40.409	ng	97
11) bis(2-Chloroethyl)ether	7.65	93	100976	38.918	ng	87
12) 1,3-Dichlorobenzene	8.11	146	140127	40.050	ng	# 89
13) 1,4-Dichlorobenzene	8.26	146	141059	39.752	ng	# 94
14) 1,2-Dichlorobenzene	8.58	146	137149	39.500	ng	# 93
15) Benzyl Alcohol	8.48	79	107095	40.949	ng	# 85
16) 2,2'-oxybis(1-Chloropropan	8.75	45	155047	39.046	ng	98
17) 2-Methylphenol	8.67	107	91942	40.958	ng	# 82
18) Hexachloroethane	9.30	117	63997	41.722	ng	96
19) n-Nitroso-di-n-propylamine	9.05	70	96471	40.185	ng	# 76
20) 3+4-Methylphenols	9.01	107	124420	38.523	ng	87
22) Acetophenone	9.07	105	177758	39.719	ng	# 84
24) Nitrobenzene	9.46	77	154651	40.206	ng	99
25) Isophorone	9.97	82	210065	40.188	ng	# 94
26) 2-Nitrophenol	10.17	139	62333	38.820	ng	# 50
27) 2,4-Dimethylphenol	10.22	122	88505	39.563	ng	# 81
28) bis(2-Chloroethoxy)methane	10.45	93	135168	39.787	ng	98
29) 2,4-Dichlorophenol	10.70	162	111554	41.965	ng	97
30) 1,2,4-Trichlorobenzene	10.91	180	129771	40.179	ng	97
31) Naphthalene	11.10	128	355173	39.484	ng	99
32) Benzoic acid	10.37	122	62276	36.519	ng	86
33) 4-Chloroaniline	11.22	127	150465	40.990	ng	# 87
34) Hexachlorobutadiene	11.36	225	93326	41.656	ng	97
35) Caprolactam	12.02	113	29259	39.740	ng	90
36) 4-Chloro-3-methylphenol	12.33	107	120462	41.201	ng	88
37) 2-Methylnaphthalene	12.69	142	244606	40.593	ng	# 90
39) 1,2,4,5-Tetrachlorobenzene	13.06	216	135764	40.633	ng	# 100
40) Hexachlorocyclopentadiene	13.02	237	60546	41.049	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072318\
 Data File : BM016048.D
 Acq On : 24 Jul 2018 04:49
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 24 06:12:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 23 16:08:38 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.30	196	86457	41.358	ng	99
43) 2,4,5-Trichlorophenol	13.37	196	86945	40.322	ng	# 83
45) 1,1'-Biphenyl	13.69	154	351340	39.668	ng	99
46) 2-Chloronaphthalene	13.74	162	279564	40.582	ng	# 89
47) 2-Nitroaniline	13.96	65	91303	40.089	ng	# 76
48) Acenaphthylene	14.58	152	411249	40.779	ng	100
49) Dimethylphthalate	14.30	163	338800	39.446	ng	99
50) 2,6-Dinitrotoluene	14.45	165	72029	41.680	ng	# 64
51) Acenaphthene	14.92	154	246809	39.793	ng	97
52) 3-Nitroaniline	14.77	138	72712	38.955	ng	# 67
53) 2,4-Dinitrophenol	15.00	184	25330	37.643	ng	# 82
54) Dibenzofuran	15.25	168	411216	40.233	ng	# 83
55) 4-Nitrophenol	15.08	139	51592	38.557	ng	85
56) 2,4-Dinitrotoluene	15.23	165	97839	39.595	ng	# 87
57) Fluorene	15.89	166	306504	40.356	ng	99
58) 2,3,4,6-Tetrachlorophenol	15.47	232	76378	41.209	ng	# 100
59) Diethylphthalate	15.65	149	374613	40.465	ng	95
60) 4-Chlorophenyl-phenylether	15.88	204	170061	40.222	ng	# 83
61) 4-Nitroaniline	15.93	138	70717	39.487	ng	# 10
62) Azobenzene	16.17	77	409750	42.027	ng	76
64) 4,6-Dinitro-2-methylphenol	15.99	198	51841	40.089	ng	85
65) n-Nitrosodiphenylamine	16.10	169	287540	40.987	ng	97
66) 4-Bromophenyl-phenylether	16.77	248	99012	41.042	ng	# 76
67) Hexachlorobenzene	16.89	284	108280	40.758	ng	# 64
68) Atrazine	17.03	200	103223	41.054	ng	97
69) Pentachlorophenol	17.23	266	65446	39.783	ng	98
70) Phenanthrene	17.63	178	473492	39.697	ng	99
71) Anthracene	17.72	178	468904	40.449	ng	98
72) Carbazole	17.99	167	485285	40.408	ng	98
73) Di-n-butylphthalate	18.51	149	618669	41.341	ng	# 96
74) Fluoranthene	19.61	202	539579	40.554	ng	99
76) Benzidine	19.79	184	272724	40.286	ng	99
77) Pyrene	19.97	202	574081	39.543	ng	99
79) Butylbenzylphthalate	20.82	149	301206	40.841	ng	# 79
80) Benzo(a)anthracene	21.69	228	589636	39.518	ng	99
81) 3,3'-Dichlorobenzidine	21.61	252	243938	40.589	ng	97
82) Chrysene	21.74	228	568935	39.750	ng	98
83) Bis(2-ethylhexyl)phthalate	21.57	149	436112	40.802	ng	# 90
84) Di-n-octyl phthalate	22.48	149	722091	40.197	ng	# 89
85) Indeno(1,2,3-cd)pyrene	26.50	276	672986	39.623	ng	# 94
87) Benzo(b)fluoranthene	23.35	252	585843	41.006	ng	# 95
88) Benzo(k)fluoranthene	23.40	252	572683	39.553	ng	98
89) Benzo(a)pyrene	23.97	252	554290	40.357	ng	100
90) Dibenzo(a,h)anthracene	26.50	278	572481	40.231	ng	97
91) Benzo(g,h,i)perylene	27.25	276	535679	39.802	ng	# 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072318\
Data File : BM016048.D
Acq On : 24 Jul 2018 04:49
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: Jul 24 06:12:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM072318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 23 16:08:38 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.23 min Scan# 840
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 152 Resp: 41778
Ion Ratio Lower Upper
152 100
150 156.2 119.5 179.3
115 67.4 43.0 64.4#




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 39.596 ng
RT: 3.52 min Scan# 39
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 88 Resp: 46659
Ion Ratio Lower Upper
88 100
58 77.0 0.0 0.0#
43 49.3 0.0 0.0#



Abundance Ion 88.00 (87.70 to 88.70): BM0
Ion 58.00 (57.70 to 58.70): BM0
Ion 43.00 (42.70 to 43.70): BM0





#3

Pyridine

Concen: 38.651 ng

RT: 3.96 min Scan# 114

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 109720

Ion Ratio Lower Upper

79 100

52 79.7 51.1 76.7#




51 56.5 26.1 39.1#



Abundance Ion 79.00 (78.70 to 79.70): BM016023.D (-111) (-)

Ion 52.00 (51.70 to 52.70): BM016023.D (-111) (-)

Ion 51.00 (50.70 to 51.70): BM016023.D (-111) (-)



#4

n-Nitrosodimethylamine

Concen: 40.073 ng

RT: 3.87 min Scan# 99

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion: 42 Resp: 88728

Ion Ratio Lower Upper

42 100

74 63.7 119.3 178.9#



44 3.7 5.4 8.2#




Abundance Ion 42.00 (41.70 to 42.70): BM016023.D (-95) (-)

Ion 74.00 (73.70 to 74.70): BM016023.D (-95) (-)

Ion 44.00 (43.70 to 44.70): BM016023.D (-95) (-)





#5

2-Fluorophenol

Concen: 40.073 ng

RT: 5.74 min Scan# 417

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 209556

Ion Ratio Lower Upper

112 100

64 62.5 38.6 57.8#




63 43.7 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#6

Aniline

Concen: 39.935 ng

RT: 7.55 min Scan# 725

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion: 93 Resp: 151024

Ion Ratio Lower Upper

93 100

66 45.2 30.8 46.2



65 25.1 16.0 24.0#




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 66.00 (65.70 to 66.70): BM0

Ion 65.00 (64.70 to 65.70): BM0





#7

Phenol-d6

Concen: 39.935 ng

RT: 7.38 min Scan# 696

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 266902

Ion Ratio Lower Upper

99 100

42 35.8 11.0 16.4#




71 51.7 23.4 35.2#



Abundance Ion 99.00 (98.70 to 99.70): BM016023.D (-690) (-)

Ion 42.00 (41.70 to 42.70): BM016023.D (-690) (-)

Ion 71.00 (70.70 to 71.70): BM016023.D (-690) (-)



#8

2-Chlorophenol

Concen: 40.564 ng

RT: 7.79 min Scan# 765

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion: 128 Resp: 121251

Ion Ratio Lower Upper

128 100

130 30.6 12.0 52.0



64 48.6 24.9 64.9




Abundance Ion 128.00 (127.70 to 128.70): BM016023.D (-758) (-)

Ion 130.00 (129.70 to 130.70): BM016023.D (-758) (-)

Ion 64.00 (63.70 to 64.70): BM016023.D (-758) (-)





#9

Benzaldehyde

Concen: 40.503 ng

RT: 7.37 min Scan# 694

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 93467

Ion Ratio Lower Upper

77 100

105 91.2 78.8 118.8

106 86.5 73.7 113.7



#10

Phenol

Concen: 40.409 ng

RT: 7.41 min Scan# 701

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49




Tgt Ion: 94 Resp: 132714


Ion Ratio Lower Upper

94 100

65 41.5 18.8 58.8

66 58.6 39.9 79.9









#11
bis(2-Chloroethyl)ether
Concen: 38.918 ng
RT: 7.65 min Scan# 741
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 100976
Ion Ratio Lower Upper
93 100
63 84.7 49.5 89.5
95 32.4 13.5 53.5



#12
1,3-Dichlorobenzene
Concen: 40.050 ng
RT: 8.11 min Scan# 820
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 146 Resp: 140127
Ion Ratio Lower Upper
146 100
148 63.6 50.9 76.3
75 46.2 21.4 32.2#





#13

1,4-Dichlorobenzene

Concen: 39.752 ng

RT: 8.26 min Scan# 846

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:146 Resp: 141059

Ion Ratio Lower Upper

146 100

148 62.9 51.9 77.9




111 44.3 29.2 43.8#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 39.500 ng

RT: 8.58 min Scan# 900

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:146 Resp: 137149

Ion Ratio Lower Upper

146 100

148 63.5 52.3 78.5



111 47.3 30.6 45.8#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 40.949 ng

RT: 8.48 min Scan# 883

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 107095

Ion Ratio Lower Upper

79 100

108 61.9 65.0 97.4#

77 72.0 53.0 79.6



#16

2,2'-oxybis(1-chloropropane)

Concen: 39.046 ng

RT: 8.75 min Scan# 929

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49




Tgt Ion: 45 Resp: 155047


Ion Ratio Lower Upper

45 100

77 15.9 0.0 35.2

79 10.9 0.0 32.0





#17

2-Methylphenol

Concen: 40.958 ng

RT: 8.67 min Scan# 916

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:107 Resp: 91942


Ion Ratio Lower Upper

107 100

108 107.3 89.8 134.8

77 65.7 32.6 48.8#

79 61.3 32.8 49.2#






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0



#18

Hexachloroethane

Concen: 41.722 ng

RT: 9.30 min Scan# 1023

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:117 Resp: 63997

Ion Ratio Lower Upper

117 100

119 92.1 79.2 118.8



201 85.7 69.8 104.6






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 40.185 ng

RT: 9.05 min Scan# 979

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 96471





Ion Ratio Lower Upper

70 100

42 80.5 44.5 66.7#

101 9.8 7.4 11.2

130 20.9 15.3 22.9



#20

3+4-Methylphenols

Concen: 38.523 ng

RT: 9.01 min Scan# 973

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Tgt Ion: 107 Resp: 124420


Ion Ratio Lower Upper

107 100

108 86.2 73.2 113.2

77 44.8 10.7 50.7

79 39.5 9.6 49.6









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.05 min Scan# 1320
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 177736
Ion Ratio Lower Upper
136 100
137 10.4 8.7 13.1
54 10.0 4.6 6.8#
68 4.0 3.7 5.5




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BM0
Ion 68.00 (67.70 to 68.70): BM0






#22
Acetophenone
Concen: 39.719 ng
RT: 9.07 min Scan# 983
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:105 Resp: 177758
Ion Ratio Lower Upper
105 100
71 0.3 0.6 1.0#
51 41.0 21.6 32.4#
120 19.9 16.1 24.1



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BM0
Ion 51.00 (50.70 to 51.70): BM0
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 39.719 ng
RT: 9.42 min Scan# 1042
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 327230
Ion Ratio Lower Upper
82 100
128 35.0 32.8 49.2
54 62.6 40.6 60.8#



#24
Nitrobenzene
Concen: 40.206 ng
RT: 9.46 min Scan# 1049
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 77 Resp: 154651
Ion Ratio Lower Upper
77 100
123 40.8 32.6 49.0
65 15.9 10.9 16.3









#25
Isophorone
Concen: 40.188 ng
RT: 9.97 min Scan# 1137
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 210065
Ion Ratio Lower Upper
82 100
95 9.8 5.8 8.6#
138 18.1 16.3 24.5



#26
2-Nitrophenol
Concen: 38.820 ng
RT: 10.17 min Scan# 1170
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 139 Resp: 62333
Ion Ratio Lower Upper
139 100
109 51.0 20.1 30.1#
65 69.2 31.5 47.3#









#27
2,4-Dimethylphenol
Concen: 39.563 ng
RT: 10.22 min Scan# 1178
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 122 Resp: 88505
Ion Ratio Lower Upper
122 100
107 135.1 84.7 127.1#
121 60.3 46.6 69.8



#28
bis(2-Chloroethoxy)methane
Concen: 39.787 ng
RT: 10.45 min Scan# 1218
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 93 Resp: 135168
Ion Ratio Lower Upper
93 100
95 32.8 25.5 38.3
123 12.2 8.6 13.0





#29

2,4-Dichlorophenol

Concen: 41.965 ng

RT: 10.70 min Scan# 1261

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 111554

Ion Ratio Lower Upper

162 100

164 64.8 42.8 82.8

98 37.0 14.5 54.5



#30

1,2,4-Trichlorobenzene

Concen: 40.179 ng

RT: 10.91 min Scan# 1296

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49




Tgt Ion:180 Resp: 129771


Ion Ratio Lower Upper

180 100

182 94.8 77.6 116.4

145 32.8 23.4 35.2









#31
Naphthalene
Concen: 39.484 ng
RT: 11.10 min Scan# 1328
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 355173
Ion Ratio Lower Upper
128 100
129 11.3 8.9 13.3
127 13.5 10.4 15.6



#32
Benzoic acid
Concen: 36.519 ng
RT: 10.37 min Scan# 1205
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:122 Resp: 62276
Ion Ratio Lower Upper
122 100
105 129.8 99.9 139.9
77 113.0 73.7 113.7









#33
4-Chloroaniline
Concen: 40.990 ng
RT: 11.22 min Scan# 1349
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 150465
Ion Ratio Lower Upper
127 100
129 31.3 25.3 37.9
65 37.0 17.6 26.4#
92 20.9 13.1 19.7#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BM
Ion 92.00 (91.70 to 92.70): BM






#34
Hexachlorobutadiene
Concen: 41.656 ng
RT: 11.36 min Scan# 1372
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:225 Resp: 93326
Ion Ratio Lower Upper
225 100
223 62.7 47.9 71.9
227 61.0 50.1 75.1



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 39.740 ng

RT: 12.02 min Scan# 1485

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:113 Resp: 29259

Ion Ratio Lower Upper

113 100

55 172.3 145.9 185.9




56 140.8 101.0 141.0



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 41.201 ng

RT: 12.33 min Scan# 1537

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:107 Resp: 120462

Ion Ratio Lower Upper

107 100

144 24.7 23.3 34.9



142 78.0 72.6 109.0




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 40.593 ng
 RT: 12.69 min Scan# 1599
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:142 Resp: 244606
 Ion Ratio Lower Upper
 142 100
 141 90.7 71.9 107.9
 115 50.9 25.4 38.0#



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.85 min Scan# 1966
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:164 Resp: 98142
 Ion Ratio Lower Upper
 164 100
 162 107.0 82.4 123.6
 160 47.7 35.8 53.6









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.633 ng
 RT: 13.06 min Scan# 1661
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Instrument :
 BNA_M
 ClientSampled :
 SSTDC040

Tgt Ion: 216 Resp: 135764
 Ion Ratio Lower Upper
 216 100
 214 78.3 0.0 0.0#
 179 25.1 0.0 0.0#
 108 24.3 0.0 0.0#




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 41.049 ng
 RT: 13.02 min Scan# 1654
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion: 237 Resp: 60546
 Ion Ratio Lower Upper
 237 100
 235 61.8 42.0 82.0
 272 10.8 0.0 33.4



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 41.049 ng
 RT: 16.34 min Scan# 2219
 Delta R.T. -0.00 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTCCCC040


Tgt Ion:330 Resp: 97709
 Ion Ratio Lower Upper
 330 100
 332 97.5 0.0 0.0#
 141 68.6 0.0 0.0#



#42
 2,4,6-Trichlorophenol
 Concen: 41.358 ng
 RT: 13.30 min Scan# 1702
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:196 Resp: 86457
 Ion Ratio Lower Upper
 196 100
 198 97.0 76.3 114.5
 200 32.1 25.6 38.4









#43
2,4,5-Trichlorophenol
Concen: 40.322 ng
RT: 13.37 min Scan# 1715
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 86945
Ion Ratio Lower Upper
196 100
198 93.7 79.4 119.0
97 64.7 26.8 40.2#
132 27.8 18.6 28.0



#44
2-Fluorobiphenyl
Concen: 40.322 ng
RT: 13.47 min Scan# 1732
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:172 Resp: 674448
Ion Ratio Lower Upper
172 100
171 36.1 28.5 42.7
170 24.3 18.8 28.2









#45
 1,1'-Biphenyl
 Concen: 39.668 ng
 RT: 13.69 min Scan# 1768
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Tgt Ion:154 Resp: 351340
 Ion Ratio Lower Upper
 154 100
 153 40.5 20.7 60.7
 76 11.6 0.0 30.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BM






#46
 2-Chloronaphthalene
 Concen: 40.582 ng
 RT: 13.74 min Scan# 1777
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:162 Resp: 279564
 Ion Ratio Lower Upper
 162 100
 127 44.6 26.9 40.3#
 164 31.8 26.8 40.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 40.089 ng
RT: 13.96 min Scan# 1814
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 65 Resp: 91303
Ion Ratio Lower Upper
65 100
92 58.6 59.1 88.7#
138 87.1 93.9 140.9#



#48
Acenaphthylene
Concen: 40.779 ng
RT: 14.58 min Scan# 1920
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 152 Resp: 411249
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.4 10.6 16.0





#49

Dimethylphthalate

Concen: 39.446 ng

RT: 14.30 min Scan# 1873

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 338800

Ion Ratio Lower Upper

163 100

194 3.6 2.7 4.1




164 10.5 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 41.680 ng

RT: 14.45 min Scan# 1897

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:165 Resp: 72029

Ion Ratio Lower Upper

165 100

63 82.2 44.1 66.1#



89 80.1 44.3 66.5#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 39.793 ng

RT: 14.92 min Scan# 1977

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 246809

Ion Ratio Lower Upper

154 100

153 114.4 90.0 135.0




152 56.3 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 38.955 ng

RT: 14.77 min Scan# 1953

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:138 Resp: 72712

Ion Ratio Lower Upper

138 100

108 11.7 7.3 10.9#



92 130.5 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53
2,4-Dinitrophenol
Concen: 37.643 ng
RT: 15.00 min Scan# 1991
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 25330
Ion Ratio Lower Upper
184 100
63 98.3 69.2 103.8
154 87.1 53.3 79.9#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BM016023.D (-1986) (-)
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 40.233 ng
RT: 15.25 min Scan# 2034
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:168 Resp: 411216
Ion Ratio Lower Upper
168 100
139 47.4 27.3 40.9#
169 13.3 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 38.557 ng
RT: 15.08 min Scan# 2005
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:139 Resp: 51592
Ion Ratio Lower Upper
139 100
109 131.4 130.0 170.0
65 131.3 130.0 170.0



#56
2,4-Dinitrotoluene
Concen: 39.595 ng
RT: 15.23 min Scan# 2031
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:165 Resp: 97839
Ion Ratio Lower Upper
165 100
63 77.5 52.4 78.6
89 101.0 72.4 108.6
182 3.0 2.0 3.0#









#57
Fluorene
 Concen: 40.356 ng
 RT: 15.89 min Scan# 2143
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:166 Resp: 306504
 Ion Ratio Lower Upper
 166 100
 165 99.8 79.0 118.4
 167 13.5 10.3 15.5



#58
2,3,4,6-Tetrachlorophenol
 Concen: 41.209 ng
 RT: 15.47 min Scan# 2072
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:232 Resp: 76378
 Ion Ratio Lower Upper
 232 100
 131 69.1 0.0 0.0#
 130 4.4 0.0 0.0#
 166 38.9 0.0 0.0#









#59
Diethylphthalate
Concen: 40.465 ng
RT: 15.65 min Scan# 2102
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 374613
Ion Ratio Lower Upper
149 100
177 19.2 18.3 27.5
150 12.4 9.8 14.8



#60
4-Chlorophenyl-phenylether
Concen: 40.222 ng
RT: 15.88 min Scan# 2141
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:204 Resp: 170061
Ion Ratio Lower Upper
204 100
206 33.0 26.6 39.8
141 76.2 45.2 67.8#









#61
4-Nitroaniline
Concen: 39.487 ng
RT: 15.93 min Scan# 2150
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 138 Resp: 70717
Ion Ratio Lower Upper
138 100
92 60.9 16.7 56.7#
108 147.6 37.6 77.6#



#62
Azobenzene
Concen: 42.027 ng
RT: 16.17 min Scan# 2190
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion: 77 Resp: 409750
Ion Ratio Lower Upper
77 100
182 18.8 6.5 46.5
105 14.7 3.8 43.8
51 45.1 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.58 min Scan# 2430
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:188 Resp: 213049
Ion Ratio Lower Upper
188 100
94 7.6 8.7 13.1#
80 9.1 9.3 13.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 40.089 ng
RT: 15.99 min Scan# 2160
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:198 Resp: 51841
Ion Ratio Lower Upper
198 100
51 67.7 28.8 68.8
105 48.1 30.5 70.5









#65
 n-Nitrosodiphenylamine
 Concen: 40.987 ng
 RT: 16.10 min Scan# 2178
 Delta R.T. -0.00 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:169 Resp: 287540
 Ion Ratio Lower Upper
 169 100
 168 69.1 53.9 80.9
 167 38.4 28.9 43.3



#66
 4-Bromophenyl-phenylether
 Concen: 41.042 ng
 RT: 16.77 min Scan# 2292
 Delta R.T. -0.00 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:248 Resp: 99012
 Ion Ratio Lower Upper
 248 100
 250 97.0 77.4 116.0
 141 104.7 45.2 67.8#









#67
Hexachlorobenzene
Concen: 40.758 ng
RT: 16.89 min Scan# 2312
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:284 Resp: 108280
Ion Ratio Lower Upper
284 100
142 58.1 20.4 30.6#
249 36.2 23.8 35.6#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 41.054 ng
RT: 17.03 min Scan# 2337
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:200 Resp: 103223
Ion Ratio Lower Upper
200 100
173 25.2 4.8 44.8
215 44.0 26.9 66.9



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 39.783 ng
 RT: 17.23 min Scan# 2371
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Tgt Ion:266 Resp: 65446
 Ion Ratio Lower Upper
 266 100
 268 64.3 52.0 78.0
 264 64.1 49.0 73.4




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E





#70
 Phenanthrene
 Concen: 39.697 ng
 RT: 17.63 min Scan# 2438
 Delta R.T. -0.00 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:178 Resp: 473492
 Ion Ratio Lower Upper
 178 100
 176 19.5 15.2 22.8
 179 14.9 12.1 18.1



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 40.449 ng
 RT: 17.72 min Scan# 2453
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:178 Resp: 468904
 Ion Ratio Lower Upper
 178 100
 176 19.6 14.6 22.0
 179 15.3 12.6 19.0



#72
 Carbazole
 Concen: 40.408 ng
 RT: 17.99 min Scan# 2499
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:167 Resp: 485285
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.2 25.8
 139 15.5 10.8 16.2









#73
Di-n-butylphthalate
Concen: 41.341 ng
RT: 18.51 min Scan# 2588
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 618669
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 7.4 3.7 5.5#



#74
Fluoranthene
Concen: 40.554 ng
RT: 19.61 min Scan# 2775
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:202 Resp: 539579
Ion Ratio Lower Upper
202 100
101 8.5 0.0 28.7
203 17.8 0.0 37.2









#75
Chrysene-d12
Concen: 20.000 ng
RT: 21.70 min Scan# 3130
Delta R.T. -0.01 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:240 Resp: 242257
Ion Ratio Lower Upper
240 100
120 9.1 8.2 12.2
236 27.2 20.0 30.0



#76
Benzidine
Concen: 40.286 ng
RT: 19.79 min Scan# 2806
Delta R.T. -0.00 min
Lab File: BM016048.D
Acq: 24 Jul 2018 04:49

Tgt Ion:184 Resp: 272724
Ion Ratio Lower Upper
184 100
185 13.9 11.3 16.9
183 12.1 8.9 13.3





#77

Pyrene

Concen: 39.543 ng

RT: 19.97 min Scan# 2836

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:202 Resp: 574081

Ion Ratio Lower Upper

202 100

200 21.8 16.9 25.3




203 17.2 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 39.543 ng

RT: 20.15 min Scan# 2867

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:244 Resp: 971177

Ion Ratio Lower Upper

244 100

212 9.2 4.9 7.3#



122 9.8 6.2 9.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 40.841 ng

RT: 20.82 min Scan# 2981

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 301206

Ion Ratio Lower Upper

149 100

91 82.0 50.1 75.1#




206 16.7 16.2 24.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BM0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 39.518 ng

RT: 21.69 min Scan# 3128

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:228 Resp: 589636

Ion Ratio Lower Upper

228 100

226 27.7 21.7 32.5



229 19.6 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#81

3,3'-Dichlorobenzidine

Concen: 40.589 ng

RT: 21.61 min Scan# 3115

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampled :

SSTDCCC040


Tgt Ion:252 Resp: 243938

Ion Ratio Lower Upper

252 100

254 62.1 51.9 77.9




126 12.0 9.8 14.8



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E



#82

Chrysene

Concen: 39.750 ng

RT: 21.74 min Scan# 3137

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:228 Resp: 568935

Ion Ratio Lower Upper

228 100

226 31.1 24.1 36.1



229 19.9 15.5 23.3




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 40.802 ng
 RT: 21.57 min Scan# 3108
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:149 Resp: 436112
 Ion Ratio Lower Upper
 149 100
 167 22.3 22.4 33.6#
 279 3.2 3.4 5.0#



#84
 Di-n-octyl phthalate
 Concen: 40.197 ng
 RT: 22.48 min Scan# 3263
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion:149 Resp: 722091
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.2 1.8
 43 13.7 7.3 10.9#









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 39.623 ng
 RT: 26.50 min Scan# 3946
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTCCCC040


Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.2	12.0	18.0#
277	25.2	20.0	30.0



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.07 min Scan# 3533
 Delta R.T. -0.01 min
 Lab File: BM016048.D
 Acq: 24 Jul 2018 04:49

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.6	27.8
265	22.5	17.3	25.9





#87

Benzo(b)fluoranthene

Concen: 41.006 ng

RT: 23.35 min Scan# 3411

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 585843

Ion Ratio Lower Upper

252 100

253 21.6 18.0 27.0




125 8.4 9.9 14.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 39.553 ng

RT: 23.40 min Scan# 3419

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion:252 Resp: 572683

Ion Ratio Lower Upper

252 100

253 22.1 17.2 25.8



125 8.5 8.1 12.1




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 40.357 ng

RT: 23.97 min Scan# 3516

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 252 Resp: 554290

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4




125 9.0 7.4 11.2



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 40.231 ng

RT: 26.50 min Scan# 3947

Delta R.T. -0.00 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49


Tgt Ion: 278 Resp: 572481

Ion Ratio Lower Upper

278 100

139 13.5 13.4 20.0



279 23.5 19.0 28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 39.802 ng

RT: 27.25 min Scan# 4074

Delta R.T. -0.01 min

Lab File: BM016048.D

Acq: 24 Jul 2018 04:49

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 276 Resp: 535679

Ion Ratio Lower Upper

276 100

277 23.5 18.5 27.7

138 17.9 19.0 28.6#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

