

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072318\
 Data File : BM016048.D
 Acq On : 24 Jul 2018 04:49
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 24 05:23:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 23 16:08:38 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.564	0.558	1.1	104	0.00
3	Pyridine	1.359	1.313	3.4	101	0.00
4	n-Nitrosodimethylamine	1.060	1.062	-0.2	104	0.00
5 S	2-Fluorophenol	1.204	1.254	-4.2	109	0.00
6	Aniline	1.810	1.807	0.2	105	0.00
7 S	Phenol-d6	1.572	1.597	-1.6	107	0.00
8	2-Chlorophenol	1.431	1.451	-1.4	107	0.00
9	Benzaldehyde	1.105	1.119	-1.3	106	0.00
10 C	Phenol	1.572	1.588	-1.0	106	0.00
11	bis(2-Chloroethyl)ether	1.242	1.208	2.7	105	0.00
12	1,3-Dichlorobenzene	1.675	1.677	-0.1	109	0.00
13 C	1,4-Dichlorobenzene	1.699	1.688	0.6	107	0.00
14	1,2-Dichlorobenzene	1.662	1.641	1.3	109	0.00
15	Benzyl Alcohol	1.252	1.282	-2.4	108	0.00
16	2,2'-oxybis(1-Chloropropane	1.901	1.856	2.4	106	0.00
17	2-Methylphenol	1.075	1.100	-2.3	106	0.00
18	Hexachloroethane	0.734	0.766	-4.4	109	0.00
19 P	n-Nitroso-di-n-propylamine	1.149	1.155	-0.5	103	0.00
20	3+4-Methylphenols	1.546	1.489	3.7	102	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.504	0.500	0.8	104	0.00
23 S	Nitrobenzene-d5	0.430	0.460	-7.0	110	0.00
24	Nitrobenzene	0.433	0.435	-0.5	102	0.00
25	Isophorone	0.588	0.591	-0.5	102	0.00
26 C	2-Nitrophenol	0.181	0.175	3.3	103	0.00
27	2,4-Dimethylphenol	0.252	0.249	1.2	102	0.00
28	bis(2-Chloroethoxy)methane	0.382	0.380	0.5	101	0.00
29 C	2,4-Dichlorophenol	0.299	0.314	-5.0	105	0.00
30	1,2,4-Trichlorobenzene	0.363	0.365	-0.6	106	0.00
31	Naphthalene	1.012	0.999	1.3	105	0.00
32	Benzoic acid	0.192	0.175	8.9	98	0.00
33	4-Chloroaniline	0.413	0.423	-2.4	103	0.00
34 C	Hexachlorobutadiene	0.252	0.263	-4.4	110	0.00
35	Caprolactam	0.083	0.082	1.2	102	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.339	-3.0	103	0.00
37	2-Methylnaphthalene	0.678	0.688	-1.5	105	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
39	1,2,4,5-Tetrachlorobenzene	0.681	0.692	-1.6	109	0.00
40 P	Hexachlorocyclopentadiene	0.278	0.308	-10.8	117	0.00
41 S	2,4,6-Tribromophenol	0.254	0.249	2.0	106	0.00
42 C	2,4,6-Trichlorophenol	0.426	0.440	-3.3	106	0.00
43	2,4,5-Trichlorophenol	0.439	0.443	-0.9	105	0.00
44 S	2-Fluorobiphenyl	1.644	1.718	-4.5	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072318\
 Data File : BM016048.D
 Acq On : 24 Jul 2018 04:49
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 24 05:23:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 23 16:08:38 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.805	1.790	0.8	105	0.00
46	2-Chloronaphthalene	1.404	1.424	-1.4	108	0.00
47	2-Nitroaniline	0.464	0.465	-0.2	104	0.00
48	Acenaphthylene	2.055	2.095	-1.9	107	0.00
49	Dimethylphthalate	1.750	1.726	1.4	105	0.00
50	2,6-Dinitrotoluene	0.352	0.367	-4.3	107	0.00
51 C	Acenaphthene	1.264	1.257	0.6	108	0.00
52	3-Nitroaniline	0.380	0.370	2.6	102	0.00
53 P	2,4-Dinitrophenol	0.131	0.129	1.5	104	0.00
54	Dibenzofuran	2.083	2.095	-0.6	108	0.00
55 P	4-Nitrophenol	0.273	0.263	3.7	101	0.00
56	2,4-Dinitrotoluene	0.504	0.498	1.2	104	0.00
57	Fluorene	1.548	1.562	-0.9	107	0.00
58	2,3,4,6-Tetrachlorophenol	0.378	0.389	-2.9	105	0.00
59	Diethylphthalate	1.887	1.909	-1.2	106	0.00
60	4-Chlorophenyl-phenylether	0.862	0.866	-0.5	107	0.00
61	4-Nitroaniline	0.365	0.360	1.4	106	0.00
62	Azobenzene	1.987	2.088	-5.1	108	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
64	4,6-Dinitro-2-methylphenol	0.121	0.122	-0.8	108	0.00
65 c	n-Nitrosodiphenylamine	0.659	0.675	-2.4	109	0.00
66	4-Bromophenyl-phenylether	0.226	0.232	-2.7	109	0.00
67	Hexachlorobenzene	0.249	0.254	-2.0	109	0.00
68	Atrazine	0.236	0.242	-2.5	107	0.00
69 C	Pentachlorophenol	0.154	0.154	0.0	108	0.00
70	Phenanthrene	1.120	1.111	0.8	108	0.00
71	Anthracene	1.088	1.100	-1.1	109	0.00
72	Carbazole	1.127	1.139	-1.1	106	0.00
73	Di-n-butylphthalate	1.405	1.452	-3.3	109	0.00
74 C	Fluoranthene	1.249	1.266	-1.4	109	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
76	Benzidine	0.559	0.563	-0.7	108	0.00
77	Pyrene	1.199	1.185	1.2	109	0.00
78 S	Terphenyl-d14	0.955	1.002	-4.9	117	0.00
79	Butylbenzylphthalate	0.609	0.622	-2.1	111	0.00
80	Benzo(a)anthracene	1.232	1.217	1.2	111	0.00
81	3,3'-Dichlorobenzidine	0.496	0.503	-1.4	112	0.00
82	Chrysene	1.182	1.174	0.7	113	0.00
83	Bis(2-ethylhexyl)phthalate	0.882	0.900	-2.0	111	0.00
84 c	Di-n-octyl phthalate	1.483	1.490	-0.5	110	0.00
85	Indeno(1,2,3-cd)pyrene	1.402	1.389	0.9	113	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	114	0.00
87	Benzo(b)fluoranthene	1.178	1.207	-2.5	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072318\
Data File : BM016048.D
Acq On : 24 Jul 2018 04:49
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jul 24 05:23:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 23 16:08:38 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.193	1.180	1.1	110	0.00
89 C	Benzo(a)pyrene	1.132	1.142	-0.9	114	0.00
90	Dibenzo(a,h)anthracene	1.173	1.180	-0.6	113	0.00
91	Benzo(a,h,i)perylene	1.109	1.104	0.5	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0