

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072320\
 Data File : BM026853.D
 Acq On : 23 Jul 2020 16:18
 Operator : JU/CG
 Sample : L3380-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 VNJ-242

Quant Time: Jul 24 04:01:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM070820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 22 14:16:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.28	152	89675	20.00	ng	-0.02
21) Naphthalene-d8	10.04	136	347549	20.00	ng	-0.01
39) Acenaphthene-d10	13.96	164	249883	20.00	ng	0.00
64) Phenanthrene-d10	16.76	188	617855	20.00	ng	0.04
76) Chrysene-d12	20.96	240	500724	20.00	ng	0.01
86) Perylene-d12	23.01	264	463363	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.92	112	540962	101.11	ng	-0.02
7) Phenol-d6	6.50	99	708333	83.10	ng	-0.02
23) Nitrobenzene-d5	8.43	82	486416	59.71	ng	-0.02
42) 2,4,6-Tribromophenol	15.49	330	340121	93.44	ng	0.02
45) 2-Fluorobiphenyl	12.55	172	936632	49.84	ng	-0.01
79) Terphenyl-d14	19.42	244	1895188	74.49	ng	0.03
Target Compounds						
50) Dimethylphthalate	13.43	163	211644	9.902	ng	# 88
78) Pyrene	19.21	202	98681	3.128	ng	# 15

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072320\
Data File : BM026853.D
Acq On : 23 Jul 2020 16:18
Operator : JU/CG
Sample : L3380-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
VNJ-242

Quant Time: Jul 24 04:01:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM070820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 22 14:16:10 2020
Response via : Initial Calibration

