

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072321\
 Data File : BM031116.D
 Acq On : 24 Jul 2021 12:13
 Operator : CG/JU
 Sample : PB137709BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS709

Quant Time: Jul 26 01:43:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 26 01:37:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.635	152	1222	0.40	ng/ul	0.00
4) Naphthalene-d8	10.401	136	4558	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.262	164	2972	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.006	188	6493	0.40	ng/ul	0.00
17) Chrysene-d12	21.190	240	5414	0.40	ng/ul	0.00
23) Perylene-d12	23.363	264	4423	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	872	0.68	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	2798	0.35	ng/ul	0.00
18) Fluoranthene-d10	19.035	212	6941	0.42	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.248	88	2199	1.75	ng/ul#	84
5) Naphthalene	10.451	128	4616	0.34	ng/ul	99
7) 2-Methylnaphthalene	12.065	142	3286	0.33	ng/ul	100
8) 1-Methylnaphthalene	12.286	142	3083	0.32	ng/ul	99
10) Acenaphthylene	13.981	152	4137	0.34	ng/ul	96
11) Acenaphthene	14.326	153	3692	0.34	ng/ul	96
12) Fluorene	15.316	166	4354	0.34	ng/ul	99
14) Pentachlorophenol	16.659	266	1239	0.55	ng/ul	98
15) Phenanthrene	17.048	178	7308	0.35	ng/ul	99
16) Anthracene	17.141	178	6708	0.35	ng/ul	98
19) Fluoranthene	19.066	202	8783	0.40	ng/ul	97
20) Pyrene	19.428	202	8830	0.40	ng/ul	99
21) Benzo(a)anthracene	21.176	228	8036	0.37	ng/ul	99
22) Chrysene	21.227	228	8657	0.37	ng/ul	99
24) Benzo(b)fluoranthene	22.716	252	8108	0.39	ng/ul	99
25) Benzo(k)fluoranthene	22.757	252	8184	0.39	ng/ul	98
26) Benzo(a)pyrene	23.270	252	6598	0.37	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	25.516	276	7610	0.32	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.531	278	6097	0.32	ng/ul	99
29) Benzo(g,h,i)perylene	26.173	276	6284	0.31	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072321\
Data File : BM031116.D
Acq On : 24 Jul 2021 12:13
Operator : CG/JU
Sample : PB137709BS
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS709

Quant Time: Jul 26 01:43:31 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 26 01:37:33 2021
Response via : Initial Calibration

