

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM072321\  
 Data File : BM031138.D  
 Acq On : 25 Jul 2021 03:04  
 Operator : CG/JU  
 Sample : PB137657BS  
 Misc :  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SLCS657

Quant Time: Jul 26 03:06:31 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM072221.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 26 03:04:43 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|------|--------|----------|
| -----                       |        |      |          |      |        |          |
| Internal Standards          |        |      |          |      |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.635  | 152  | 936      | 0.40 | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.401 | 136  | 3541     | 0.40 | ng/ul  | 0.00     |
| 9) Acenaphthene-d10         | 14.258 | 164  | 2309     | 0.40 | ng/ul  | -0.01    |
| 13) Phenanthrene-d10        | 17.006 | 188  | 5098     | 0.40 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.190 | 240  | 4671     | 0.40 | ng/ul  | 0.00     |
| 23) Perylene-d12            | 23.360 | 264  | 4200     | 0.40 | ng/ul  | -0.01    |
| System Monitoring Compounds |        |      |          |      |        |          |
| 3) 1,4-Dioxane-d8           | 3.217  | 96   | 760      | 0.77 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 11.993 | 152  | 2400     | 0.39 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.035 | 212  | 5953     | 0.42 | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |      |        |          |
|                             |        |      |          |      |        | Qvalue   |
| 2) 1,4-Dioxane              | 3.248  | 88   | 1849     | 1.92 | ng/ul# | 81       |
| 5) Naphthalene              | 10.446 | 128  | 3911     | 0.37 | ng/ul  | 99       |
| 7) 2-Methylnaphthalene      | 12.065 | 142  | 2750     | 0.36 | ng/ul  | 99       |
| 8) 1-Methylnaphthalene      | 12.285 | 142  | 2560     | 0.34 | ng/ul  | 100      |
| 10) Acenaphthylene          | 13.981 | 152  | 3540     | 0.37 | ng/ul  | 98       |
| 11) Acenaphthene            | 14.322 | 153  | 3015     | 0.36 | ng/ul  | 96       |
| 12) Fluorene                | 15.315 | 166  | 3599     | 0.37 | ng/ul  | 98       |
| 14) Pentachlorophenol       | 16.659 | 266  | 1232     | 0.70 | ng/ul  | 98       |
| 15) Phenanthrene            | 17.047 | 178  | 6035     | 0.37 | ng/ul  | 98       |
| 16) Anthracene              | 17.138 | 178  | 5597     | 0.37 | ng/ul  | 99       |
| 19) Fluoranthene            | 19.065 | 202  | 7235     | 0.38 | ng/ul  | 97       |
| 20) Pyrene                  | 19.428 | 202  | 7295     | 0.38 | ng/ul  | 98       |
| 21) Benzo(a)anthracene      | 21.173 | 228  | 7127     | 0.38 | ng/ul  | 99       |
| 22) Chrysene                | 21.226 | 228  | 7590     | 0.38 | ng/ul  | 98       |
| 24) Benzo(b)fluoranthene    | 22.713 | 252  | 7608     | 0.39 | ng/ul  | 99       |
| 25) Benzo(k)fluoranthene    | 22.757 | 252  | 7598     | 0.38 | ng/ul  | 98       |
| 26) Benzo(a)pyrene          | 23.268 | 252  | 6440     | 0.38 | ng/ul# | 96       |
| 27) Indeno(1,2,3-cd)pyrene  | 25.511 | 276  | 7719     | 0.35 | ng/ul# | 97       |
| 28) Dibenzo(a,h)anthracene  | 25.528 | 278  | 6214     | 0.35 | ng/ul  | 99       |
| 29) Benzo(g,h,i)perylene    | 26.166 | 276  | 6459     | 0.34 | ng/ul  | 96       |
| -----                       |        |      |          |      |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM072321\  
Data File : BM031138.D  
Acq On : 25 Jul 2021 03:04  
Operator : CG/JU  
Sample : PB137657BS  
Misc :  
ALS Vial : 52 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SLCS657

Quant Time: Jul 26 03:06:31 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM072221.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jul 26 03:04:43 2021  
Response via : Initial Calibration

