

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072321\
 Data File : BM031122.D
 Acq On : 24 Jul 2021 16:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4065

Quant Time: Jul 26 02:07:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 18:04:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.639	152	1009	0.40	ng/ul	0.00
4) Naphthalene-d8	10.401	136	3797	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.262	164	2522	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.009	188	5602	0.40	ng/ul	0.00
17) Chrysene-d12	21.190	240	5090	0.40	ng/ul	0.00
23) Perylene-d12	23.365	264	4287	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.220	96	388	0.37	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	2536	0.38	ng/ul	0.00
18) Fluoranthene-d10	19.039	212	6270	0.40	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.251	88	419	0.40	ng/ul	96
5) Naphthalene	10.451	128	4287	0.38	ng/ul	99
7) 2-Methylnaphthalene	12.065	142	3053	0.37	ng/ul	100
8) 1-Methylnaphthalene	12.290	142	3005	0.38	ng/ul	98
10) Acenaphthylene	13.986	152	4000	0.39	ng/ul	97
11) Acenaphthene	14.326	153	3450	0.38	ng/ul	98
12) Fluorene	15.316	166	4038	0.38	ng/ul	99
14) Pentachlorophenol	16.659	266	669	0.35	ng/ul	99
15) Phenanthrene	17.051	178	6785	0.38	ng/ul	99
16) Anthracene	17.141	178	6290	0.38	ng/ul	98
19) Fluoranthene	19.066	202	7974	0.39	ng/ul	97
20) Pyrene	19.432	202	8108	0.39	ng/ul	99
21) Benzo(a)anthracene	21.176	228	7553	0.37	ng/ul	98
22) Chrysene	21.229	228	8066	0.37	ng/ul	99
24) Benzo(b)fluoranthene	22.716	252	7849	0.39	ng/ul	98
25) Benzo(k)fluoranthene	22.759	252	7883	0.39	ng/ul	98
26) Benzo(a)pyrene	23.270	252	6367	0.37	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.519	276	7379	0.32	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.533	278	5929	0.32	ng/ul	99
29) Benzo(g,h,i)perylene	26.175	276	6217	0.32	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072321\
Data File : BM031122.D
Acq On : 24 Jul 2021 16:30
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4065

Quant Time: Jul 26 02:07:20 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 23 18:04:27 2021
Response via : Initial Calibration

