

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072423\
 Data File : BM041091.D
 Acq On : 25 Jul 2023 01:03
 Operator : MA/JU
 Sample : 03534-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A46Y4

Quant Time: Jul 25 02:05:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 00:13:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.832	152	3374	0.400	ng/ul	0.00
4) Naphthalene-d8	10.636	136	11138	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.490	164	5541	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.248	188	11405	0.400	ng/ul	0.00
17) Chrysene-d12	21.451	240	10466	0.400	ng/ul	0.00
23) Perylene-d12	23.833	264	13875	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.250	96	23928	4.597	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.242	152	4799	0.302	ng/ul	0.00
18) Fluoranthene-d10	19.287	212	8942	0.314	ng/ul	0.00
Target Compounds						
					Qvalue	
21) Benzo(a)anthracene	21.439	228	736	0.021	ng/ul#	64
22) Chrysene	21.489	228	957	0.024	ng/ul#	73
24) Benzo(b)fluoranthene	23.108	252	1653	0.030	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072423\
Data File : BM041091.D
Acq On : 25 Jul 2023 01:03
Operator : MA/JU
Sample : 03534-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A46Y4

Quant Time: Jul 25 02:05:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 00:13:11 2023
Response via : Initial Calibration

