

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072621\
 Data File : BM031192.D
 Acq On : 27 Jul 2021 10:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4071

Quant Time: Jul 27 11:04:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 01:52:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.632	152	1077	0.40	ng/ul	-0.01
4) Naphthalene-d8	10.391	136	4089	0.40	ng/ul	-0.01
9) Acenaphthene-d10	14.257	164	2763	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.002	188	6073	0.40	ng/ul	-0.01
17) Chrysene-d12	21.190	240	4733	0.40	ng/ul	-0.01
23) Perylene-d12	23.360	264	3655	0.40	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	429	0.38	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.988	152	2782	0.39	ng/ul	-0.01
18) Fluoranthene-d10	19.031	212	6553	0.45	ng/ul	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.248	88	426	0.39	ng/ul	92
5) Naphthalene	10.445	128	4631	0.38	ng/ul	99
7) 2-Methylnaphthalene	12.060	142	3367	0.38	ng/ul	99
8) 1-Methylnaphthalene	12.285	142	3314	0.39	ng/ul	99
10) Acenaphthylene	13.980	152	4374	0.39	ng/ul	97
11) Acenaphthene	14.322	153	3739	0.37	ng/ul	99
12) Fluorene	15.311	166	4401	0.37	ng/ul	98
14) Pentachlorophenol	16.655	266	659	0.31	ng/ul	98
15) Phenanthrene	17.044	178	7269	0.38	ng/ul	99
16) Anthracene	17.134	178	6764	0.37	ng/ul	99
19) Fluoranthene	19.061	202	8272	0.43	ng/ul	98
20) Pyrene	19.424	202	8246	0.42	ng/ul	97
21) Benzo(a)anthracene	21.173	228	7064	0.37	ng/ul	99
22) Chrysene	21.226	228	7399	0.36	ng/ul	98
24) Benzo(b)fluoranthene	22.713	252	6583	0.38	ng/ul	98
25) Benzo(k)fluoranthene	22.756	252	6685	0.38	ng/ul	100
26) Benzo(a)pyrene	23.265	252	5376	0.36	ng/ul#	97
27) Indeno(1,2,3-cd)pyrene	25.511	276	6262	0.32	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.526	278	5020	0.32	ng/ul	98
29) Benzo(g,h,i)perylene	26.168	276	5262	0.32	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072621\
Data File : BM031192.D
Acq On : 27 Jul 2021 10:24
Operator : CG/JU
Sample : SSTDC00.4
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4071

Quant Time: Jul 27 11:04:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 01:52:16 2021
Response via : Initial Calibration

