

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072621\
 Data File : BM031154.D
 Acq On : 26 Jul 2021 09:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4068

Quant Time: Jul 26 16:00:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 26 15:58:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.632	152	1141	0.40	ng/ul	0.00
4) Naphthalene-d8	10.396	136	4260	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.257	164	2732	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.006	188	6035	0.40	ng/ul	0.00
17) Chrysene-d12	21.190	240	4984	0.40	ng/ul	0.00
23) Perylene-d12	23.360	264	4181	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	419	0.35	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.988	152	2812	0.38	ng/ul	0.00
18) Fluoranthene-d10	19.035	212	6592	0.43	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.248	88	376	0.32	ng/ul#	76
5) Naphthalene	10.445	128	4822	0.38	ng/ul	100
7) 2-Methylnaphthalene	12.064	142	3406	0.37	ng/ul	100
8) 1-Methylnaphthalene	12.285	142	3374	0.38	ng/ul	98
10) Acenaphthylene	13.980	152	4278	0.38	ng/ul	97
11) Acenaphthene	14.322	153	3716	0.38	ng/ul	97
12) Fluorene	15.311	166	4363	0.38	ng/ul	99
14) Pentachlorophenol	16.659	266	705	0.34	ng/ul	99
15) Phenanthrene	17.047	178	7298	0.38	ng/ul	99
16) Anthracene	17.137	178	6671	0.37	ng/ul	98
19) Fluoranthene	19.065	202	8428	0.42	ng/ul	98
20) Pyrene	19.427	202	8492	0.41	ng/ul	100
21) Benzo(a)anthracene	21.175	228	7430	0.37	ng/ul	99
22) Chrysene	21.226	228	7922	0.37	ng/ul	100
24) Benzo(b)fluoranthene	22.713	252	7639	0.39	ng/ul	99
25) Benzo(k)fluoranthene	22.756	252	7616	0.38	ng/ul	98
26) Benzo(a)pyrene	23.265	252	6199	0.36	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.511	276	7275	0.33	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.528	278	5909	0.33	ng/ul	99
29) Benzo(g,h,i)perylene	26.168	276	6177	0.33	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072621\
Data File : BM031154.D
Acq On : 26 Jul 2021 09:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4068

Quant Time: Jul 26 16:00:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 26 15:58:44 2021
Response via : Initial Calibration

