


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072621\
 Data File : BM031153.D
 Acq On : 26 Jul 2021 08:47
 Operator : CG/JU
 Sample : DFTPP008
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP008

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 27 05:01:15 2021



AutoFind: Scans 2443, 2444, 2445; Background Corrected with Scan 2436

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.6	54816	PASS
68	69	0.00	2	1.9	1088	PASS
69	198	0.00	100	51.9	56232	PASS
70	69	0.00	2	0.7	366	PASS
127	198	10	80	56.2	60944	PASS
197	198	0.00	2	0.7	730	PASS
198	198	100	100	100.0	108355	PASS
199	198	5	9	6.9	7487	PASS
275	198	10	60	28.2	30533	PASS
365	198	1	100	5.2	5640	PASS
441	198	0.01	100	11.4	12349	PASS
442	198	50	100	68.1	73803	PASS
443	442	15	24	18.4	13586	PASS

m/z	Abundance
39.90	198.0
41.10	211.0
43.00	533.0
44.00	1025.0
49.00	201.0
71.00	371.0
73.00	933.0
76.90	197.0
147.00	501.0
193.10	195.0
206.90	637.0
207.90	166.0
221.00	307.0
281.10	308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

m/z	Abundance
37.00	561.0
38.00	1104.0
39.10	5128.0
40.10	614.0
41.10	313.0
43.00	551.0
43.90	855.0
44.90	239.0
49.00	856.0
50.00	20096.0
51.00	66664.0
52.10	3032.0
52.90	203.0
54.90	454.0
56.00	2206.0
57.00	4305.0
58.00	176.0
61.00	1059.0
62.00	1211.0
63.00	3182.0
64.00	397.0
65.10	1475.0
67.20	236.0
68.00	1324.0
69.00	67792.0
70.00	328.0
71.00	349.0
73.10	1100.0
74.00	8005.0
75.00	11747.0
76.10	3970.0
77.00	73344.0
78.00	4561.0
79.00	6254.0
80.00	4000.0
81.00	6131.0
82.00	1437.0
82.90	995.0
85.00	931.0
86.00	2214.0
87.00	820.0
87.90	345.0
91.00	1280.0
92.00	1665.0
93.00	12119.0
93.90	578.0
95.00	248.0
95.90	473.0
97.00	328.0
98.00	9596.0
99.00	7094.0
100.00	605.0
101.00	3037.0
101.90	159.0
103.00	938.0
104.00	2216.0
105.00	2281.0
105.90	625.0
107.00	24584.0
108.00	3700.0
109.10	502.0
110.00	35096.0
111.00	6151.0
111.90	973.0
112.90	263.0
115.90	1217.0
117.00	29720.0
118.00	2189.0
119.00	300.0
120.00	282.0
121.10	163.0
122.00	2058.0
123.00	2662.0
124.00	1489.0
125.00	805.0
126.20	524.0
127.00	69704.0
128.00	5969.0
129.00	32384.0
130.00	2812.0
131.10	808.0
132.00	336.0
132.80	165.0
134.00	914.0
134.90	2605.0
136.00	1264.0
136.90	1638.0
137.80	211.0
139.10	266.0
140.00	404.0
141.00	4749.0
142.00	1444.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

143.00	969.0
145.90	898.0
147.00	2183.0
148.00	5864.0
148.90	1176.0
150.00	264.0
151.00	562.0
152.00	467.0
153.00	1249.0
153.90	812.0
155.00	2004.0
156.00	3075.0
157.10	754.0
157.90	620.0
158.90	560.0
159.90	1399.0
161.00	1885.0
162.00	649.0
163.10	190.0
165.00	1719.0
166.00	1317.0
167.00	9843.0
168.00	5094.0
168.90	906.0
170.00	251.0
171.00	308.0
171.90	733.0
172.90	885.0
173.90	1539.0
175.00	3365.0
175.90	773.0
177.00	1388.0
178.00	552.0
179.00	6159.0
180.00	3845.0
181.00	1673.0
182.00	361.0
184.00	389.0
185.00	2872.0
186.00	20016.0
187.00	5734.0
187.90	798.0
189.00	1853.0
190.00	470.0
191.00	781.0
192.00	2020.0
193.00	2627.0
194.10	347.0
194.90	226.0
195.90	3658.0
197.10	861.0
197.90	116680.0
198.90	8717.0
200.00	688.0
201.30	424.0
203.00	1070.0
204.00	5449.0
205.00	9204.0
206.00	35672.0
207.00	5926.0
208.00	1703.0
208.90	426.0
209.90	598.0
210.90	1492.0
211.80	302.0
214.90	376.0
216.00	903.0
217.00	10695.0
217.90	1378.0
218.90	176.0
219.90	306.0
221.00	4042.0
222.00	1469.0
223.00	2122.0
224.00	20160.0
225.00	5071.0
226.00	775.0
227.00	11212.0
228.00	1738.0
229.00	2094.0
230.00	284.0
231.00	712.0
231.90	184.0
232.90	169.0
233.90	648.0
235.00	897.0
236.00	530.0
236.90	894.0
237.90	171.0
239.10	480.0
239.90	197.0
240.90	555.0
242.00	722.0
243.00	894.0
244.00	14354.0
245.00	2003.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

246.00	3781.0
247.00	712.0
249.00	464.0
251.00	203.0
252.90	408.0
254.10	579.0
255.00	75176.0
256.00	11650.0
257.00	860.0
258.00	5750.0
259.00	909.0
260.00	243.0
264.00	267.0
265.00	2404.0
265.90	459.0
270.90	257.0
272.00	188.0
273.00	2013.0
274.00	6336.0
275.00	30008.0
276.00	4288.0
277.00	3789.0
277.90	590.0
279.00	197.0
283.10	237.0
284.20	219.0
285.00	575.0
289.00	172.0
293.00	529.0
294.00	159.0
294.80	178.0
296.00	10878.0
297.00	1784.0
301.90	200.0
303.00	1172.0
303.90	419.0
312.90	174.0
313.90	412.0
314.90	1273.0
316.00	566.0
321.00	316.0
322.00	200.0
323.00	2753.0
324.10	526.0
327.00	729.0
328.00	365.0
331.90	275.0
333.00	353.0
334.10	1734.0
335.00	479.0
341.00	195.0
345.90	640.0
346.90	163.0
352.00	875.0
352.90	383.0
354.00	833.0
355.10	159.0
365.00	5350.0
365.90	439.0
371.00	234.0
372.10	1234.0
372.90	398.0
383.00	427.0
390.10	187.0
391.00	172.0
401.90	607.0
403.00	654.0
403.90	282.0
421.00	741.0
422.10	498.0
423.00	3758.0
424.00	1058.0
441.10	10052.0
442.10	59088.0
443.10	11450.0
444.00	1062.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

m/z	Abundance
36.90	409.0
38.00	984.0
39.00	4817.0
40.00	385.0
41.00	363.0
42.90	666.0
43.90	1047.0
45.00	235.0
49.10	523.0
50.00	18864.0
51.00	62832.0
52.00	3119.0
53.00	177.0
55.00	424.0
56.00	2013.0
57.00	4410.0
58.00	190.0
61.00	743.0
62.00	1158.0
63.00	2510.0
64.00	296.0
65.00	991.0
67.00	204.0
68.10	1186.0
69.00	64992.0
70.00	568.0
71.00	461.0
73.10	1229.0
74.00	8381.0
75.00	11312.0
76.10	4164.0
77.10	70624.0
78.00	4539.0
79.00	5318.0
80.00	3873.0
81.00	5490.0
82.00	1338.0
83.00	977.0
83.90	214.0
84.90	842.0
86.00	1840.0
86.90	878.0
88.00	334.0
89.00	209.0
91.00	1331.0
92.00	1493.0
93.00	11034.0
93.90	630.0
96.00	405.0
97.10	221.0
98.00	8782.0
99.00	6616.0
99.90	584.0
101.00	3039.0
102.00	244.0
102.90	979.0
104.00	2009.0
105.00	2079.0
106.00	591.0
107.00	25816.0
108.00	3331.0
109.10	665.0
110.00	34408.0
111.00	5895.0
112.00	632.0
112.90	217.0
116.00	1470.0
117.00	31592.0
118.00	1882.0
119.00	346.0
119.90	291.0
121.10	194.0
122.00	1974.0
123.00	2829.0
124.00	1178.0
125.00	1056.0
126.20	451.0
127.00	72296.0
128.00	5737.0
129.00	32440.0
130.00	2751.0
130.90	447.0
132.10	332.0
133.00	197.0
134.00	989.0
135.00	2675.0
136.00	1170.0
137.00	1467.0
138.00	399.0
140.00	425.0
141.00	4624.0
142.00	1329.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

143.00	1085.0
144.00	183.0
145.00	188.0
146.10	910.0
147.00	2530.0
148.00	6531.0
148.90	1452.0
149.90	334.0
150.90	629.0
151.80	281.0
153.00	1244.0
154.00	989.0
155.00	2601.0
156.00	3628.0
157.00	719.0
157.90	845.0
158.90	624.0
160.00	1619.0
161.00	2072.0
162.00	654.0
163.00	157.0
164.00	239.0
165.00	1922.0
166.00	1454.0
167.00	9901.0
168.10	4591.0
169.00	636.0
170.00	343.0
171.10	305.0
172.00	834.0
172.90	866.0
174.00	2180.0
175.00	3828.0
176.00	963.0
177.00	1533.0
178.10	602.0
178.90	6390.0
180.00	4011.0
181.00	1694.0
181.90	337.0
184.00	585.0
185.00	3185.0
186.00	22552.0
187.00	6754.0
188.00	732.0
188.90	1675.0
190.00	565.0
190.90	867.0
192.00	2200.0
193.00	2452.0
194.10	520.0
195.00	260.0
196.00	3297.0
196.90	837.0
197.90	128128.0
198.90	8484.0
199.90	611.0
201.40	514.0
203.00	1451.0
204.00	6208.0
205.00	10297.0
206.00	38360.0
207.00	5928.0
208.00	1538.0
208.90	724.0
210.10	766.0
211.00	1559.0
211.60	482.0
215.00	769.0
215.90	1150.0
216.90	12451.0
217.90	1411.0
218.80	183.0
221.00	3913.0
221.90	1573.0
223.00	2339.0
224.00	24144.0
225.00	5699.0
225.90	528.0
227.00	12744.0
228.00	1557.0
229.00	2312.0
230.10	186.0
231.00	884.0
232.90	179.0
234.00	817.0
235.00	979.0
235.90	640.0
237.00	915.0
239.00	504.0
240.00	584.0
241.00	555.0
242.00	1164.0
243.00	1219.0
244.00	16960.0
245.00	2117.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

246.00	4411.0
247.00	847.0
247.90	342.0
249.00	642.0
250.90	257.0
252.10	283.0
253.00	481.0
254.20	934.0
255.00	91832.0
256.00	12724.0
257.00	927.0
258.00	6461.0
258.90	1015.0
265.00	2671.0
266.00	276.0
271.00	248.0
272.10	287.0
273.00	2516.0
274.00	6889.0
275.00	36872.0
276.00	5440.0
277.00	3993.0
278.00	779.0
278.90	160.0
280.90	244.0
281.90	150.0
283.00	470.0
284.00	405.0
284.90	802.0
288.90	219.0
289.90	249.0
291.90	180.0
292.90	771.0
293.80	353.0
294.80	171.0
295.20	184.0
296.00	14971.0
297.00	1883.0
298.00	170.0
301.90	269.0
303.00	1336.0
303.90	579.0
307.90	157.0
314.00	637.0
315.00	1526.0
316.00	1022.0
321.00	432.0
323.10	3605.0
324.10	657.0
326.90	727.0
327.80	269.0
332.00	301.0
333.00	347.0
334.10	2347.0
335.00	752.0
341.00	614.0
346.00	851.0
347.00	266.0
352.00	1142.0
353.00	809.0
354.10	1111.0
355.00	185.0
365.00	6852.0
366.00	742.0
371.00	432.0
372.00	1963.0
373.00	290.0
382.90	368.0
389.90	244.0
391.10	217.0
402.00	716.0
403.00	1115.0
404.10	317.0
421.00	943.0
422.10	565.0
423.10	5101.0
424.10	1147.0
440.00	246.0
441.10	15217.0
442.10	92568.0
443.10	16345.0
444.10	1798.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

m/z	Abundance
37.10	181.0
38.00	637.0
39.10	3172.0
40.00	404.0
40.90	294.0
43.00	602.0
44.00	825.0
49.00	356.0
50.00	9925.0
51.00	34952.0
52.00	1788.0
55.00	314.0
56.00	1035.0
57.00	2557.0
60.90	463.0
62.00	514.0
63.00	1597.0
64.00	200.0
65.00	620.0
68.00	754.0
69.00	35912.0
69.90	203.0
71.00	589.0
73.00	992.0
74.00	4217.0
75.00	6571.0
76.10	2197.0
77.10	39856.0
78.00	2758.0
79.00	3180.0
80.00	2310.0
81.00	3307.0
82.00	691.0
82.90	632.0
85.00	577.0
86.00	1301.0
86.90	364.0
88.00	210.0
91.00	580.0
92.00	1038.0
93.00	7204.0
93.90	416.0
96.10	248.0
98.00	5688.0
99.00	3595.0
100.00	391.0
101.00	1623.0
102.90	718.0
104.00	1146.0
104.90	1260.0
106.00	414.0
107.00	14316.0
108.00	2074.0
108.90	390.0
110.00	20136.0
110.90	3469.0
112.00	501.0
112.90	191.0
115.10	164.0
116.00	997.0
117.00	17496.0
117.90	1161.0
118.90	169.0
121.90	1127.0
123.00	1753.0
124.00	878.0
124.90	529.0
126.10	363.0
127.00	40832.0
128.00	3208.0
129.00	18696.0
129.90	1594.0
131.10	338.0
131.90	187.0
133.00	196.0
134.00	446.0
135.00	1544.0
135.90	438.0
137.00	708.0
138.90	165.0
139.90	416.0
141.00	2925.0
142.10	665.0
143.00	920.0
144.00	192.0
145.90	473.0
146.90	1732.0
148.00	3895.0
149.00	540.0
150.00	189.0
150.90	505.0
151.80	154.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

153.00	861.0
153.90	503.0
155.00	1587.0
156.00	1949.0
157.00	260.0
158.00	444.0
159.00	347.0
159.90	941.0
161.00	1188.0
161.80	474.0
164.90	1079.0
166.00	740.0
167.00	6445.0
168.00	3256.0
169.00	476.0
170.20	236.0
171.00	369.0
171.90	561.0
173.00	760.0
174.00	955.0
175.00	2278.0
176.00	472.0
177.00	951.0
177.90	302.0
179.00	4281.0
180.00	2674.0
181.00	1267.0
182.00	159.0
183.90	436.0
185.00	1825.0
186.00	12763.0
187.00	4060.0
188.00	396.0
189.00	1371.0
190.00	235.0
191.00	569.0
192.00	1302.0
193.00	1846.0
194.00	328.0
195.10	228.0
196.00	2338.0
197.10	493.0
197.90	80256.0
199.00	5260.0
199.90	505.0
201.30	386.0
201.80	239.0
202.90	1053.0
204.00	3933.0
205.00	6927.0
206.00	26568.0
207.00	4291.0
208.00	1020.0
209.00	582.0
210.10	558.0
211.00	1105.0
212.10	240.0
214.90	248.0
215.90	837.0
216.90	8646.0
218.00	1345.0
220.90	2367.0
221.90	973.0
223.00	1890.0
224.00	15415.0
225.00	3402.0
226.10	578.0
227.00	8344.0
228.00	1188.0
228.90	1402.0
229.90	252.0
230.90	637.0
234.00	547.0
235.00	319.0
236.00	304.0
237.10	360.0
239.00	319.0
240.00	291.0
241.00	315.0
242.00	681.0
243.00	905.0
244.00	11434.0
245.10	1464.0
246.00	2653.0
246.90	575.0
248.80	399.0
250.90	258.0
253.00	366.0
254.10	460.0
255.00	60848.0
256.00	9321.0
257.00	593.0
258.00	4211.0
259.00	624.0
265.00	2291.0
265.90	353.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008

271.00	243.0
271.90	172.0
273.00	2099.0
274.00	5107.0
275.00	24720.0
276.00	3476.0
277.00	2863.0
278.00	566.0
280.90	188.0
283.10	291.0
284.00	209.0
285.00	537.0
288.80	183.0
292.90	546.0
293.90	163.0
296.00	9696.0
297.00	1383.0
303.00	982.0
304.00	324.0
314.10	308.0
314.90	1028.0
316.10	528.0
321.00	199.0
321.90	174.0
323.00	2323.0
324.10	505.0
327.00	552.0
328.00	333.0
332.10	312.0
332.90	297.0
334.00	1697.0
335.10	460.0
341.10	300.0
346.00	546.0
346.90	189.0
351.90	838.0
353.00	841.0
353.90	873.0
355.00	320.0
365.00	4719.0
366.10	788.0
372.00	1154.0
373.10	305.0
382.90	273.0
389.90	183.0
402.10	505.0
403.00	532.0
403.80	253.0
420.90	550.0
422.00	523.0
423.00	3669.0
424.10	1069.0
424.90	192.0
440.20	167.0
441.10	11777.0
442.10	69752.0
443.10	12962.0
444.10	1192.0

Instrument :
BNA_M
ClientSampleId :
DFTPP008