

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072623\
 Data File : BM041137.D
 Acq On : 26 Jul 2023 16:03
 Operator : MA/JU
 Sample : 03553-17
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A46Z2

Quant Time: Jul 27 01:18:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 26 04:20:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.823	152	2827	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.631	136	9681	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.486	164	4871	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.244	188	9588	0.400	ng/ul	0.00
17) Chrysene-d12	21.454	240	7938	0.400	ng/ul	0.00
23) Perylene-d12	23.833	264	9907	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.250	96	19745	4.527	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.231	152	5012	0.363	ng/ul	0.00
18) Fluoranthene-d10	19.283	212	10607	0.491	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072623\
Data File : BM041137.D
Acq On : 26 Jul 2023 16:03
Operator : MA/JU
Sample : 03553-17
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A46Z2

Quant Time: Jul 27 01:18:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 26 04:20:13 2023
Response via : Initial Calibration

