

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072820\
 Data File : BM026889.D
 Acq On : 28 Jul 2020 14:02
 Operator : JU/CG
 Sample : PB130524BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK24

Quant Time: Jul 28 14:36:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072720MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 28 10:40:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	48632	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	180900	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.31	164	115879	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.05	188	248800	20.00	ng/ul	0.00
78) Chrysene-d12	21.25	240	255029	20.00	ng/ul	0.00
86) Perylene-d12	23.47	264	279818	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	7964	7.47	ng/uL	0.00
5) Phenol-d5	6.84	99	126556	29.71	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.01	67	84874	29.43	ng/ul	0.00
9) 2-Chlorophenol-d4	7.21	132	97011	30.42	ng/ul	0.00
13) 4-Methylphenol-d8	8.37	113	97793	29.83	ng/ul	0.00
19) Nitrobenzene-d5	8.81	128	40931	29.09	ng/ul	0.00
22) 2-Nitrophenol-d4	9.53	143	37364	28.97	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.07	165	90562	29.41	ng/ul	0.00
29) 4-Chloroaniline-d4	10.58	131	123413	33.61	ng/ul	0.00
44) Dimethylphthalate-d6	13.72	166	283106	31.16	ng/ul	0.00
47) Acenaphthylene-d8	14.00	160	334994	28.97	ng/ul	0.00
52) 4-Nitrophenol-d4	14.49	143	36506	23.75	ng/ul	0.00
58) Fluorene-d10	15.30	176	246341	31.17	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.41	200	22563	20.34	ng/ul	0.00
71) Anthracene-d10	17.15	188	384950	30.96	ng/ul	0.00
79) Pyrene-d10	19.45	212	412517	30.32	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.32	264	464804	29.44	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072820\
Data File : BM026889.D
Acq On : 28 Jul 2020 14:02
Operator : JU/CG
Sample : PB130524BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK24

Quant Time: Jul 28 14:36:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072720MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 28 10:40:31 2020
Response via : Initial Calibration

