

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM073119\
 Data File : BM021749.D
 Acq On : 31 Jul 2019 15:08
 Operator : HP/JU
 Sample : SSTD3.212
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.212

Quant Time: Jul 31 15:47:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM073119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 14:07:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	1600	0.40	ng/ul	0.00
2) Naphthalene-d8	10.45	136	3858	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.31	164	1642	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	4111	0.40	ng/ul	-0.01
16) Chrysene-d12	21.27	240	7960	0.40	ng/ul	0.00
20) Perylene-d12	23.51	264	7264	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.73	266	5137	5.088	ng/ul	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM073119\
Data File : BM021749.D
Acq On : 31 Jul 2019 15:08
Operator : HP/JU
Sample : SST3.212
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.212

Quant Time: Jul 31 15:47:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM073119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 14:07:07 2019
Response via : Initial Calibration

