

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM073119\
 Data File : BM021750.D
 Acq On : 31 Jul 2019 15:45
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV13

Quant Time: Jul 31 16:24:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM073119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:40:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
3	Naphthalene	1.010	1.017	-0.7	103	0.00
4 SURR	2-Methylnaphthalene-d10	0.477	0.469	1.7	100	0.00
5	2-Methylnaphthalene	0.533	0.510	4.3	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
7	Acenaphthylene	2.186	2.200	-0.6	97	0.00
8 C	Acenaphthene	1.611	1.613	-0.1	96	0.00
9	Fluorene	1.808	1.742	3.7	94	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
11	Pentachlorophenol	0.137	0.133	2.9	91	0.00
12	Phenanthrene	1.106	1.112	-0.5	91	0.00
13	Anthracene	0.952	0.972	-2.1	91	0.00
14 SURR	Fluoranthene-d10	1.150	1.106	3.8	86	0.00
15 C	Fluoranthene	1.382	1.296	6.2	86	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
17	Pyrene	1.133	1.258	-11.0	85	0.00
18	Benzo(a)anthracene	1.246	1.288	-3.4	78	0.00
19	Chrysene	1.571	1.620	-3.1	79	0.00
20 I	Perylene-d12	1.000	1.000	0.0	75	0.00
21	Benzo(b)fluoranthene	1.477	1.453	1.6	73	0.00
22	Benzo(k)fluoranthene	1.496	1.429	4.5	73	0.00
23 C	Benzo(a)pyrene	1.327	1.302	1.9	72	0.00
24	Indeno(1,2,3-cd)pyrene	1.776	1.859	-4.7	72	0.00
25	Dibenzo(a,h)anthracene	1.392	1.449	-4.1	72	0.00
26	Benzo(g,h,i)perylene	1.702	1.791	-5.2	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0