

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM073119\
Data File : BM021756.D
Acq On : 31 Jul 2019 22:21
Operator : HP/JU
Sample : PB120381BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK81

Manual Integrations
APPROVED

mohammad
8/1/2019 8:56:47 AM

Quant Time: Aug 01 02:55:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM073119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:40:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	2091	0.40	ng/ul	0.00
2) Naphthalene-d8	10.44	136	5398	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.31	164	2085	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	5087m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.28	240	5464m	0.40	ng/ul	0.00
20) Perylene-d12	23.50	264	6082m	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.06	152	1882	0.29	ng/ul	-0.01
14) Fluoranthene-d10	19.10	212	4640m	0.32	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM073119\
Data File : BM021756.D
Acq On : 31 Jul 2019 22:21
Operator : HP/JU
Sample : PB120381BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK81

Manual Integrations
APPROVED

mohammad
8/1/2019 8:56:47 AM

Quant Time: Aug 01 02:55:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM073119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:40:15 2019
Response via : Initial Calibration

