

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM073119\
Data File : BM021758.D
Acq On : 31 Jul 2019 23:34
Operator : HP/JU
Sample : K4128-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DB8L0

Manual Integrations
APPROVED

Sohil
8/1/2019 5:44:50 PM

Quant Time: Aug 01 03:10:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM073119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:40:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	2235	0.40	ng/ul	0.00
2) Naphthalene-d8	10.45	136	5818	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.31	164	2305	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.07	188	5450m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.27	240	6035m	0.40	ng/ul	0.00
20) Perylene-d12	23.50	264	5908m	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.07	152	1676	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.10	212	5492m	0.35	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM073119\
Data File : BM021758.D
Acq On : 31 Jul 2019 23:34
Operator : HP/JU
Sample : K4128-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DB8L0

Quant Time: Aug 01 03:10:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM073119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:40:15 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
8/1/2019 5:44:50 PM

