

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/31/2023	BNA_M	<u>BM041180.D</u>
Compound Name	Response	Retention Time
DDT	2268488	20.721
DDD	273647	20.28
DDE	0	19.768
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
273647	2542135	10.76