

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/31/2023	BNA_M	<u>BM041180.D</u>
Compound Name	Response	Retention Time
DDT	226804	20.715
DDD	428448	20.28
DDE	1719	19.762
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
430167	656971	65.48